

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q2150	OrderDate:	5/28/2025 3:53:00 PM
Client:	CDM Smith	Project:	South River WM Replacement
Contact:	Marcie Ann Encinas	Location:	L41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2150-01	TP-44	SOIL	PCB	8082A	05/27/25	05/30/25	05/30/25	05/28/25
Q2150-02	TP-42	SOIL	PCB	8082A	05/27/25	05/30/25	05/30/25	05/28/25
Q2150-03	TP-39	SOIL	PCB	8082A	05/27/25	05/30/25	05/30/25	05/28/25
Q2150-04	TP-48	SOIL	PCB	8082A	05/27/25	05/30/25	05/30/25	05/28/25
Q2150-05	TP-47	SOIL	PCB	8082A	05/27/25	05/30/25	05/30/25	05/28/25
Q2150-06	TP-50	SOIL	PCB	8082A	05/27/25	05/30/25	05/30/25	05/28/25
Q2150-07	TP-51	SOIL	PCB	8082A	05/27/25	05/30/25	05/30/25	05/28/25
Q2150-08	TP-52	SOIL	PCB	8082A	05/28/25	05/30/25	05/30/25	05/28/25
Q2150-09	TP-54	SOIL	PCB	8082A	05/28/25	05/30/25	05/30/25	05/28/25
Q2150-10	TP-53	SOIL	PCB	8082A	05/28/25	05/30/25	05/30/25	05/28/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2150

Order ID: Q2150

Client: CDM Smith

Project ID: South River WM Replacement

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: **Q2150**

Client: **CDM Smith**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072163.D	PIBLK-PP072163.D	Tetrachloro-m-xylene	1	20	16.3	82		60	140
		Decachlorobiphenyl	1	20	16.7	84		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
		Decachlorobiphenyl	2	20	17.0	85		60	140
I.BLK-PP072465.D	PIBLK-PP072465.D	Tetrachloro-m-xylene	1	20	18.1	91		60	140
		Decachlorobiphenyl	1	20	19.5	98		60	140
		Tetrachloro-m-xylene	2	20	20.0	100		60	140
		Decachlorobiphenyl	2	20	21.7	108		60	140
PB168208BL	PB168208BL	Tetrachloro-m-xylene	1	20	19.8	99		32	144
		Decachlorobiphenyl	1	20	21.2	106		32	175
		Tetrachloro-m-xylene	2	20	21.6	108		32	144
		Decachlorobiphenyl	2	20	24.3	122		32	175
PB168208BS	PB168208BS	Tetrachloro-m-xylene	1	20	20.2	101		32	144
		Decachlorobiphenyl	1	20	21.4	107		32	175
		Tetrachloro-m-xylene	2	20	20.2	101		32	144
		Decachlorobiphenyl	2	20	23.6	118		32	175
Q2150-01	TP-44	Tetrachloro-m-xylene	1	20	17.1	85		32	144
		Decachlorobiphenyl	1	20	14.5	72		32	175
		Tetrachloro-m-xylene	2	20	19.8	99		32	144
		Decachlorobiphenyl	2	20	16.3	81		32	175
Q2150-02	TP-42	Tetrachloro-m-xylene	1	20	17.7	89		32	144
		Decachlorobiphenyl	1	20	12.6	63		32	175
		Tetrachloro-m-xylene	2	20	20.4	102		32	144
		Decachlorobiphenyl	2	20	14.9	75		32	175
Q2150-02MS	TP-42MS	Tetrachloro-m-xylene	1	20	18.1	90		32	144
		Decachlorobiphenyl	1	20	14.3	72		32	175
		Tetrachloro-m-xylene	2	20	18.8	94		32	144
		Decachlorobiphenyl	2	20	16.1	81		32	175
Q2150-02MSD	TP-42MSD	Tetrachloro-m-xylene	1	20	18.9	94		32	144
		Decachlorobiphenyl	1	20	15.2	76		32	175
		Tetrachloro-m-xylene	2	20	19.8	99		32	144
		Decachlorobiphenyl	2	20	16.8	84		32	175
Q2150-03	TP-39	Tetrachloro-m-xylene	1	20	16.4	82		32	144
		Decachlorobiphenyl	1	20	10.9	55		32	175
		Tetrachloro-m-xylene	2	20	18.7	94		32	144
		Decachlorobiphenyl	2	20	12.0	60		32	175
Q2150-04	TP-48	Tetrachloro-m-xylene	1	20	16.1	81		32	144
		Decachlorobiphenyl	1	20	14.3	71		32	175
		Tetrachloro-m-xylene	2	20	17.5	88		32	144
		Decachlorobiphenyl	2	20	16.3	81		32	175
Q2150-05	TP-47	Tetrachloro-m-xylene	1	20	17.7	88		32	144

Surrogate Summary

SDG No.: **Q2150**

Client: **CDM Smith**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2150-05	TP-47	Decachlorobiphenyl	1	20	13.1	66		32	175
		Tetrachloro-m-xylene	2	20	20.3	102		32	144
Q2150-06	TP-50	Decachlorobiphenyl	2	20	14.0	70		32	175
		Tetrachloro-m-xylene	1	20	18.0	90		32	144
		Decachlorobiphenyl	1	20	16.9	85		32	175
		Tetrachloro-m-xylene	2	20	20.3	101		32	144
		Decachlorobiphenyl	2	20	17.8	89		32	175
		Tetrachloro-m-xylene	1	20	18.7	93		60	140
I.BLK-PP072480.D	PIBLK-PP072480.D	Decachlorobiphenyl	1	20	19.4	97		60	140
		Tetrachloro-m-xylene	2	20	20.1	100		60	140
		Decachlorobiphenyl	2	20	21.4	107		60	140
		Tetrachloro-m-xylene	1	20	18.1	91		32	144
Q2150-07	TP-51	Decachlorobiphenyl	1	20	11.8	59		32	175
		Tetrachloro-m-xylene	2	20	20.6	103		32	144
		Decachlorobiphenyl	2	20	14.1	70		32	175
		Tetrachloro-m-xylene	1	20	17.7	89		32	144
Q2150-08	TP-52	Decachlorobiphenyl	1	20	11.9	60		32	175
		Tetrachloro-m-xylene	2	20	20.1	101		32	144
		Decachlorobiphenyl	2	20	14.7	74		32	175
		Tetrachloro-m-xylene	1	20	15.8	79		32	144
Q2150-09	TP-54	Decachlorobiphenyl	1	20	12.2	61		32	175
		Tetrachloro-m-xylene	2	20	17.1	86		32	144
		Decachlorobiphenyl	2	20	13.7	69		32	175
		Tetrachloro-m-xylene	1	20	18.4	92		32	144
Q2150-10	TP-53	Decachlorobiphenyl	1	20	14.1	70		32	175
		Tetrachloro-m-xylene	2	20	20.2	101		32	144
		Decachlorobiphenyl	2	20	16.0	80		32	175
		Tetrachloro-m-xylene	1	20	18.4	92		60	140
I.BLK-PP072495.D	PIBLK-PP072495.D	Decachlorobiphenyl	1	20	19.5	98		60	140
		Tetrachloro-m-xylene	2	20	21.1	106		60	140
		Decachlorobiphenyl	2	20	21.1	105		60	140
		Tetrachloro-m-xylene	1	20	18.4	92		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2150</u>	Analytical Method:	<u>8082A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PP072470.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2150-02MS (Column 1)	TP-42MS AR1016	193.3	0	72.6	ug/kg	38	*			55	146	
	AR1260	193.3	0	64.4	ug/kg	33				31	146	
Client Sample ID: Q2150-02MS (Column 2)	TP-42MS AR1016	193.3	0	85.6	ug/kg	44	*			55	146	
	AR1260	193.3	0	77.2	ug/kg	40				31	146	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2150</u>	Analytical Method:	<u>8082A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PP072471.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2150-02MSD (Column 1)	TP-42MSD AR1016	193.2	0	74.5	ug/kg	39	*	3		55	146	20
	AR1260	193.2	0	67.3	ug/kg	35		6		31	146	20
Client Sample ID: Q2150-02MSD (Column 2)	TP-42MSD AR1016	193.2	0	86.5	ug/kg	45	*	2		55	146	20
	AR1260	193.2	0	79.3	ug/kg	41		2		31	146	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2150</u>	Analytical Method:	<u>8082A</u>
Client:	<u>CDM Smith</u>	Datafile :	<u>PP072467.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168208BS (Column 1)	AR1016	166.5	121	ug/kg	73				71	120	
	AR1260	166.5	123	ug/kg	74				65	130	
PB168208BS (Column 2)	AR1016	166.5	129	ug/kg	77				71	120	
	AR1260	166.5	134	ug/kg	80				65	130	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168208BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150

SAS No.: Q2150 SDG NO.: Q2150

Lab Sample ID: PB168208BL

Lab File ID: PP072466.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 05/30/2025

Date Analyzed (1): 05/30/2025

Date Analyzed (2): 05/30/2025

Time Analyzed (1): 12:59

Time Analyzed (2): 12:59

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
PB168208BS	PB168208BS	PP072467.D	05/30/2025	05/30/2025
TP-44	Q2150-01	PP072468.D	05/30/2025	05/30/2025
TP-42	Q2150-02	PP072469.D	05/30/2025	05/30/2025
TP-42MS	Q2150-02MS	PP072470.D	05/30/2025	05/30/2025
TP-42MSD	Q2150-02MSD	PP072471.D	05/30/2025	05/30/2025
TP-39	Q2150-03	PP072472.D	05/30/2025	05/30/2025
TP-48	Q2150-04	PP072473.D	05/30/2025	05/30/2025
TP-47	Q2150-05	PP072474.D	05/30/2025	05/30/2025
TP-50	Q2150-06	PP072475.D	05/30/2025	05/30/2025
TP-51	Q2150-07	PP072481.D	05/30/2025	05/30/2025
TP-52	Q2150-08	PP072482.D	05/30/2025	05/30/2025
TP-54	Q2150-09	PP072483.D	05/30/2025	05/30/2025
TP-53	Q2150-10	PP072484.D	05/30/2025	05/30/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-44	SDG No.:	Q2150
Lab Sample ID:	Q2150-01	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	78.9
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072468.D	1	05/30/25 08:50	05/30/25 13:32	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.00	U	5.00	21.5	ug/kg
11104-28-2	Aroclor-1221	5.10	U	5.10	21.5	ug/kg
11141-16-5	Aroclor-1232	4.70	U	4.70	21.5	ug/kg
53469-21-9	Aroclor-1242	5.10	U	5.10	21.5	ug/kg
12672-29-6	Aroclor-1248	7.50	U	7.50	21.5	ug/kg
11097-69-1	Aroclor-1254	4.10	U	4.10	21.5	ug/kg
37324-23-5	Aroclor-1262	6.40	U	6.40	21.5	ug/kg
11100-14-4	Aroclor-1268	4.60	U	4.60	21.5	ug/kg
11096-82-5	Aroclor-1260	4.10	U	4.10	21.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.8		32 - 144	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.3		32 - 175	81%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072468.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 13:32
 Operator : YP\AJ
 Sample : Q2150-01
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-44

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 14:38:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.796	34341788	31646236	17.079m	19.799
2) SA Decachlor...	10.210	8.817	23560625	17224699	14.462	16.271

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072468.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 13:32
Operator : YP\AJ
Sample : Q2150-01
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-44

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 14:38:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.500 min
Delta R.T.: 0.002 min
Response: 34341788
Conc: 17.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-44




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.002 min
Response: 31646236
Conc: 19.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.002 min
Response: 23560625
Conc: 14.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 17224699
Conc: 16.27 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-42	SDG No.:	Q2150
Lab Sample ID:	Q2150-02	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.2
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072469.D	1	05/30/25 08:50	05/30/25 13:48	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.7	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.7	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.7	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.7	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.7	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.7	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.7	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.7	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.4		32 - 144	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.9		32 - 175	75%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072469.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 13:48
Operator : YP\AJ
Sample : Q2150-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-42

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 14:38:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	35639212	32633005	17.724	20.417
2) SA Decachlor...	10.208	8.817	20572602	15817434	12.628	14.942

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072469.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 13:48
Operator : YP\AJ
Sample : Q2150-02
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-42

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 14:38:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.501 min
Delta R.T.: 0.003 min
Response: 35639212
Conc: 17.72 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-42




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.004 min
Response: 32633005
Conc: 20.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.208 min
Delta R.T.: -0.004 min
Response: 20572602
Conc: 12.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 15817434
Conc: 14.94 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-39	SDG No.:	Q2150
Lab Sample ID:	Q2150-03	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	83.7
Sample Wt/Vol:	30.08	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072472.D	1	05/30/25 08:50	05/30/25 14:37	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.3	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.3	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.3	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.3	ug/kg
12672-29-6	Aroclor-1248	7.10	U	7.10	20.3	ug/kg
11097-69-1	Aroclor-1254	3.80	U	3.80	20.3	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.3	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.3	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.7		32 - 144	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	12.0		32 - 175	60%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072472.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 14:37
 Operator : YP\AJ
 Sample : Q2150-03
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-39

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 16:53:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	32976165	29923480	16.400	18.721
2) SA Decachlor...	10.211	8.819	17833442	12678101	10.947	11.976

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072472.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 14:37
Operator : YP\AJ
Sample : Q2150-03
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-39

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 16:53:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.503 min
Delta R.T.: 0.005 min
Response: 32976165
Conc: 16.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-39




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.004 min
Response: 29923480
Conc: 18.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.211 min
Delta R.T.: -0.001 min
Response: 17833442
Conc: 10.95 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.819 min
Delta R.T.: -0.002 min
Response: 12678101
Conc: 11.98 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-48	SDG No.:	Q2150
Lab Sample ID:	Q2150-04	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	80.7
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072473.D	1	05/30/25 08:50	05/30/25 14:53	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.0	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.0	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	21.0	ug/kg
53469-21-9	Aroclor-1242	5.00	U	5.00	21.0	ug/kg
12672-29-6	Aroclor-1248	7.30	U	7.30	21.0	ug/kg
11097-69-1	Aroclor-1254	4.00	U	4.00	21.0	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	21.0	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.0	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.5		32 - 144	88%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.3		32 - 175	81%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072473.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 14:53
Operator : YP\AJ
Sample : Q2150-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-48

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 16:53:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.797	32460405	28030465	16.143	17.537
2) SA Decachlor...	10.212	8.817	23231046	17205688	14.260	16.253

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072473.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 14:53
Operator : YP\AJ
Sample : Q2150-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-48

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 16:53:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.504 min
Delta R.T.: 0.006 min
Response: 32460405
Conc: 16.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-48




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.004 min
Response: 28030465
Conc: 17.54 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.212 min
Delta R.T.: 0.000 min
Response: 23231046
Conc: 14.26 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 17205688
Conc: 16.25 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-47	SDG No.:	Q2150
Lab Sample ID:	Q2150-05	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	81
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072474.D	1	05/30/25 08:50	05/30/25 15:10	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.0	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.0	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	21.0	ug/kg
53469-21-9	Aroclor-1242	4.90	U	4.90	21.0	ug/kg
12672-29-6	Aroclor-1248	7.30	U	7.30	21.0	ug/kg
11097-69-1	Aroclor-1254	4.00	U	4.00	21.0	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	21.0	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	21.0	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.3		32 - 144	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.0		32 - 175	70%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 15:10
Operator : YP\AJ
Sample : Q2150-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-47

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 16:54:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.798	35563776	32446850	17.686	20.300
2) SA Decachlor...	10.210	8.818	21411289	14804442	13.143	13.985

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072474.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 15:10
Operator : YP\AJ
Sample : Q2150-05
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-47

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 16:54:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.502 min
Delta R.T.: 0.004 min
Response: 35563776
Conc: 17.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-47




#1 Tetrachloro-m-xylene

R.T.: 3.798 min
Delta R.T.: 0.004 min
Response: 32446850
Conc: 20.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.002 min
Response: 21411289
Conc: 13.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.818 min
Delta R.T.: -0.003 min
Response: 14804442
Conc: 13.98 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-50	SDG No.:	Q2150
Lab Sample ID:	Q2150-06	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072475.D	1	05/30/25 08:50	05/30/25 15:26	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.8	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.8	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.8	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	19.8	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.8	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.8	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.8	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.8	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	19.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.3		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.8		32 - 175	89%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 15:26
Operator : YP\AJ
Sample : Q2150-06
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-50

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 16:54:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.797	36131864	32392944	17.969m	20.266
2) SA Decachlor...	10.209	8.817	27566306	18859144	16.921	17.815

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 15:26
Operator : YP\AJ
Sample : Q2150-06
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-50

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 16:54:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.500 min
Delta R.T.: 0.002 min
Response: 36131864
Conc: 17.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-50




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.004 min
Response: 32392944
Conc: 20.27 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: -0.003 min
Response: 27566306
Conc: 16.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 18859144
Conc: 17.82 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-51	SDG No.:	Q2150
Lab Sample ID:	Q2150-07	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	83.7
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072481.D	1	05/30/25 08:50	05/30/25 17:37	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.3	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.3	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.3	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.3	ug/kg
12672-29-6	Aroclor-1248	7.10	U	7.10	20.3	ug/kg
11097-69-1	Aroclor-1254	3.80	U	3.80	20.3	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.3	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.3	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.6		32 - 144	103%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.1		32 - 175	70%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072481.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 17:37
Operator : YP\AJ
Sample : Q2150-07
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-51

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:53:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.796	36425596	32880737	18.115	20.572
2) SA Decachlor...	10.210	8.817	19273907	14870776	11.831	14.048

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072481.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 17:37
Operator : YP\AJ
Sample : Q2150-07
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-51

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:53:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.503 min
Delta R.T.: 0.005 min
Response: 36425596
Conc: 18.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-51




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.003 min
Response: 32880737
Conc: 20.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.002 min
Response: 19273907
Conc: 11.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.005 min
Response: 14870776
Conc: 14.05 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/28/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-52	SDG No.:	Q2150
Lab Sample ID:	Q2150-08	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	85.7
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072482.D	1	05/30/25 08:50	05/30/25 17:53	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.8	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.8	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.8	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	19.8	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.8	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.8	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	19.8	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.8	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	19.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.1		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.7		32 - 175	74%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072482.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 17:53
Operator : YP\AJ
Sample : Q2150-08
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-52

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:54:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.498	3.797	35653212	32134261	17.731m	20.105
2) SA Decachlor...	10.207	8.817	19468810	15600392	11.950	14.737

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072482.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 17:53
Operator : YP\AJ
Sample : Q2150-08
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-52

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:54:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 35653212
Conc: 17.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-52




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.004 min
Response: 32134261
Conc: 20.10 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.207 min
Delta R.T.: -0.005 min
Response: 19468810
Conc: 11.95 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 15600392
Conc: 14.74 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/28/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-54	SDG No.:	Q2150
Lab Sample ID:	Q2150-09	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	83.6
Sample Wt/Vol:	30.09	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072483.D	1	05/30/25 08:50	05/30/25 18:09	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.3	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.3	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.3	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.3	ug/kg
12672-29-6	Aroclor-1248	7.10	U	7.10	20.3	ug/kg
11097-69-1	Aroclor-1254	3.80	U	3.80	20.3	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.3	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.3	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		32 - 144	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.7		32 - 175	69%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072483.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 18:09
 Operator : YP\AJ
 Sample : Q2150-09
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-54

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 22:54:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.503	3.797	31690463	27413314	15.760	17.151
2)	SA Decachlor...	10.210	8.817	19942262	14543074	12.241	13.738

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072483.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 18:09
Operator : YP\AJ
Sample : Q2150-09
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-54

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:54:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.503 min
Delta R.T.: 0.005 min
Response: 31690463
Conc: 15.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-54




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.004 min
Response: 27413314
Conc: 17.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.002 min
Response: 19942262
Conc: 12.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 14543074
Conc: 13.74 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/28/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-53	SDG No.:	Q2150
Lab Sample ID:	Q2150-10	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	85.3
Sample Wt/Vol:	30.05	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072484.D	1	05/30/25 08:50	05/30/25 18:26	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.9	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.9	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	19.9	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	19.9	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.9	ug/kg
11097-69-1	Aroclor-1254	3.80	U	3.80	19.9	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	19.9	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.9	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	19.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.0		32 - 175	80%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072484.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 18:26
Operator : YP\AJ
Sample : Q2150-10
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-53

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:55:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	37059545	32282154	18.430	20.197
2) SA Decachlor...	10.209	8.818	22959967	16975587	14.093	16.036

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072484.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 18:26
Operator : YP\AJ
Sample : Q2150-10
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-53

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:55:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.501 min
Delta R.T.: 0.003 min
Response: 37059545
Conc: 18.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-53




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.004 min
Response: 32282154
Conc: 20.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: -0.003 min
Response: 22959967
Conc: 14.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.818 min
Delta R.T.: -0.003 min
Response: 16975587
Conc: 16.04 ng/ml



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Contract: **CAMP02**

Lab Code: CHEM

Case No.: Q2150

SAS No.: Q2150

SDG NO.: Q2150

Instrument ID: ECD_P

Calibration Date(s):

05/19/2025

05/19/2025

Calibration Times:

09:42

18:25

GC Column: ZB-MR1

ID: 0.32 (mm)

LAB FILE ID:

RT 1000 = PP072164.D

RT 750 = PP072165.D

RT 500 = PP072166.D

RT 250 = PP072167.D

RT 050 = PP072168.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.22	10.22	10.22	10.22	10.22	10.22	10.12	10.32
Tetrachloro-m-xylene	4.50	4.50	4.50	4.50	4.50	4.50	4.40	4.60
Aroclor-1268-1 (1)	8.72	8.72	8.72	8.72	8.72	8.72	8.62	8.82
Aroclor-1268-2 (2)	8.81	8.82	8.81	8.81	8.81	8.81	8.71	8.91
Aroclor-1268-3 (3)	9.04	9.05	9.04	9.04	9.05	9.04	8.94	9.14
Aroclor-1268-4 (4)	9.46	9.46	9.46	9.46	9.46	9.46	9.36	9.56
Aroclor-1268-5 (5)	9.88	9.88	9.88	9.88	9.88	9.88	9.78	9.98
Decachlorobiphenyl	10.22	10.22	10.22	10.21	10.22	10.22	10.12	10.32
Tetrachloro-m-xylene	4.50	4.51	4.50	4.50	4.50	4.50	4.40	4.60

RETENTION TIMES OF INITIAL CALIBRATION

Contract: **CAMP02**

Lab Code: CHEM

Case No.: Q2150

SAS No.: Q2150

SDG NO.: Q2150

Instrument ID: ECD_P

Calibration Date(s):

05/19/2025

05/19/2025

Calibration Times:

09:42

18:25

GC Column: ZB-MR2

ID: 0.32 (mm)

LAB FILE ID:

RT 1000 = PP072164.D

RT 750 = PP072165.D

RT 500 = PP072166.D

RT 250 = PP072167.D

RT 050 = PP072168.D

[illegible]

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02
Lab Code: CHEM **Case No.:** Q2150 **SAS No.:** Q2150 **SDG NO.:** Q2150
Instrument ID: ECD_P **Calibration Date(s):** 05/19/2025 05/19/2025
Calibration Times: 09:42 18:25

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP072164.D</u>		CF 750 = <u>PP072165.D</u>			
CF 500 = <u>PP072166.D</u>		CF 250 = <u>PP072167.D</u>		CF 050 = <u>PP072168.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	66878491	69828851	76010044	83211968	79488900	75083651
Aroclor-1016-2	(2)	99468528	104467549	110518222	118579572	101924480	106991670
Aroclor-1016-3	(3)	61412187	64234915	67714060	73550892	60912680	65564947
Aroclor-1016-4	(4)	50363803	53404009	55311866	58220708	49049560	53269989
Aroclor-1016-5	(5)	45437165	47941741	49914606	51189784	50062640	48909187
Aroclor-1260-1	(1)	85553757	90346417	93957158	98060904	100055640	93594775
Aroclor-1260-2	(2)	135650987	137755076	143609204	149400496	151032820	143489717
Aroclor-1260-3	(3)	109042220	112821207	117388622	121236852	108865120	113870804
Aroclor-1260-4	(4)	102101329	107107725	110940768	114591832	101771400	107302611
Aroclor-1260-5	(5)	228876095	233706888	240834266	249937120	229956720	236662218
Decachlorobiphenyl		1538266220	1617591187	1654764640	1712186920	1622812600	1629124313
Tetrachloro-m-xylene		1944699760	2026562000	2072807240	2135906680	1874026800	2010800496
Aroclor-1232-1	(1)	44075653	45896815	48002922	48247276	49319600	47108453
Aroclor-1232-2	(2)	21717379	22928696	23244388	23029808	25495660	23283186
Aroclor-1232-3	(3)	46772523	50148184	51844656	54382920	47167260	50063109
Aroclor-1232-4	(4)	23238504	24146713	24804390	26183196	27122440	25099049
Aroclor-1232-5	(5)	15752983	16470224	16442412	18109320	20822780	17519544
Decachlorobiphenyl		1545279190	1600827560	1672901480	1703701720	1499866000	1604515190
Tetrachloro-m-xylene		1886122400	2000973440	2054641840	2122290120	1923172600	1997440080
Aroclor-1242-1	(1)	56211553	57323072	62083360	65843868	71524840	62597339
Aroclor-1242-2	(2)	82973080	85971496	91883334	97584776	75486540	86779845
Aroclor-1242-3	(3)	51559127	52121972	55496172	57080948	59071260	55065896
Aroclor-1242-4	(4)	41207603	43030657	45948576	46781836	44776460	44349026
Aroclor-1242-5	(5)	47977894	48696228	50879554	52435600	54934800	50984815
Decachlorobiphenyl		1526214180	1567133680	1645977640	1728993240	1487418400	1591147428
Tetrachloro-m-xylene		1827808490	1928927520	1991275140	2114690400	1713029600	1915146230
Aroclor-1248-1	(1)	44038397	47472432	47912646	51445364	51108500	48395468
Aroclor-1248-2	(2)	56984732	59400859	61165396	63476680	62172680	60640069
Aroclor-1248-3	(3)	64677132	65813748	68767932	70638012	61003720	66180109
Aroclor-1248-4	(4)	81606824	85304428	88412180	92196600	89440360	87392078
Aroclor-1248-5	(5)	77587697	81003728	83211240	86851844	86474060	83025714
Decachlorobiphenyl		1573368300	1599269360	1663196880	1714969280	1478724200	1605905604
Tetrachloro-m-xylene		1880240370	1909383627	1990709860	2112434560	1856348600	1949823403
Aroclor-1254-1	(1)	79792123	82783284	86922424	91684028	88834440	86003260

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1254-2	(2)	120142096	124850076	131105324	134547092	142741880	130677294	7
Aroclor-1254-3	(3)	124307316	127832729	133782854	136984312	137108200	132003082	4
Aroclor-1254-4	(4)	116304358	118550812	124418220	136069712	157210520	130510724	13
Aroclor-1254-5	(5)	107938622	112056604	116315218	117138120	102784720	111246657	5
Decachlorobiphenyl		1576500070	1614032093	1688035380	1708021160	1475084400	1612334621	6
Tetrachloro-m-xylene		1869066560	1948875347	2032435420	2069900400	1617536400	1907562825	9
Aroclor-1268-1	(1)	315692357	322966827	337626016	351356792	338115380	333151474	4
Aroclor-1268-2	(2)	265886956	272371317	285215132	296487980	279979560	279988189	4
Aroclor-1268-3	(3)	227961300	235081659	247250176	255747588	241284760	241465097	4
Aroclor-1268-4	(4)	101392203	104765767	107968348	108360296	102254740	104948271	3
Aroclor-1268-5	(5)	651541042	667177075	686960040	706355632	672967440	677000246	3
Decachlorobiphenyl		2733685450	2815599213	2937966700	3038738360	2636754200	2832548785	6
Tetrachloro-m-xylene		1849618000	1900065360	1995699220	2054630680	1744336400	1908869932	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02
Lab Code: CHEM **Case No.:** Q2150 **SAS No.:** Q2150 **SDG NO.:** Q2150
Instrument ID: ECD_P **Calibration Date(s):** 05/19/2025 05/19/2025
Calibration Times: 09:42 18:25

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP072164.D</u>		CF 750 = <u>PP072165.D</u>			
CF 500 = <u>PP072166.D</u>		CF 250 = <u>PP072167.D</u>		CF 050 = <u>PP072168.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	54525663	58849481	59632288	64515432	69624760	61429525
Aroclor-1016-2	(2)	79772717	86154421	85288758	92453296	95277560	87789350
Aroclor-1016-3	(3)	42641318	46352395	47330430	50989192	50959120	47654491
Aroclor-1016-4	(4)	33654246	36754908	37811202	41521120	40485160	38045327
Aroclor-1016-5	(5)	43791070	47744144	49105368	53701332	54780200	49824423
Aroclor-1260-1	(1)	72244144	75307759	80373306	84844232	85999160	79753720
Aroclor-1260-2	(2)	87664004	90365408	98490088	104924488	125666920	101422182
Aroclor-1260-3	(3)	79909091	81326363	88937536	92494016	91844780	86902357
Aroclor-1260-4	(4)	64349922	67685321	72647720	77711116	78060640	72090944
Aroclor-1260-5	(5)	161315528	166937577	177635780	187517596	174713540	173624004
Decachlorobiphenyl		920208980	1044964960	1093136160	1163579000	1071118200	1058601460
Tetrachloro-m-xylene		1547649930	1584243107	1593704220	1639561360	1626633000	1598358323
Aroclor-1232-1	(1)	37616932	40592059	45290150	47667724	45555860	43344545
Aroclor-1232-2	(2)	39358353	42024527	44532022	47402620	47585980	44180700
Aroclor-1232-3	(3)	20796080	22347549	23808552	25172964	24790340	23383097
Aroclor-1232-4	(4)	17897124	19415229	20680720	21689292	23918420	20720157
Aroclor-1232-5	(5)	19472370	21237964	22534450	24219412	28307940	23154427
Decachlorobiphenyl		991061470	1069964400	1120483020	1147415120	1151969600	1096178722
Tetrachloro-m-xylene		1503650850	1560846907	1685979140	1671096560	1578260800	1599966851
Aroclor-1242-1	(1)	46123670	49902476	52843000	57237560	54435320	52108405
Aroclor-1242-2	(2)	66003789	71039361	75471286	80940564	74026140	73496228
Aroclor-1242-3	(3)	34394227	38777200	41064482	44714476	36677280	39125533
Aroclor-1242-4	(4)	32146418	36854119	38826456	43192656	36894340	37582798
Aroclor-1242-5	(5)	41776015	46613400	48754570	54226988	49917340	48257663
Decachlorobiphenyl		967171680	1074707920	1046164720	1163123680	1078294400	1065892480
Tetrachloro-m-xylene		1545621850	1670448987	1600956100	1623877080	1377345600	1563649923
Aroclor-1248-1	(1)	38942472	38080832	42980810	46736940	49867020	43321615
Aroclor-1248-2	(2)	50685144	49386624	55814958	61079164	65420760	56477330
Aroclor-1248-3	(3)	53612101	51937139	58301874	64007240	67677000	59107071
Aroclor-1248-4	(4)	63542852	61081784	68799052	75621092	79219440	69652844
Aroclor-1248-5	(5)	63751699	62374025	69241268	74501112	78829180	69739457
Decachlorobiphenyl		1077930460	1066864453	1129616840	1218519160	1187538600	1136093903
Tetrachloro-m-xylene		1585370290	1544149787	1686359520	1807980440	1717995000	1668371007
Aroclor-1254-1	(1)	88394708	92134828	103452628	109580892	106208260	99954263

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1254-2	(2)	75158218	78808232	90151838	93786952	93629340	86306916	10
Aroclor-1254-3	(3)	118588697	123091027	137573488	140955076	122895140	128620686	8
Aroclor-1254-4	(4)	76184182	79817879	90049888	92363636	88066400	85296397	8
Aroclor-1254-5	(5)	103321222	111630831	120507404	120857332	110441220	113351602	7
Decachlorobiphenyl		1048618240	1081081133	1110321040	1211638680	1138223000	1117976419	6
Tetrachloro-m-xylene		1515554540	1677026120	1758978840	1703246040	1394379800	1609837068	9
Aroclor-1268-1	(1)	212587803	222334836	235771254	243664628	256729280	234217560	7
Aroclor-1268-2	(2)	188526781	196925783	208389682	217661736	222594360	206819668	7
Aroclor-1268-3	(3)	155195263	161035703	174669460	176674312	179279020	169370752	6
Aroclor-1268-4	(4)	67370219	71019111	74855072	76562592	74977520	72956903	5
Aroclor-1268-5	(5)	431064337	442819956	457037218	461770552	492119380	456962289	5
Decachlorobiphenyl		1753973260	1820277787	1902401960	1966032280	2130248600	1914586777	8
Tetrachloro-m-xylene		1581658390	1604094987	1649636500	1662342440	1498230000	1599192463	4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Instrument ID: ECD_P Date(s) Analyzed: 05/19/2025 05/19/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.70	4.60	4.80	25133000
		2	4.79	4.69	4.89	17981400
		3	4.86	4.76	4.96	58509600
		4	0.00			0
		5	0.00			0
Aroclor-1262	500	1	8.09	7.99	8.19	133245000
		2	8.41	8.31	8.51	272386000
		3	8.72	8.62	8.82	185530000
		4	8.81	8.71	8.91	134281000
		5	9.46	9.36	9.56	92460600



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Instrument ID: ECD_P Date(s) Analyzed: 05/19/2025 05/19/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.01	3.91	4.11	23124400
		2	4.09	3.99	4.19	17698400
		3	4.17	4.07	4.27	52270400
		4	0.00			0
		5	0.00			0
Aroclor-1262	500	1	6.92	6.82	7.02	113028000
		2	7.18	7.08	7.28	97127000
		3	7.70	7.60	7.80	81531200
		4	7.77	7.67	7.87	140624000
		5	8.27	8.17	8.37	65083800

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072164.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 09:42
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 10:58:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.498	3.791	194.5E6	154.8E6	96.811	98.534
2) SA Decachlor...	10.214	8.822	153.8E6	92020898	96.351	91.411
Target Compounds						
3) L1 AR-1016-1	5.653	4.877	66878491	54525663	936.093	955.267
4) L1 AR-1016-2	5.675	4.895	99468528	79772717	947.379	966.582
5) L1 AR-1016-3	5.737	5.072	61412187	42641318	951.196	947.882
6) L1 AR-1016-4	5.835	5.114	50363803	33654246	953.177	941.833
7) L1 AR-1016-5	6.128	5.329	45437165	43791070	953.043	942.793
31) L7 AR-1260-1	7.249	6.364	85553757	72244144	953.187	946.735
32) L7 AR-1260-2	7.503	6.553	135.7E6	87664004	971.503	941.843
33) L7 AR-1260-3	7.861	6.705	109.0E6	79909091	963.139	946.529
34) L7 AR-1260-4	8.085	7.177	102.1E6	64349922	958.508	939.431
35) L7 AR-1260-5	8.404	7.419	228.9E6	161.3E6	974.541	951.851



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072164.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 09:42
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 10:58:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072165.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 09:59
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 11:01:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.789	152.0E6	118.8E6	75.442	75.431
2) SA Decachlor...	10.213	8.822	121.3E6	78372372	75.657	76.878
Target Compounds						
3) L1 AR-1016-1	5.652	4.876	52371638	44137111	738.609	765.351
4) L1 AR-1016-2	5.673	4.894	78350662	64615816	747.492	771.637
5) L1 AR-1016-3	5.736	5.071	48176186	34764296	747.454	765.036
6) L1 AR-1016-4	5.833	5.114	40053007	27566181	755.339	764.168
7) L1 AR-1016-5	6.127	5.328	35956306	35808108	752.783	763.822
31) L7 AR-1260-1	7.247	6.363	67759813	56480819	753.285	743.413
32) L7 AR-1260-2	7.500	6.552	103.3E6	67774056	743.256	735.291
33) L7 AR-1260-3	7.860	6.705	84615905	60994772	748.257	731.431
34) L7 AR-1260-4	8.084	7.176	80330794	50763991	752.749	744.038
35) L7 AR-1260-5	8.404	7.418	175.3E6	125.2E6	747.551	742.474



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 09:59
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 11:01:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072166.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 10:15
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 10:55:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.791	103.6E6	79685211	50.000	50.000
2) SA Decachlor...	10.213	8.823	82738232	54656808	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.652	4.877	38005022	29816144	500.000	500.000
4) L1 AR-1016-2	5.674	4.895	55259111	42644379	500.000	500.000
5) L1 AR-1016-3	5.736	5.073	33857030	23665215	500.000	500.000
6) L1 AR-1016-4	5.834	5.115	27655933	18905601	500.000	500.000
7) L1 AR-1016-5	6.127	5.329	24957303	24552684	500.000	500.000
31) L7 AR-1260-1	7.247	6.364	46978579	40186653	500.000	500.000
32) L7 AR-1260-2	7.501	6.553	71804602	49245044	500.000	500.000
33) L7 AR-1260-3	7.860	6.706	58694311	44468768	500.000	500.000
34) L7 AR-1260-4	8.085	7.177	55470384	36323860	500.000	500.000
35) L7 AR-1260-5	8.403	7.419	120.4E6	88817890	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072166.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 10:15
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 10:55:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072167.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 10:32
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 11:04:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.792	53397667	40989034	26.111	25.758
2) SA Decachlor...	10.213	8.823	42804673	29089475	26.249	27.561
Target Compounds						
3) L1 AR-1016-1	5.655	4.878	20802992	16128858	281.189	271.618
4) L1 AR-1016-2	5.676	4.896	29644893	23113324	273.834	269.018
5) L1 AR-1016-3	5.739	5.073	18387723	12747298	275.562	272.213
6) L1 AR-1016-4	5.837	5.115	14555177	10380280	267.927	277.285
7) L1 AR-1016-5	6.130	5.330	12797446	13425333	263.209	276.324
31) L7 AR-1260-1	7.248	6.364	24515226	21211058	266.529	271.268
32) L7 AR-1260-2	7.504	6.553	37350124	26231122	263.765	275.072
33) L7 AR-1260-3	7.862	6.705	30309213	23123504	263.279	269.924
34) L7 AR-1260-4	8.086	7.177	28647958	19427779	263.586	275.187
35) L7 AR-1260-5	8.406	7.418	62484280	46879399	262.166	270.430



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072167.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 10:32
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 11:04:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072168.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 10:48
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 11:09:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 10:54:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.498	3.793	9370134	8133165	4.660	5.088
2) SA Decachlor...	10.212	8.821	8114063	5355591	4.981	5.059
Target Compounds						
3) L1 AR-1016-1	5.650	4.878	3974445	3481238	53.538m	56.670
4) L1 AR-1016-2	5.672	4.896	5096224	4763878	47.816m	54.265
5) L1 AR-1016-3	5.734	5.073	3045634	2547956	47.236m	53.467
6) L1 AR-1016-4	5.832	5.115	2452478	2024258	45.948m	53.206
7) L1 AR-1016-5	6.127	5.330	2503132	2739010	51.179	54.973
31) L7 AR-1260-1	7.247	6.364	5002782	4299958	53.452	53.915
32) L7 AR-1260-2	7.500	6.553	7551641	6283346	52.628	61.952
33) L7 AR-1260-3	7.859	6.705	5443256	4592239	47.802	52.844
34) L7 AR-1260-4	8.083	7.176	5088570	3903032	47.423	54.140
35) L7 AR-1260-5	8.403	7.418	11497836	8735677	48.583	50.314

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072168.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 10:48
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 11:09:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 10:54:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072169.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:05
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 11:34:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 11:34:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.794	100.3E6	78966679	50.000	50.000
2) SA Decachlor...	10.214	8.824	80876112	53101134	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.703	4.005	12566489	11562215	500.000	500.000
9) L2 AR-1221-2	4.788	4.091	8990719	8849205	500.000	500.000
10) L2 AR-1221-3	4.864	4.167	29254822	26135247	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072169.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:05
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 11:34:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 11:34:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072170.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:21
 Operator : YP\AJ
 Sample : AR1232ICC1000
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:10:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 12:05:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.795	188.6E6	150.4E6	95.724	94.284
2) SA Decachlor...	10.213	8.823	154.5E6	99106147	96.034	93.871
Target Compounds						
11) L3 AR-1232-1	4.863	4.168	44075653	37616932	957.349	907.448
12) L3 AR-1232-2	5.389	4.898	21717379	39358353	966.038	938.328
13) L3 AR-1232-3	5.675	5.074	46772523	20796080	948.567	932.463
14) L3 AR-1232-4	5.835	5.159	23238504	17897124	967.407	927.845
15) L3 AR-1232-5	5.925	5.331	15752983	19472370	978.586	927.105



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072170.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:21
Operator : YP\AJ
Sample : AR1232ICC1000
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:10:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 12:05:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:38
 Operator : YP\AJ
 Sample : AR1232ICC750
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:15:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 12:12:58 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.796	150.1E6	117.1E6	75.772	73.927
2) SA Decachlor...	10.215	8.824	120.1E6	80247330	74.743	75.669
Target Compounds						
11) L3 AR-1232-1	4.864	4.168	34422611	30444044	748.451	739.537
12) L3 AR-1232-2	5.390	4.898	17196522	31518395	759.894	750.945
13) L3 AR-1232-3	5.676	5.075	37611138	16760662	758.466	751.013
14) L3 AR-1232-4	5.836	5.159	18110035	14561422	752.603	753.267
15) L3 AR-1232-5	5.926	5.332	12352668	15928473	761.482	755.563



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:38
Operator : YP\AJ
Sample : AR1232ICC750
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:15:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 12:12:58 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072172.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 11:54
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:07:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 12:05:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.795	102.7E6	84298957	50.000	50.000
2) SA Decachlor...	10.214	8.823	83645074	56024151	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.864	4.168	24001461	22645075	500.000	500.000
12) L3 AR-1232-2	5.390	4.897	11622194	22266011	500.000	500.000
13) L3 AR-1232-3	5.676	5.075	25922328	11904276	500.000	500.000
14) L3 AR-1232-4	5.836	5.159	12402195	10340360	500.000	500.000
15) L3 AR-1232-5	5.926	5.331	8221206	11267225	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072172.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 11:54
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:07:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 12:05:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072173.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 12:10
 Operator : YP\AJ
 Sample : AR1232ICC250
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 12:57:16 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 12:57:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.795	53057253	41777414	26.318	26.023
2) SA Decachlor...	10.217	8.824	42592543	28685378	26.120	26.506
Target Compounds						
11) L3 AR-1232-1	4.865	4.167	12061819	11916931	259.084	278.487
12) L3 AR-1232-2	5.391	4.897	5757452	11850655	253.297	273.502
13) L3 AR-1232-3	5.678	5.074	13595730	6293241	267.701	273.247
14) L3 AR-1232-4	5.838	5.158	6545799	5422323	266.163	272.197
15) L3 AR-1232-5	5.928	5.331	4527330	6054853	271.199	276.907



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072173.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 12:10
Operator : YP\AJ
Sample : AR1232ICC250
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 12:57:16 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 12:57:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072174.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 12:27
 Operator : YP\AJ
 Sample : AR1232ICC050
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/21/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:01:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:01:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.499	3.796	9615863	7891304	4.814	4.932
2) SA Decachlor...	10.211	8.824	7499330	5759848	4.674	5.254
Target Compounds						
11) L3 AR-1232-1	4.863	4.168	2465980	2277793	52.347	52.551
12) L3 AR-1232-2	5.389	4.898	1274783	2379299	54.751	53.854
13) L3 AR-1232-3	5.674	5.075	2358363	1239517	46.317m	53.009
14) L3 AR-1232-4	5.832	5.160	1356122	1195921	55.264m	57.718
15) L3 AR-1232-5	5.924	5.331	1041139	1415397	59.427	61.129

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072174.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 12:27
Operator : YP\AJ
Sample : AR1232ICC050
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1232ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/21/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:01:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:01:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072175.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 12:43
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:42:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:38:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.797	182.8E6	154.6E6	95.720	98.241
2) SA Decachlor...	10.216	8.825	152.6E6	96717168	96.225	96.077
Target Compounds						
16) L4 AR-1242-1	5.656	4.881	56211553	46123670	950.363	932.105
17) L4 AR-1242-2	5.678	4.898	82973080	66003789	949.042	933.080
18) L4 AR-1242-3	5.740	5.076	51559127	34394227	963.224	911.604
19) L4 AR-1242-4	5.837	5.160	41207603	32146418	945.604	905.879
20) L4 AR-1242-5	6.568	5.683	47977894	41776015	970.648	922.915



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072175.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 12:43
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:42:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:38:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072176.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 12:59
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:45:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:38:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	144.7E6	125.3E6	75.506	78.026
2) SA Decachlor...	10.215	8.825	117.5E6	80603094	74.400	78.305
Target Compounds						
16) L4 AR-1242-1	5.656	4.881	42992304	37426857	734.417	754.223
17) L4 AR-1242-2	5.677	4.899	64478622	53279521	741.623	752.130
18) L4 AR-1242-3	5.739	5.076	39091479	29082900	736.754	763.759
19) L4 AR-1242-4	5.837	5.161	32272993	27640589	743.693	769.026
20) L4 AR-1242-5	6.568	5.684	36522171	34960050	742.554	764.745



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072176.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 12:59
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:45:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:38:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072177.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 13:15
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:39:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:38:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.505	3.798	99563757	80047805	50.000	50.000
2) SA Decachlor...	10.221	8.827	82298882	52308236	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.659	4.883	31041680	26421500	500.000	500.000
17) L4 AR-1242-2	5.681	4.901	45941667	37735643	500.000	500.000
18) L4 AR-1242-3	5.743	5.078	27748086	20532241	500.000	500.000
19) L4 AR-1242-4	5.841	5.162	22974288	19413228	500.000	500.000
20) L4 AR-1242-5	6.572	5.685	25439777	24377285	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072177.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 13:15
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:39:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:38:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072178.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 13:32
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 13:48:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 13:38:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.797	52867260	40596927	26.895	25.212
2) SA Decachlor...	10.214	8.825	43224831	29078092	26.730	27.360
Target Compounds						
16) L4 AR-1242-1	5.654	4.881	16460967	14309390	272.688	277.708
17) L4 AR-1242-2	5.675	4.899	24396194	20235141	272.269	275.819
18) L4 AR-1242-3	5.738	5.076	14270237	11178619	263.948	281.311
19) L4 AR-1242-4	5.836	5.161	11695459	10798164	264.351	286.007
20) L4 AR-1242-5	6.566	5.684	13108900	13556747	262.192	283.361



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072178.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 13:32
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 13:48:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 13:38:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072179.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 13:48
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 14:11:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:10:51 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	8565148	6886728	4.472	4.404
2) SA Decachlor...	10.216	8.824	7437092	5391472	4.674	5.058
Target Compounds						
16) L4 AR-1242-1	5.654	4.882	3576242	2721766	56.040m	52.233
17) L4 AR-1242-2	5.677	4.900	3774327	3701307	42.861m	50.361
18) L4 AR-1242-3	5.740	5.076	2953563	1833864	53.891m	46.871
19) L4 AR-1242-4	5.836	5.161	2238823	1844717	51.018m	49.084
20) L4 AR-1242-5	6.568	5.683	2746740	2495867	53.225m	52.475m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072179.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 13:48
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 14:11:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:10:51 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 14:04
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 14:59:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:54:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.506	3.798	188.0E6	158.5E6	97.146	96.913
2) SA Decachlor...	10.219	8.825	157.3E6	107.8E6	97.225	97.659
Target Compounds						
21) L5 AR-1248-1	5.659	4.881	44038397	38942472	957.866	950.706
22) L5 AR-1248-2	5.931	5.118	56984732	50685144	964.616	951.833
23) L5 AR-1248-3	6.134	5.161	64677132	53612101	969.345	958.095
24) L5 AR-1248-4	6.533	5.333	81606824	63542852	959.973	960.283
25) L5 AR-1248-5	6.572	5.725	77587697	63751699	965.027	958.723



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 14:04
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 14:59:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:54:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072181.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 14:20
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 15:02:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:54:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.799	143.2E6	115.8E6	74.323	72.143
2) SA Decachlor...	10.214	8.826	119.9E6	80014834	74.410	73.309
Target Compounds						
21) L5 AR-1248-1	5.655	4.882	35604324	28560624	766.105	713.991
22) L5 AR-1248-2	5.927	5.119	44550644	37039968	752.752	712.825
23) L5 AR-1248-3	6.130	5.161	49360311	38952854	743.159	713.200
24) L5 AR-1248-4	6.529	5.333	63978321	45811338	751.733	710.534
25) L5 AR-1248-5	6.568	5.726	60752796	46780519	753.748	718.348



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072181.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 14:20
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 15:02:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:54:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072182.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 14:37
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 14:56:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:54:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.798	99535493	84317976	50.000	50.000
2) SA Decachlor...	10.215	8.825	83159844	56480842	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.656	4.881	23956323	21490405	500.000	500.000
22) L5 AR-1248-2	5.928	5.119	30582698	27907479	500.000	500.000
23) L5 AR-1248-3	6.131	5.161	34383966	29150937	500.000	500.000
24) L5 AR-1248-4	6.530	5.333	44206090	34399526	500.000	500.000
25) L5 AR-1248-5	6.569	5.726	41605620	34620634	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072182.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 14:37
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 14:56:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:54:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072183.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 14:53
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 15:05:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 14:54:54 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.798	52810864	45199511	26.764	27.295
2) SA Decachlor...	10.217	8.825	42874232	30462979	26.180	27.121
Target Compounds						
21) L5 AR-1248-1	5.657	4.882	12861341	11684235	269.533	280.297
22) L5 AR-1248-2	5.929	5.119	15869170	15269791	263.358	281.515
23) L5 AR-1248-3	6.132	5.161	17659503	16001810	261.722	280.908
24) L5 AR-1248-4	6.531	5.333	23049150	18905273	265.299	281.073
25) L5 AR-1248-5	6.570	5.726	21712961	18625278	264.265	276.065



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072183.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 14:53
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 15:05:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 14:54:54 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 15:25
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 15:39:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 15:20:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.500	3.797	9281743	8589975	4.760	5.149
2) SA Decachlor...	10.214	8.824	7393621	5937693	4.604	5.226
Target Compounds						
21) L5 AR-1248-1	5.652	4.881	2555425	2493351	52.477m	57.554
22) L5 AR-1248-2	5.926	5.118	3108634	3271038	51.264	57.918
23) L5 AR-1248-3	6.128	5.160	3050186	3383850	46.089	57.249
24) L5 AR-1248-4	6.528	5.332	4472018	3960972	51.172	56.978m
25) L5 AR-1248-5	6.567	5.725	4323703	3941459	52.077	56.517

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 15:25
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 15:39:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 15:20:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072185.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 15:42
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 17:56:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:50:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.799	186.9E6	151.6E6	95.813	92.566
2) SA Decachlor...	10.217	8.825	157.7E6	104.9E6	96.583	97.142
Target Compounds						
26) L6 AR-1254-1	6.507	5.685	79792123	88394708	957.230	921.511
27) L6 AR-1254-2	6.724	5.834	120.1E6	75158218	956.365	909.300
28) L6 AR-1254-3	7.088	6.236	124.3E6	118.6E6	963.286	925.888
29) L6 AR-1254-4	7.370	6.465	116.3E6	76184182	966.294	916.589
30) L6 AR-1254-5	7.787	6.883	107.9E6	103.3E6	962.647	923.217



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072185.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 15:42
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 17:56:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:50:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072186.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 15:58
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 18:00:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:59:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.799	146.2E6	125.8E6	74.952	76.204
2) SA Decachlor...	10.217	8.825	121.1E6	81081085	74.439	75.075
Target Compounds						
26) L6 AR-1254-1	6.507	5.685	62087463	69101121	746.549	729.987
27) L6 AR-1254-2	6.723	5.834	93637557	59106174	746.914	726.363
28) L6 AR-1254-3	7.087	6.237	95874547	92318270	745.288	730.264
29) L6 AR-1254-4	7.370	6.465	88913109	59863409	742.441	729.887
30) L6 AR-1254-5	7.786	6.883	84042453	83723123	749.686	748.732



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072186.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 15:58
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 18:00:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:59:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072187.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 16:14
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 17:52:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:50:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.798	101.6E6	87948942	50.000	50.000
2) SA Decachlor...	10.215	8.825	84401769	55516052	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.506	5.685	43461212	51726314	500.000	500.000
27) L6 AR-1254-2	6.723	5.833	65552662	45075919	500.000	500.000
28) L6 AR-1254-3	7.086	6.236	66891427	68786744	500.000	500.000
29) L6 AR-1254-4	7.369	6.465	62209110	45024944	500.000	500.000
30) L6 AR-1254-5	7.785	6.883	58157609	60253702	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072187.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 16:14
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 17:52:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:50:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072188.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 16:31
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 18:03:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:59:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.798	51747510	42581151	26.134	25.594
2) SA Decachlor...	10.215	8.825	42700529	30290967	25.932	27.218
Target Compounds						
26) L6 AR-1254-1	6.506	5.685	22921007	27395223	268.725	278.433
27) L6 AR-1254-2	6.722	5.833	33636773	23446738	263.485	277.554
28) L6 AR-1254-3	7.085	6.237	34246078	35238769	261.967	270.959
29) L6 AR-1254-4	7.369	6.465	34017428	23090909	274.698	272.930
30) L6 AR-1254-5	7.785	6.883	29284530	30214333	258.327	264.854



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072188.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 16:31
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 18:03:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:59:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 16:47
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 18:06:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 19 17:59:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	8087682	6971899	4.231m	4.331m
2) SA Decachlor...	10.218	8.826	7375422	5691115	4.574	5.091
Target Compounds						
26) L6 AR-1254-1	6.507	5.686	4441722	5310413	51.935m	53.128
27) L6 AR-1254-2	6.724	5.834	7137094	4681467	54.616	54.242
28) L6 AR-1254-3	7.088	6.237	6855410	6144757	51.934	47.774
29) L6 AR-1254-4	7.370	6.466	7860526	4403320	60.049m	52.100m
30) L6 AR-1254-5	7.788	6.883	5139236	5522061	46.197	48.716

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 16:47
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 18:06:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 19 17:59:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072190.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 17:03
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 02:51:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 02:50:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.798	91392014	72543788	50.000	50.000
2) SA Decachlor...	10.218	8.826	79498650	53689400	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.088	6.920	66622555	56514187	500.000	500.000
37) L8 AR-1262-2	8.408	7.179	136.2E6	48563462	500.000	500.000
38) L8 AR-1262-3	8.724	7.703	92764853	40765600	500.000	500.000
39) L8 AR-1262-4	8.809	7.767	67140472	70312088	500.000	500.000
40) L8 AR-1262-5	9.464	8.267	46230270	32541912	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072190.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 17:03
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 02:51:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 02:50:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072191.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 17:20
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:11:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	185.0E6	158.2E6	92.680	95.879
2) SA Decachlor...	10.215	8.825	273.4E6	175.4E6	93.047	92.198
Target Compounds						
41) L9 AR-1268-1	8.718	7.702	315.7E6	212.6E6	935.036	901.670
42) L9 AR-1268-2	8.810	7.768	265.9E6	188.5E6	932.233	904.684
43) L9 AR-1268-3	9.041	7.971	228.0E6	155.2E6	921.986	888.508
44) L9 AR-1268-4	9.460	8.265	101.4E6	67370219	939.092	900.009
45) L9 AR-1268-5	9.876	8.565	651.5E6	431.1E6	948.441	943.171



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072191.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 17:20
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:11:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072192.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 17:36
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:11:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.506	3.799	142.5E6	120.3E6	71.406	72.929
2) SA Decachlor...	10.220	8.826	211.2E6	136.5E6	71.876	71.762
Target Compounds						
41) L9 AR-1268-1	8.722	7.704	242.2E6	166.8E6	717.436	707.258
42) L9 AR-1268-2	8.815	7.770	204.3E6	147.7E6	716.226	708.741
43) L9 AR-1268-3	9.047	7.973	176.3E6	120.8E6	713.088	691.459
44) L9 AR-1268-4	9.464	8.267	78574325	53264333	727.753	711.566
45) L9 AR-1268-5	9.880	8.568	500.4E6	332.1E6	728.402	726.669



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072192.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 17:36
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:11:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072193.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 17:53
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:11:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.799	99784961	82481825	50.000	50.000
2) SA Decachlor...	10.215	8.826	146.9E6	95120098	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.719	7.703	168.8E6	117.9E6	500.000	500.000
42) L9 AR-1268-2	8.812	7.769	142.6E6	104.2E6	500.000	500.000
43) L9 AR-1268-3	9.042	7.972	123.6E6	87334730	500.000	500.000
44) L9 AR-1268-4	9.461	8.267	53984174	37427536	500.000	500.000
45) L9 AR-1268-5	9.876	8.566	343.5E6	228.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072193.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 17:53
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:11:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072194.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 18:09
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:12:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	51365767	41558561	25.738	25.193
2) SA Decachlor...	10.214	8.824	75968459	49150807	25.857	25.836
Target Compounds						
41) L9 AR-1268-1	8.717	7.702	87839198	60916157	260.167	258.370
42) L9 AR-1268-2	8.811	7.767	74121995	54415434	259.881	261.123
43) L9 AR-1268-3	9.040	7.971	63936897	44168578	258.592	252.869
44) L9 AR-1268-4	9.459	8.265	27090074	19140648	250.908	255.703
45) L9 AR-1268-5	9.875	8.565	176.6E6	115.4E6	257.058	252.589



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072194.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 18:09
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:12:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072195.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 18:25
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/20/2025
 Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:12:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:10:31 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	8721682	7491150	4.370m	4.541
2) SA Decachlor...	10.218	8.823	13183771	10651243	4.487	5.599
Target Compounds						
41) L9 AR-1268-1	8.720	7.701	16905769	12836464	50.072m	54.445
42) L9 AR-1268-2	8.814	7.768	13998978	11129718	49.082	53.408
43) L9 AR-1268-3	9.045	7.971	12064238	8963951	48.794	51.320
44) L9 AR-1268-4	9.464	8.265	5112737	3748876	47.354	50.082
45) L9 AR-1268-5	9.878	8.565	33648372	24605969	48.982	53.838

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072195.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 18:25
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/20/2025
Supervised By :mohammad ahmed 05/22/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:12:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:10:31 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 18:42
 Operator : YP\AJ
 Sample : PP051925ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:36:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.798	95813360	75129300	47.649	47.004
2) SA Decachlor...	10.216	8.825	78897811	55159923	48.430	52.106
Target Compounds						
3) L1 AR-1016-1	5.656	4.882	34275021	28415442	456.491	462.570
4) L1 AR-1016-2	5.677	4.900	52043054	40409419	486.422	460.300
5) L1 AR-1016-3	5.739	5.077	30972988	22233290	472.402	466.552
6) L1 AR-1016-4	5.837	5.120	26002231	17703522	488.122	465.327
7) L1 AR-1016-5	6.130	5.333	22892152	23633981	468.054	474.345
31) L7 AR-1260-1	7.250	6.367	45618683	37214670	487.406	466.620
32) L7 AR-1260-2	7.503	6.556	68966981	45298985	480.641	446.638
33) L7 AR-1260-3	7.861	6.709	56600304	41944639	497.057	482.664
34) L7 AR-1260-4	8.086	7.179	54236593	34998898	505.455	485.483
35) L7 AR-1260-5	8.405	7.421	116.3E6	86272130	491.560	496.891



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 18:42
Operator : YP\AJ
Sample : PP051925ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:36:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072197.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 19:14
 Operator : YP\AJ
 Sample : AR1232ICV500
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1232

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:37:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.798	99088078	76822954	49.278	48.064
2) SA Decachlor...	10.217	8.824	80208843	53006313	49.234	50.072
Target Compounds						
11) L3 AR-1232-1	4.867	4.169	22860124	20991950	485.266	484.304
12) L3 AR-1232-2	5.392	4.899	11426080	20405975	490.744	461.875
13) L3 AR-1232-3	5.679	5.076	24392843	11207066	487.242	479.281
14) L3 AR-1232-4	5.839	5.159	11984156	9670115	477.474	466.701
15) L3 AR-1232-5	5.929	5.332	8016086	10482162	457.551	452.707



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072197.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 19:14
Operator : YP\AJ
Sample : AR1232ICV500
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1232

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:37:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 19:47
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:37:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.798	96787545	76508111	48.134	47.867
2) SA Decachlor...	10.213	8.825	81177144	55473433	49.829	52.403
Target Compounds						
16) L4 AR-1242-1	5.655	4.882	30009936	25988486	479.412	498.739
17) L4 AR-1242-2	5.676	4.900	44873042	37823387	517.091	514.630
18) L4 AR-1242-3	5.739	5.077	26969959	20511520	489.776	524.249
19) L4 AR-1242-4	5.836	5.161	22146348	19605635	499.365	521.665
20) L4 AR-1242-5	6.567	5.684	24841123	25039100	487.226	518.863



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 19:47
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:37:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072199.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 20:19
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:38:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.798	99024800	84569869	49.246	52.910
2) SA Decachlor...	10.212	8.824	83034822	56507020	50.969	53.379
Target Compounds						
21) L5 AR-1248-1	5.656	4.881	23569496	21026666	487.019	485.362
22) L5 AR-1248-2	5.927	5.119	30714899	27827558	506.512	492.721
23) L5 AR-1248-3	6.130	5.161	34359439	29174304	519.181	493.584
24) L5 AR-1248-4	6.529	5.333	44131890	34133080	504.987	490.046
25) L5 AR-1248-5	6.568	5.725	42053839	34467498	506.516	494.232



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072199.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 20:19
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:38:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072200.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 20:52
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:39:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:29:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	100.3E6	87142783	49.887	54.520
2) SA Decachlor...	10.214	8.823	83987433	56189590	51.554	53.079
Target Compounds						
26) L6 AR-1254-1	6.506	5.684	43414677	50851971	504.803	508.752
27) L6 AR-1254-2	6.722	5.832	65079227	43668465	498.015	505.967
28) L6 AR-1254-3	7.085	6.235	66155023	67680378	501.163	526.201
29) L6 AR-1254-4	7.367	6.463	61672158	43977452	472.545	515.584
30) L6 AR-1254-5	7.784	6.881	58110470	59298415	522.357	523.137



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 20:52
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:39:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
 Data File : PP072201.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2025 21:25
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP051925AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 03:20:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 20 03:19:33 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.504	3.798	100.4E6	87828193	52.612	54.920
2) SA Decachlor...	10.216	8.824	147.5E6	102.0E6	52.061	53.249
Target Compounds						
41) L9 AR-1268-1	8.720	7.702	171.0E6	124.0E6	513.323	529.555
42) L9 AR-1268-2	8.812	7.768	144.0E6	108.9E6	514.169	526.775
43) L9 AR-1268-3	9.044	7.971	123.7E6	91750131	512.196	541.712
44) L9 AR-1268-4	9.463	8.266	54675516	40276953	520.976	552.065
45) L9 AR-1268-5	9.877	8.565	344.9E6	244.7E6	509.466	535.577



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072201.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 21:25
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP051925AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:20:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:19:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 05/30/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 10:21 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.65	5.55	5.75	-0.01
Aroclor-1016-2	(2)	5.68	5.67	5.57	5.77	-0.01
Aroclor-1016-3	(3)	5.74	5.74	5.64	5.84	0.00
Aroclor-1016-4	(4)	5.84	5.83	5.73	5.93	-0.01
Aroclor-1016-5	(5)	6.13	6.13	6.03	6.23	0.00
Aroclor-1260-1	(1)	7.25	7.25	7.15	7.35	0.00
Aroclor-1260-2	(2)	7.50	7.50	7.40	7.60	0.00
Aroclor-1260-3	(3)	7.86	7.86	7.76	7.96	0.00
Aroclor-1260-4	(4)	8.08	8.09	7.99	8.19	0.01
Aroclor-1260-5	(5)	8.40	8.40	8.30	8.50	0.00
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.21	10.21	10.11	10.31	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 05/30/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 10:21 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.12	5.02	5.22	0.01
Aroclor-1016-5	(5)	5.33	5.33	5.23	5.43	0.00
Aroclor-1260-1	(1)	6.36	6.36	6.26	6.46	0.00
Aroclor-1260-2	(2)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-3	(3)	6.70	6.71	6.61	6.81	0.01
Aroclor-1260-4	(4)	7.17	7.18	7.08	7.28	0.01
Aroclor-1260-5	(5)	7.42	7.42	7.32	7.52	0.01
Tetrachloro-m-xylene		3.80	3.79	3.69	3.89	-0.01
Decachlorobiphenyl		8.82	8.82	8.72	8.92	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072461.D Time Analyzed: 10:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.655	5.552	5.752	422.630	500.000	-15.5
Aroclor-1016-2	5.676	5.574	5.774	470.170	500.000	-6.0
Aroclor-1016-3	5.738	5.636	5.836	468.320	500.000	-6.3
Aroclor-1016-4	5.836	5.734	5.934	488.430	500.000	-2.3
Aroclor-1016-5	6.128	6.027	6.227	478.290	500.000	-4.3
Aroclor-1260-1	7.247	7.147	7.347	455.460	500.000	-8.9
Aroclor-1260-2	7.501	7.401	7.601	471.800	500.000	-5.6
Aroclor-1260-3	7.859	7.760	7.960	493.710	500.000	-1.3
Aroclor-1260-4	8.083	7.985	8.185	484.210	500.000	-3.2
Aroclor-1260-5	8.402	8.303	8.503	506.790	500.000	1.4
Decachlorobiphenyl	10.210	10.113	10.313	52.720	50.000	5.4
Tetrachloro-m-xylene	4.502	4.396	4.596	47.950	50.000	-4.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072461.D Time Analyzed: 10:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	498.390	500.000	-0.3
Aroclor-1016-2	4.896	4.795	4.995	499.930	500.000	0.0
Aroclor-1016-3	5.073	4.973	5.173	502.840	500.000	0.6
Aroclor-1016-4	5.114	5.015	5.215	506.060	500.000	1.2
Aroclor-1016-5	5.329	5.229	5.429	496.780	500.000	-0.6
Aroclor-1260-1	6.361	6.264	6.464	523.800	500.000	4.8
Aroclor-1260-2	6.550	6.453	6.653	499.500	500.000	-0.1
Aroclor-1260-3	6.702	6.606	6.806	530.710	500.000	6.1
Aroclor-1260-4	7.173	7.077	7.277	518.770	500.000	3.8
Aroclor-1260-5	7.415	7.319	7.519	517.690	500.000	3.5
Decachlorobiphenyl	8.816	8.723	8.923	55.760	50.000	11.5
Tetrachloro-m-xylene	3.796	3.691	3.891	49.680	50.000	-0.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072461.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 10:21
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 10:54:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.796	96412594	79404874	47.947	49.679
2) SA Decachlor...	10.210	8.816	85883109	59029368	52.717	55.762
Target Compounds						
3) L1 AR-1016-1	5.655	4.878	31732582	30615644	422.630	498.386
4) L1 AR-1016-2	5.676	4.896	50304118	43888127	470.169	499.925
5) L1 AR-1016-3	5.738	5.073	30705307	23962492	468.319	502.838
6) L1 AR-1016-4	5.836	5.114	26018804	19253134	488.433	506.058
7) L1 AR-1016-5	6.128	5.329	23392629	24751772	478.287	496.780
31) L7 AR-1260-1	7.247	6.361	42628401	41774855	455.457	523.798
32) L7 AR-1260-2	7.501	6.550	67698563	50660537	471.801	499.502
33) L7 AR-1260-3	7.859	6.702	56219658	46119898	493.714	530.709
34) L7 AR-1260-4	8.083	7.173	51957296	37398436	484.213	518.767
35) L7 AR-1260-5	8.402	7.415	119.9E6	89883010	506.790	517.688



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072461.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 10:21
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 10:54:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 05/30/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 16:15 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.65	5.55	5.75	-0.01
Aroclor-1016-2	(2)	5.68	5.67	5.57	5.77	-0.01
Aroclor-1016-3	(3)	5.74	5.74	5.64	5.84	0.00
Aroclor-1016-4	(4)	5.84	5.83	5.73	5.93	-0.01
Aroclor-1016-5	(5)	6.13	6.13	6.03	6.23	0.00
Aroclor-1260-1	(1)	7.25	7.25	7.15	7.35	0.00
Aroclor-1260-2	(2)	7.50	7.50	7.40	7.60	0.00
Aroclor-1260-3	(3)	7.86	7.86	7.76	7.96	0.00
Aroclor-1260-4	(4)	8.08	8.09	7.99	8.19	0.01
Aroclor-1260-5	(5)	8.40	8.40	8.30	8.50	0.00
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.21	10.21	10.11	10.31	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 05/30/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 16:15 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.12	5.12	5.02	5.22	0.01
Aroclor-1016-5	(5)	5.33	5.33	5.23	5.43	0.00
Aroclor-1260-1	(1)	6.36	6.36	6.26	6.46	0.00
Aroclor-1260-2	(2)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-3	(3)	6.70	6.71	6.61	6.81	0.01
Aroclor-1260-4	(4)	7.18	7.18	7.08	7.28	0.00
Aroclor-1260-5	(5)	7.42	7.42	7.32	7.52	0.00
Tetrachloro-m-xylene		3.80	3.79	3.69	3.89	-0.01
Decachlorobiphenyl		8.82	8.82	8.72	8.92	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072476.D Time Analyzed: 16:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.655	5.552	5.752	416.340	500.000	-16.7
Aroclor-1016-2	5.677	5.574	5.774	449.250	500.000	-10.2
Aroclor-1016-3	5.739	5.636	5.836	436.950	500.000	-12.6
Aroclor-1016-4	5.837	5.734	5.934	456.260	500.000	-8.7
Aroclor-1016-5	6.129	6.027	6.227	446.620	500.000	-10.7
Aroclor-1260-1	7.248	7.147	7.347	438.650	500.000	-12.3
Aroclor-1260-2	7.501	7.401	7.601	447.180	500.000	-10.6
Aroclor-1260-3	7.860	7.760	7.960	462.100	500.000	-7.6
Aroclor-1260-4	8.084	7.985	8.185	456.590	500.000	-8.7
Aroclor-1260-5	8.403	8.303	8.503	467.390	500.000	-6.5
Decachlorobiphenyl	10.211	10.113	10.313	48.580	50.000	-2.8
Tetrachloro-m-xylene	4.503	4.396	4.596	46.790	50.000	-6.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072476.D Time Analyzed: 16:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	482.170	500.000	-3.6
Aroclor-1016-2	4.896	4.795	4.995	495.180	500.000	-1.0
Aroclor-1016-3	5.074	4.973	5.173	495.170	500.000	-1.0
Aroclor-1016-4	5.115	5.015	5.215	501.040	500.000	0.2
Aroclor-1016-5	5.330	5.229	5.429	493.810	500.000	-1.2
Aroclor-1260-1	6.362	6.264	6.464	516.350	500.000	3.3
Aroclor-1260-2	6.551	6.453	6.653	491.560	500.000	-1.7
Aroclor-1260-3	6.704	6.606	6.806	527.110	500.000	5.4
Aroclor-1260-4	7.175	7.077	7.277	515.270	500.000	3.1
Aroclor-1260-5	7.416	7.319	7.519	509.650	500.000	1.9
Decachlorobiphenyl	8.819	8.723	8.923	51.310	50.000	2.6
Tetrachloro-m-xylene	3.796	3.691	3.891	50.540	50.000	1.1

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072476.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 16:15
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 16:55:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.796	94075645	80780012	46.785	50.539
2) SA Decachlor...	10.211	8.819	79138448	54314812	48.577	51.308
Target Compounds						
3) L1 AR-1016-1	5.655	4.878	31260640	29619561	416.344	482.171
4) L1 AR-1016-2	5.677	4.896	48065533	43471682	449.246	495.182
5) L1 AR-1016-3	5.739	5.074	28648757	23597245	436.952	495.174
6) L1 AR-1016-4	5.837	5.115	24305064	19062099	456.262	501.037
7) L1 AR-1016-5	6.129	5.330	21843738	24604001	446.618	493.814
31) L7 AR-1260-1	7.248	6.362	41055446	41180890	438.651	516.351
32) L7 AR-1260-2	7.501	6.551	64165978	49854892	447.182	491.558
33) L7 AR-1260-3	7.860	6.704	52619234	45807371	462.096	527.113
34) L7 AR-1260-4	8.084	7.175	48992828	37146046	456.586	515.266
35) L7 AR-1260-5	8.403	7.416	110.6E6	88486794	467.387	509.646



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072476.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 16:15
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 16:55:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 05/30/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 21:09 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.65	5.55	5.75	-0.01
Aroclor-1016-2	(2)	5.68	5.67	5.57	5.77	-0.01
Aroclor-1016-3	(3)	5.74	5.74	5.64	5.84	0.00
Aroclor-1016-4	(4)	5.84	5.83	5.73	5.93	-0.01
Aroclor-1016-5	(5)	6.13	6.13	6.03	6.23	0.00
Aroclor-1260-1	(1)	7.25	7.25	7.15	7.35	0.00
Aroclor-1260-2	(2)	7.50	7.50	7.40	7.60	0.00
Aroclor-1260-3	(3)	7.86	7.86	7.76	7.96	0.00
Aroclor-1260-4	(4)	8.08	8.09	7.99	8.19	0.01
Aroclor-1260-5	(5)	8.40	8.40	8.30	8.50	0.00
Tetrachloro-m-xylene		4.50	4.50	4.40	4.60	0.00
Decachlorobiphenyl		10.21	10.21	10.11	10.31	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 05/30/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 21:09 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.88	4.88	4.78	4.98	0.00
Aroclor-1016-2	(2)	4.90	4.90	4.80	5.00	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.12	5.12	5.02	5.22	0.00
Aroclor-1016-5	(5)	5.33	5.33	5.23	5.43	0.00
Aroclor-1260-1	(1)	6.36	6.36	6.26	6.46	0.00
Aroclor-1260-2	(2)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-3	(3)	6.71	6.71	6.61	6.81	0.01
Aroclor-1260-4	(4)	7.18	7.18	7.08	7.28	0.00
Aroclor-1260-5	(5)	7.42	7.42	7.32	7.52	0.00
Tetrachloro-m-xylene		3.80	3.79	3.69	3.89	-0.01
Decachlorobiphenyl		8.82	8.82	8.72	8.92	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072491.D Time Analyzed: 21:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.656	5.552	5.752	414.050	500.000	-17.2
Aroclor-1016-2	5.677	5.574	5.774	443.170	500.000	-11.4
Aroclor-1016-3	5.739	5.636	5.836	435.990	500.000	-12.8
Aroclor-1016-4	5.837	5.734	5.934	447.390	500.000	-10.5
Aroclor-1016-5	6.129	6.027	6.227	438.220	500.000	-12.4
Aroclor-1260-1	7.248	7.147	7.347	426.950	500.000	-14.6
Aroclor-1260-2	7.501	7.401	7.601	419.840	500.000	-16.0
Aroclor-1260-3	7.860	7.760	7.960	435.350	500.000	-12.9
Aroclor-1260-4	8.084	7.985	8.185	424.930	500.000	-15.0
Aroclor-1260-5	8.403	8.303	8.503	444.970	500.000	-11.0
Decachlorobiphenyl	10.211	10.113	10.313	46.890	50.000	-6.2
Tetrachloro-m-xylene	4.503	4.396	4.596	46.710	50.000	-6.6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072491.D Time Analyzed: 21:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.880	4.777	4.977	500.510	500.000	0.1
Aroclor-1016-2	4.897	4.795	4.995	505.890	500.000	1.2
Aroclor-1016-3	5.074	4.973	5.173	512.590	500.000	2.5
Aroclor-1016-4	5.116	5.015	5.215	515.110	500.000	3.0
Aroclor-1016-5	5.330	5.229	5.429	506.240	500.000	1.2
Aroclor-1260-1	6.363	6.264	6.464	499.960	500.000	0.0
Aroclor-1260-2	6.552	6.453	6.653	476.940	500.000	-4.6
Aroclor-1260-3	6.705	6.606	6.806	508.980	500.000	1.8
Aroclor-1260-4	7.175	7.077	7.277	507.230	500.000	1.4
Aroclor-1260-5	7.417	7.319	7.519	506.090	500.000	1.2
Decachlorobiphenyl	8.819	8.723	8.923	52.090	50.000	4.2
Tetrachloro-m-xylene	3.797	3.691	3.891	50.530	50.000	1.1

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072491.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 21:09
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 22:57:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	93917012	80769793	46.706	50.533
2) SA Decachlor...	10.211	8.819	76396638	55141600	46.894	52.089
Target Compounds						
3) L1 AR-1016-1	5.656	4.880	31088760	30745968	414.055	500.508
4) L1 AR-1016-2	5.677	4.897	47415520	44411990	443.170	505.893
5) L1 AR-1016-3	5.739	5.074	28585605	24427445	435.989	512.595
6) L1 AR-1016-4	5.837	5.116	23832598	19597437	447.393	515.108
7) L1 AR-1016-5	6.129	5.330	21433055	25223147	438.221	506.241
31) L7 AR-1260-1	7.248	6.363	39960057	39873449	426.948	499.957
32) L7 AR-1260-2	7.501	6.552	60242657	48372108	419.840	476.938
33) L7 AR-1260-3	7.860	6.705	49574162	44231499	435.354	508.979
34) L7 AR-1260-4	8.084	7.175	45595993	36566661	424.929	507.230
35) L7 AR-1260-5	8.403	7.417	105.3E6	87870123	444.974	506.094



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072491.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 21:09
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 22:57:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Analytical Sequence

Client: CDM Smith

SDG No.: Q2150

Project: South River WM Replacement

Instrument ID: ECD_P

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 05/19/2025

05/19/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	05/19/2025	09:26	PP072163.D	10.21	4.50
AR1660ICC1000	AR1660ICC1000	05/19/2025	09:42	PP072164.D	10.21	4.50
AR1660ICC750	AR1660ICC750	05/19/2025	09:59	PP072165.D	10.21	4.50
AR1660ICC500	AR1660ICC500	05/19/2025	10:15	PP072166.D	10.21	4.50
AR1660ICC250	AR1660ICC250	05/19/2025	10:32	PP072167.D	10.21	4.50
AR1660ICC050	AR1660ICC050	05/19/2025	10:48	PP072168.D	10.21	4.50
AR1221ICC500	AR1221ICC500	05/19/2025	11:05	PP072169.D	10.21	4.50
AR1232ICC1000	AR1232ICC1000	05/19/2025	11:21	PP072170.D	10.21	4.50
AR1232ICC750	AR1232ICC750	05/19/2025	11:38	PP072171.D	10.22	4.50
AR1232ICC500	AR1232ICC500	05/19/2025	11:54	PP072172.D	10.21	4.50
AR1232ICC250	AR1232ICC250	05/19/2025	12:10	PP072173.D	10.22	4.50
AR1232ICC050	AR1232ICC050	05/19/2025	12:27	PP072174.D	10.21	4.50
AR1242ICC1000	AR1242ICC1000	05/19/2025	12:43	PP072175.D	10.22	4.50
AR1242ICC750	AR1242ICC750	05/19/2025	12:59	PP072176.D	10.22	4.50
AR1242ICC500	AR1242ICC500	05/19/2025	13:15	PP072177.D	10.22	4.51
AR1242ICC250	AR1242ICC250	05/19/2025	13:32	PP072178.D	10.21	4.50
AR1242ICC050	AR1242ICC050	05/19/2025	13:48	PP072179.D	10.22	4.50
AR1248ICC1000	AR1248ICC1000	05/19/2025	14:04	PP072180.D	10.22	4.51
AR1248ICC750	AR1248ICC750	05/19/2025	14:20	PP072181.D	10.21	4.50
AR1248ICC500	AR1248ICC500	05/19/2025	14:37	PP072182.D	10.22	4.50
AR1248ICC250	AR1248ICC250	05/19/2025	14:53	PP072183.D	10.22	4.50
AR1248ICC050	AR1248ICC050	05/19/2025	15:25	PP072184.D	10.21	4.50
AR1254ICC1000	AR1254ICC1000	05/19/2025	15:42	PP072185.D	10.22	4.50
AR1254ICC750	AR1254ICC750	05/19/2025	15:58	PP072186.D	10.22	4.50
AR1254ICC500	AR1254ICC500	05/19/2025	16:14	PP072187.D	10.22	4.50
AR1254ICC250	AR1254ICC250	05/19/2025	16:31	PP072188.D	10.22	4.50
AR1254ICC050	AR1254ICC050	05/19/2025	16:47	PP072189.D	10.22	4.50
AR1262ICC500	AR1262ICC500	05/19/2025	17:03	PP072190.D	10.22	4.50
AR1268ICC1000	AR1268ICC1000	05/19/2025	17:20	PP072191.D	10.22	4.50
AR1268ICC750	AR1268ICC750	05/19/2025	17:36	PP072192.D	10.22	4.51
AR1268ICC500	AR1268ICC500	05/19/2025	17:53	PP072193.D	10.22	4.50
AR1268ICC250	AR1268ICC250	05/19/2025	18:09	PP072194.D	10.21	4.50
AR1268ICC050	AR1268ICC050	05/19/2025	18:25	PP072195.D	10.22	4.50
AR1660CCC500	AR1660CCC500	05/30/2025	10:21	PP072461.D	10.21	4.50
IBLK	IBLK	05/30/2025	11:29	PP072465.D	10.21	4.50
PB168208BL	PB168208BL	05/30/2025	12:59	PP072466.D	10.21	4.51
PB168208BS	PB168208BS	05/30/2025	13:15	PP072467.D	10.21	4.50
TP-44	Q2150-01	05/30/2025	13:32	PP072468.D	10.21	4.50
TP-42	Q2150-02	05/30/2025	13:48	PP072469.D	10.21	4.50
TP-42MS	Q2150-02MS	05/30/2025	14:04	PP072470.D	10.21	4.50
TP-42MSD	Q2150-02MSD	05/30/2025	14:20	PP072471.D	10.21	4.50
TP-39	Q2150-03	05/30/2025	14:37	PP072472.D	10.21	4.50

Analytical Sequence

TP-48	Q2150-04	05/30/2025	14:53	PP072473.D	10.21	4.50
TP-47	Q2150-05	05/30/2025	15:10	PP072474.D	10.21	4.50
TP-50	Q2150-06	05/30/2025	15:26	PP072475.D	10.21	4.50
AR1660CCC500	AR1660CCC500	05/30/2025	16:15	PP072476.D	10.21	4.50
LBLK	LBLK	05/30/2025	17:20	PP072480.D	10.21	4.50
TP-51	Q2150-07	05/30/2025	17:37	PP072481.D	10.21	4.50
TP-52	Q2150-08	05/30/2025	17:53	PP072482.D	10.21	4.50
TP-54	Q2150-09	05/30/2025	18:09	PP072483.D	10.21	4.50
TP-53	Q2150-10	05/30/2025	18:26	PP072484.D	10.21	4.50
AR1660CCC500	AR1660CCC500	05/30/2025	21:09	PP072491.D	10.21	4.50
LBLK	LBLK	05/30/2025	22:31	PP072495.D	10.21	4.50

Analytical Sequence

Client: CDM Smith

SDG No.: Q2150

Project: South River WM Replacement

Instrument ID: ECD_P

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 05/19/2025

05/19/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
LBLK	LBLK	05/19/2025	09:26	PP072163.D	8.82	3.79
AR1660ICC1000	AR1660ICC1000	05/19/2025	09:42	PP072164.D	8.82	3.79
AR1660ICC750	AR1660ICC750	05/19/2025	09:59	PP072165.D	8.82	3.79
AR1660ICC500	AR1660ICC500	05/19/2025	10:15	PP072166.D	8.82	3.79
AR1660ICC250	AR1660ICC250	05/19/2025	10:32	PP072167.D	8.82	3.79
AR1660ICC050	AR1660ICC050	05/19/2025	10:48	PP072168.D	8.82	3.79
AR1221ICC500	AR1221ICC500	05/19/2025	11:05	PP072169.D	8.82	3.79
AR1232ICC1000	AR1232ICC1000	05/19/2025	11:21	PP072170.D	8.82	3.80
AR1232ICC750	AR1232ICC750	05/19/2025	11:38	PP072171.D	8.82	3.80
AR1232ICC500	AR1232ICC500	05/19/2025	11:54	PP072172.D	8.82	3.80
AR1232ICC250	AR1232ICC250	05/19/2025	12:10	PP072173.D	8.82	3.80
AR1232ICC050	AR1232ICC050	05/19/2025	12:27	PP072174.D	8.82	3.80
AR1242ICC1000	AR1242ICC1000	05/19/2025	12:43	PP072175.D	8.83	3.80
AR1242ICC750	AR1242ICC750	05/19/2025	12:59	PP072176.D	8.83	3.80
AR1242ICC500	AR1242ICC500	05/19/2025	13:15	PP072177.D	8.83	3.80
AR1242ICC250	AR1242ICC250	05/19/2025	13:32	PP072178.D	8.83	3.80
AR1242ICC050	AR1242ICC050	05/19/2025	13:48	PP072179.D	8.82	3.80
AR1248ICC1000	AR1248ICC1000	05/19/2025	14:04	PP072180.D	8.83	3.80
AR1248ICC750	AR1248ICC750	05/19/2025	14:20	PP072181.D	8.83	3.80
AR1248ICC500	AR1248ICC500	05/19/2025	14:37	PP072182.D	8.83	3.80
AR1248ICC250	AR1248ICC250	05/19/2025	14:53	PP072183.D	8.83	3.80
AR1248ICC050	AR1248ICC050	05/19/2025	15:25	PP072184.D	8.82	3.80
AR1254ICC1000	AR1254ICC1000	05/19/2025	15:42	PP072185.D	8.83	3.80
AR1254ICC750	AR1254ICC750	05/19/2025	15:58	PP072186.D	8.83	3.80
AR1254ICC500	AR1254ICC500	05/19/2025	16:14	PP072187.D	8.83	3.80
AR1254ICC250	AR1254ICC250	05/19/2025	16:31	PP072188.D	8.83	3.80
AR1254ICC050	AR1254ICC050	05/19/2025	16:47	PP072189.D	8.83	3.80
AR1262ICC500	AR1262ICC500	05/19/2025	17:03	PP072190.D	8.83	3.80
AR1268ICC1000	AR1268ICC1000	05/19/2025	17:20	PP072191.D	8.83	3.80
AR1268ICC750	AR1268ICC750	05/19/2025	17:36	PP072192.D	8.83	3.80
AR1268ICC500	AR1268ICC500	05/19/2025	17:53	PP072193.D	8.83	3.80
AR1268ICC250	AR1268ICC250	05/19/2025	18:09	PP072194.D	8.82	3.80
AR1268ICC050	AR1268ICC050	05/19/2025	18:25	PP072195.D	8.82	3.80
AR1660CCC500	AR1660CCC500	05/30/2025	10:21	PP072461.D	8.82	3.80
LBLK	LBLK	05/30/2025	11:29	PP072465.D	8.82	3.80
PB168208BL	PB168208BL	05/30/2025	12:59	PP072466.D	8.82	3.80
PB168208BS	PB168208BS	05/30/2025	13:15	PP072467.D	8.82	3.80
TP-44	Q2150-01	05/30/2025	13:32	PP072468.D	8.82	3.80
TP-42	Q2150-02	05/30/2025	13:48	PP072469.D	8.82	3.80
TP-42MS	Q2150-02MS	05/30/2025	14:04	PP072470.D	8.82	3.80
TP-42MSD	Q2150-02MSD	05/30/2025	14:20	PP072471.D	8.82	3.80
TP-39	Q2150-03	05/30/2025	14:37	PP072472.D	8.82	3.80

Analytical Sequence

TP-48	Q2150-04	05/30/2025	14:53	PP072473.D	8.82	3.80
TP-47	Q2150-05	05/30/2025	15:10	PP072474.D	8.82	3.80
TP-50	Q2150-06	05/30/2025	15:26	PP072475.D	8.82	3.80
AR1660CCC500	AR1660CCC500	05/30/2025	16:15	PP072476.D	8.82	3.80
LBLK	LBLK	05/30/2025	17:20	PP072480.D	8.82	3.80
TP-51	Q2150-07	05/30/2025	17:37	PP072481.D	8.82	3.80
TP-52	Q2150-08	05/30/2025	17:53	PP072482.D	8.82	3.80
TP-54	Q2150-09	05/30/2025	18:09	PP072483.D	8.82	3.80
TP-53	Q2150-10	05/30/2025	18:26	PP072484.D	8.82	3.80
AR1660CCC500	AR1660CCC500	05/30/2025	21:09	PP072491.D	8.82	3.80
LBLK	LBLK	05/30/2025	22:31	PP072495.D	8.82	3.80

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB168208BS

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Lab Sample ID: PB168208BS Date(s) Analyzed: 05/30/2025 05/30/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072467.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.655	5.605	5.705	112	121	6.4
COLUMN 1	2	5.677	5.627	5.727	126		
	3	5.74	5.69	5.79	123		
	4	5.837	5.787	5.887	125		
	5	6.129	6.079	6.179	120		
	1	4.879	4.829	4.929	128		
COLUMN 2	2	4.897	4.847	4.947	131	129	
	3	5.074	5.024	5.124	131		
	4	5.115	5.065	5.165	130		
	5	5.329	5.279	5.379	128		
Aroclor-1260	1	7.248	7.198	7.298	131	123	8.56
COLUMN 1	2	7.501	7.451	7.551	129		
	3	7.86	7.81	7.91	115		
	4	8.084	8.034	8.134	121		
	5	8.404	8.354	8.454	120		
	1	6.362	6.312	6.412	140	134	
COLUMN 2	2	6.55	6.5	6.6	133		
	3	6.703	6.653	6.753	144		
	4	7.173	7.123	7.223	128		
	5	7.416	7.366	7.466	125		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TP-42MS

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Lab Sample ID: Q2150-02MS Date(s) Analyzed: 05/30/2025 05/30/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072470.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.654	5.604	5.704	70.2	72.6	16.43
COLUMN 1	2	5.676	5.626	5.726	73.6		
	3	5.737	5.687	5.787	72.1		
	4	5.835	5.785	5.885	76.2		
	5	6.128	6.078	6.178	70.6		
COLUMN 2	1	4.879	4.829	4.929	84.9	85.6	
	2	4.897	4.847	4.947	85.9		
	3	5.075	5.025	5.125	87.1		
	4	5.116	5.066	5.166	85.8		
	5	5.33	5.28	5.38	84.5		
Aroclor-1260	1	7.245	7.195	7.295	69.8	64.4	18.08
COLUMN 1	2	7.5	7.45	7.55	69.1		
	3	7.858	7.808	7.908	57.9		
	4	8.081	8.031	8.131	66.0		
	5	8.4	8.35	8.45	59.4		
COLUMN 2	1	6.363	6.313	6.413	87.7	77.2	
	2	6.552	6.502	6.602	77.8		
	3	6.704	6.654	6.754	82.5		
	4	7.175	7.125	7.225	70.2		
	5	7.417	7.367	7.467	67.7		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

TP-42MSD

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Lab Sample ID: Q2150-02MSD Date(s) Analyzed: 05/30/2025 05/30/2025

Instrument ID (1): ECD_P Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP072471.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.655	5.605	5.705	69.0	74.5	14.91
COLUMN 1	2	5.676	5.626	5.726	78.8		
	3	5.739	5.689	5.789	73.6		
	4	5.836	5.786	5.886	77.5		
	5	6.129	6.079	6.179	73.4		
COLUMN 2	1	4.88	4.83	4.93	86.0	86.5	
	2	4.898	4.848	4.948	85.7		
	3	5.075	5.025	5.125	87.7		
	4	5.116	5.066	5.166	87.5		
	5	5.331	5.281	5.381	85.7		
Aroclor-1260	1	7.246	7.196	7.296	73.3	67.3	16.37
COLUMN 1	2	7.501	7.451	7.551	71.1		
	3	7.859	7.809	7.909	60.9		
	4	8.084	8.034	8.134	69.5		
	5	8.404	8.354	8.454	61.5		
COLUMN 2	1	6.363	6.313	6.413	89.3	79.3	
	2	6.552	6.502	6.602	79.1		
	3	6.704	6.654	6.754	84.7		
	4	7.175	7.125	7.225	73.4		
	5	7.417	7.367	7.467	69.9		



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	PB168208BL	SDG No.:	Q2150
Lab Sample ID:	PB168208BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	Injection Volume :	
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072466.D	1	05/30/25 08:50	05/30/25 12:59	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.6		32 - 144	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.3		32 - 175	122%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072466.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 12:59
Operator : YP\AJ
Sample : PB168208BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB168208BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 14:37:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.505	3.796	39897071	34561223	19.841	21.623
2) SA Decachlor...	10.214	8.818	34550231	25733695	21.208	24.309

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072466.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 12:59
Operator : YP\AJ
Sample : PB168208BL
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB168208BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 14:37:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.505 min
Delta R.T.: 0.007 min
Response: 39897071
Conc: 19.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB168208BL




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.003 min
Response: 34561223
Conc: 21.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.214 min
Delta R.T.: 0.002 min
Response: 34550231
Conc: 21.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.818 min
Delta R.T.: -0.003 min
Response: 25733695
Conc: 24.31 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/19/25
Project:	South River WM Replacement	Date Received:	05/19/25
Client Sample ID:	PIBLK-PP072163.D	SDG No.:	Q2150
Lab Sample ID:	I.BLK-PP072163.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072163.D	1		05/19/25	PP051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.3		60 - 140	82%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.7		60 - 140	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072163.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 09:26
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:35:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.495	3.788	32787184	26552565	16.306	16.612
2) SA Decachlor...	10.213	8.822	27208113	18027788	16.701	17.030

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP051925\
Data File : PP072163.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2025 09:26
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 03:35:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 20 03:29:33 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.495 min
Delta R.T.: -0.003 min
Response: 32787184
Conc: 16.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.788 min
Delta R.T.: -0.004 min
Response: 26552565
Conc: 16.61 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.213 min
Delta R.T.: 0.000 min
Response: 27208113
Conc: 16.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.822 min
Delta R.T.: 0.000 min
Response: 18027788
Conc: 17.03 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/30/25
Project:	South River WM Replacement	Date Received:	05/30/25
Client Sample ID:	PIBLK-PP072465.D	SDG No.:	Q2150
Lab Sample ID:	I.BLK-PP072465.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072465.D	1		05/30/25	pp053025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.1		60 - 140	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.5		60 - 140	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072465.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 11:29
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 11:39:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.503	3.797	36428540	32017402	18.116	20.031
2) SA Decachlor...	10.210	8.817	31762850	22931451	19.497	21.662

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072465.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 11:29
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 11:39:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Response_ Signal: PP072465.D\ECD1A.ch




#1 Tetrachloro-m-xylene

R.T.: 4.503 min
Delta R.T.: 0.005 min
Response: 36428540
Conc: 18.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK


Response_ Signal: PP072465.D\ECD2B.ch



#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.004 min
Response: 32017402
Conc: 20.03 ng/ml


Response_ Signal: PP072465.D\ECD1A.ch



#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: -0.002 min
Response: 31762850
Conc: 19.50 ng/ml

Response_ Signal: PP072465.D\ECD2B.ch



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 22931451
Conc: 21.66 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/30/25
Project:	South River WM Replacement	Date Received:	05/30/25
Client Sample ID:	PIBLK-PP072480.D	SDG No.:	Q2150
Lab Sample ID:	I.BLK-PP072480.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072480.D	1		05/30/25	pp053025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.7		60 - 140	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.4		60 - 140	97%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 17:20
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 19:13:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.501	3.797	37550594	32095375	18.674	20.080
2) SA Decachlor...	10.207	8.818	31609389	22629892	19.403	21.377

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072480.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 17:20
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 19:13:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.501 min
Delta R.T.: 0.003 min
Response: 37550594
Conc: 18.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.797 min
Delta R.T.: 0.003 min
Response: 32095375
Conc: 20.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.207 min
Delta R.T.: -0.005 min
Response: 31609389
Conc: 19.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.818 min
Delta R.T.: -0.004 min
Response: 22629892
Conc: 21.38 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	05/30/25
Project:	South River WM Replacement	Date Received:	05/30/25
Client Sample ID:	PIBLK-PP072495.D	SDG No.:	Q2150
Lab Sample ID:	I.BLK-PP072495.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072495.D	1		05/30/25	pp053025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11104-28-2	Aroclor-1221	0.13	U	0.13	0.50	ug/L
11141-16-5	Aroclor-1232	0.096	U	0.096	0.50	ug/L
53469-21-9	Aroclor-1242	0.12	U	0.12	0.50	ug/L
12672-29-6	Aroclor-1248	0.071	U	0.071	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
37324-23-5	Aroclor-1262	0.14	U	0.14	0.50	ug/L
11100-14-4	Aroclor-1268	0.11	U	0.11	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.4		60 - 140	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.5		60 - 140	98%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072495.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 22:31
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 01:29:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.796	36990832	33785481	18.396	21.138
2) SA Decachlor...	10.209	8.817	31816117	22291137	19.530	21.057

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072495.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 22:31
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 01:29:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.502 min
Delta R.T.: 0.004 min
Response: 36990832
Conc: 18.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.796 min
Delta R.T.: 0.003 min
Response: 33785481
Conc: 21.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.209 min
Delta R.T.: -0.003 min
Response: 31816117
Conc: 19.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.817 min
Delta R.T.: -0.004 min
Response: 22291137
Conc: 21.06 ng/ml

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	PB168208BS	SDG No.:	Q2150
Lab Sample ID:	PB168208BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072467.D	1	05/30/25 08:50	05/30/25 13:15	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	129		3.90	17.0	ug/kg
11104-28-2	Aroclor-1221	4.00	U	4.00	17.0	ug/kg
11141-16-5	Aroclor-1232	3.70	U	3.70	17.0	ug/kg
53469-21-9	Aroclor-1242	4.00	U	4.00	17.0	ug/kg
12672-29-6	Aroclor-1248	5.90	U	5.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
37324-23-5	Aroclor-1262	5.00	U	5.00	17.0	ug/kg
11100-14-4	Aroclor-1268	3.60	U	3.60	17.0	ug/kg
11096-82-5	Aroclor-1260	134		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.6		32 - 175	118%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072467.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 13:15
 Operator : YP\AJ
 Sample : PB168208BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB168208BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 14:37:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.503	3.797	40600987	32347143	20.191	20.238
2)	SA Decachlor...	10.211	8.817	34940595	24942907	21.447	23.562
Target Compounds							
3)	L1 AR-1016-1	5.655	4.879	25342147	23593187	337.519	384.069
4)	L1 AR-1016-2	5.677	4.897	40534084	34426586	378.853	392.150
5)	L1 AR-1016-3	5.740	5.074	24252393	18689847	369.899	392.195
6)	L1 AR-1016-4	5.837	5.115	19964068	14895177	374.771	391.511
7)	L1 AR-1016-5	6.129	5.329	17606509	19107202	359.984	383.491
31)	L7 AR-1260-1	7.248	6.362	36724421	33579849	392.377	421.044
32)	L7 AR-1260-2	7.501	6.550	55483664	40642782	386.673	400.729
33)	L7 AR-1260-3	7.860	6.703	39322171	37691790	345.323	433.726 #
34)	L7 AR-1260-4	8.084	7.173	39129847	27783719	364.668	385.398
35)	L7 AR-1260-5	8.404	7.416	85600897	65393382	361.701	376.638



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072467.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 13:15
Operator : YP\AJ
Sample : PB168208BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB168208BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 14:37:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-42MS	SDG No.:	Q2150
Lab Sample ID:	Q2150-02MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.2
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072470.D	1	05/30/25 08:50	05/30/25 14:04	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	85.6		4.60	19.7	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.7	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.7	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	19.7	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.7	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.7	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.7	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.7	ug/kg
11096-82-5	Aroclor-1260	77.2		3.70	19.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.8		32 - 144	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.1		32 - 175	81%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072470.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 14:04
 Operator : YP\AJ
 Sample : Q2150-02MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-42MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 14:39:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.502	3.798	36366117	29999123	18.085	18.769
2)	SA Decachlor...	10.210	8.819	23299460	17086679	14.302	16.141
Target Compounds							
3)	L1 AR-1016-1	5.654	4.879	13637863	13494876	181.636	219.681
4)	L1 AR-1016-2	5.676	4.897	20368415	19497777	190.374	222.097
5)	L1 AR-1016-3	5.737	5.075	12228267	10733005	186.506	225.225
6)	L1 AR-1016-4	5.835	5.116	10502513	8443389	197.156	221.930
7)	L1 AR-1016-5	6.128	5.330	8938075	10885976	182.748	218.487
31)	L7 AR-1260-1	7.245	6.363	16893593	18100231	180.497m	226.952 #
32)	L7 AR-1260-2	7.500	6.552	25645069	20404808	178.724	201.187
33)	L7 AR-1260-3	7.858	6.704	17046078	18537355	149.697	213.312 #
34)	L7 AR-1260-4	8.081	7.175	18326610	13096194	170.794m	181.662
35)	L7 AR-1260-5	8.400	7.417	36349425	30393370	153.592m	175.053



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072470.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 14:04
Operator : YP\AJ
Sample : Q2150-02MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-42MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 14:39:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Report of Analysis

Client:	CDM Smith	Date Collected:	05/27/25
Project:	South River WM Replacement	Date Received:	05/28/25
Client Sample ID:	TP-42MSD	SDG No.:	Q2150
Lab Sample ID:	Q2150-02MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	86.2
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072471.D	1	05/30/25 08:50	05/30/25 14:20	PB168208

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	86.5		4.60	19.7	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.7	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.7	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.7	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.7	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.7	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.7	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.7	ug/kg
11096-82-5	Aroclor-1260	79.3		3.70	19.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.8		32 - 144	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.8		32 - 175	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
 Data File : PP072471.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2025 14:20
 Operator : YP\AJ
 Sample : Q2150-02MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-42MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 14:39:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat May 24 03:32:20 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.502	3.798	37967707	31693917	18.882	19.829
2) SA Decachlor...	10.212	8.819	24822927	17755240	15.237	16.772
Target Compounds						
3) L1 AR-1016-1	5.655	4.880	13407620	13664758	178.569	222.446
4) L1 AR-1016-2	5.676	4.898	21818116	19464055	203.923	221.713
5) L1 AR-1016-3	5.739	5.075	12494305	10814309	190.564	226.932
6) L1 AR-1016-4	5.836	5.116	10684673	8612907	200.576	226.385
7) L1 AR-1016-5	6.129	5.331	9286117	11044932	189.864	221.677
31) L7 AR-1260-1	7.246	6.363	17761827	18437406	189.774m	231.179
32) L7 AR-1260-2	7.501	6.552	26384208	20748512	183.875	204.576
33) L7 AR-1260-3	7.859	6.704	17953838	19051220	157.668	219.226 #
34) L7 AR-1260-4	8.084	7.175	19303720	13694219	179.900	189.958
35) L7 AR-1260-5	8.404	7.417	37692966	31398527	159.269	180.842



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP053025\
Data File : PP072471.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2025 14:20
Operator : YP\AJ
Sample : Q2150-02MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
TP-42MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 14:39:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP051925.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat May 24 03:32:20 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm



Manual Integration Report

Sequence:	PP051925	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PP072168.D	AR-1016-1	yogesh	5/20/2025 7:28:08 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1660ICC050	PP072168.D	AR-1016-2	yogesh	5/20/2025 7:28:08 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1660ICC050	PP072168.D	AR-1016-3	yogesh	5/20/2025 7:28:08 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1660ICC050	PP072168.D	AR-1016-4	yogesh	5/20/2025 7:28:08 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1232ICC050	PP072174.D	AR-1232-3	yogesh	5/21/2025 8:50:30 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1232ICC050	PP072174.D	AR-1232-4	yogesh	5/21/2025 8:50:30 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-1	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-2	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-3	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-4	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-5	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1242ICC050	PP072179.D	AR-1242-5 #2	yogesh	5/20/2025 7:28:10 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1248ICC050	PP072184.D	AR-1248-1	yogesh	5/20/2025 7:28:12 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software

Manual Integration Report

Sequence:

PP051925

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1248ICC050	PP072184.D	AR-1248-4 #2	yogesh	5/20/2025 7:28:12 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	AR-1254-1	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	AR-1254-4	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	AR-1254-4 #2	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	Tetrachloro-m-xylene	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1254ICC050	PP072189.D	Tetrachloro-m-xylene #2	yogesh	5/20/2025 7:28:14 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1268ICC050	PP072195.D	AR-1268-1	yogesh	5/20/2025 7:28:15 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software
AR1268ICC050	PP072195.D	Tetrachloro-m-xylene	yogesh	5/20/2025 7:28:15 AM	mohammad	5/22/2025 6:37:35	Peak Integrated by Software

Manual Integration Report

Sequence:

pp053025

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2150-01	PP072468.D	Tetrachloro-m-xylene	yogesh	6/2/2025 7:25:13 AM	 	 	Peak Integrated by Software
Q2150-02MS	PP072470.D	AR-1260-1	yogesh	6/2/2025 7:25:15 AM	 	 	Peak Integrated by Software
Q2150-02MS	PP072470.D	AR-1260-4	yogesh	6/2/2025 7:25:15 AM	 	 	Peak Integrated by Software
Q2150-02MS	PP072470.D	AR-1260-5	yogesh	6/2/2025 7:25:15 AM	 	 	Peak Integrated by Software
Q2150-02MSD	PP072471.D	AR-1260-1	yogesh	6/2/2025 7:25:17 AM	 	 	Peak Integrated by Software
Q2150-06	PP072475.D	Tetrachloro-m-xylene	yogesh	6/2/2025 7:25:18 AM	 	 	Peak Integrated by Software
Q2150-08	PP072482.D	Tetrachloro-m-xylene	yogesh	6/2/2025 7:25:20 AM	 	 	Peak Integrated by Software

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344, PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359, PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP072162.D	19 May 2025 09:09	YP\AJ	Ok
2	I.BLK	PP072163.D	19 May 2025 09:26	YP\AJ	Ok
3	AR1660ICC1000	PP072164.D	19 May 2025 09:42	YP\AJ	Ok
4	AR1660ICC750	PP072165.D	19 May 2025 09:59	YP\AJ	Ok
5	AR1660ICC500	PP072166.D	19 May 2025 10:15	YP\AJ	Ok
6	AR1660ICC250	PP072167.D	19 May 2025 10:32	YP\AJ	Ok
7	AR1660ICC050	PP072168.D	19 May 2025 10:48	YP\AJ	Ok,M
8	AR1221ICC500	PP072169.D	19 May 2025 11:05	YP\AJ	Ok
9	AR1232ICC1000	PP072170.D	19 May 2025 11:21	YP\AJ	Ok
10	AR1232ICC750	PP072171.D	19 May 2025 11:38	YP\AJ	Ok
11	AR1232ICC500	PP072172.D	19 May 2025 11:54	YP\AJ	Ok
12	AR1232ICC250	PP072173.D	19 May 2025 12:10	YP\AJ	Ok
13	AR1232ICC050	PP072174.D	19 May 2025 12:27	YP\AJ	Ok,M
14	AR1242ICC1000	PP072175.D	19 May 2025 12:43	YP\AJ	Ok
15	AR1242ICC750	PP072176.D	19 May 2025 12:59	YP\AJ	Ok
16	AR1242ICC500	PP072177.D	19 May 2025 13:15	YP\AJ	Ok
17	AR1242ICC250	PP072178.D	19 May 2025 13:32	YP\AJ	Ok
18	AR1242ICC050	PP072179.D	19 May 2025 13:48	YP\AJ	Ok,M
19	AR1248ICC1000	PP072180.D	19 May 2025 14:04	YP\AJ	Ok
20	AR1248ICC750	PP072181.D	19 May 2025 14:20	YP\AJ	Ok
21	AR1248ICC500	PP072182.D	19 May 2025 14:37	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC			
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1248ICC250	PP072183.D	19 May 2025 14:53	YP\AJ	Ok
23	AR1248ICC050	PP072184.D	19 May 2025 15:25	YP\AJ	Ok,M
24	AR1254ICC1000	PP072185.D	19 May 2025 15:42	YP\AJ	Ok
25	AR1254ICC750	PP072186.D	19 May 2025 15:58	YP\AJ	Ok
26	AR1254ICC500	PP072187.D	19 May 2025 16:14	YP\AJ	Ok
27	AR1254ICC250	PP072188.D	19 May 2025 16:31	YP\AJ	Ok
28	AR1254ICC050	PP072189.D	19 May 2025 16:47	YP\AJ	Ok,M
29	AR1262ICC500	PP072190.D	19 May 2025 17:03	YP\AJ	Ok
30	AR1268ICC1000	PP072191.D	19 May 2025 17:20	YP\AJ	Ok
31	AR1268ICC750	PP072192.D	19 May 2025 17:36	YP\AJ	Ok
32	AR1268ICC500	PP072193.D	19 May 2025 17:53	YP\AJ	Ok
33	AR1268ICC250	PP072194.D	19 May 2025 18:09	YP\AJ	Ok
34	AR1268ICC050	PP072195.D	19 May 2025 18:25	YP\AJ	Ok,M
35	PP051925ICV500	PP072196.D	19 May 2025 18:42	YP\AJ	Ok
36	AR1232ICV500	PP072197.D	19 May 2025 19:14	YP\AJ	Ok
37	AR1242ICV500	PP072198.D	19 May 2025 19:47	YP\AJ	Ok
38	AR1248ICV500	PP072199.D	19 May 2025 20:19	YP\AJ	Ok
39	AR1254ICV500	PP072200.D	19 May 2025 20:52	YP\AJ	Ok
40	AR1268ICV500	PP072201.D	19 May 2025 21:25	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP053025

Review By	yogesh	Review On	5/30/2025 1:38:04 PM
Supervise By		Supervise On	
SubDirectory	PP053025	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP072460.D	30 May 2025 10:05	YP\AJ	Ok
2	AR1660CCC500	PP072461.D	30 May 2025 10:21	YP\AJ	Ok
3	AR1242CCC500	PP072462.D	30 May 2025 10:40	YP\AJ	Ok
4	AR1248CCC500	PP072463.D	30 May 2025 10:57	YP\AJ	Ok
5	AR1254CCC500	PP072464.D	30 May 2025 11:13	YP\AJ	Ok
6	I.BLK	PP072465.D	30 May 2025 11:29	YP\AJ	Ok
7	PB168208BL	PP072466.D	30 May 2025 12:59	YP\AJ	Ok
8	PB168208BS	PP072467.D	30 May 2025 13:15	YP\AJ	Ok
9	Q2150-01	PP072468.D	30 May 2025 13:32	YP\AJ	Ok,NS
10	Q2150-02	PP072469.D	30 May 2025 13:48	YP\AJ	Ok
11	Q2150-02MS	PP072470.D	30 May 2025 14:04	YP\AJ	Ok,NS
12	Q2150-02MSD	PP072471.D	30 May 2025 14:20	YP\AJ	Ok,NS
13	Q2150-03	PP072472.D	30 May 2025 14:37	YP\AJ	Ok
14	Q2150-04	PP072473.D	30 May 2025 14:53	YP\AJ	Ok
15	Q2150-05	PP072474.D	30 May 2025 15:10	YP\AJ	Ok
16	Q2150-06	PP072475.D	30 May 2025 15:26	YP\AJ	Ok,NS
17	AR1660CCC500	PP072476.D	30 May 2025 16:15	YP\AJ	Ok
18	AR1242CCC500	PP072477.D	30 May 2025 16:31	YP\AJ	Ok
19	AR1248CCC500	PP072478.D	30 May 2025 16:48	YP\AJ	Ok
20	AR1254CCC500	PP072479.D	30 May 2025 17:04	YP\AJ	Ok
21	I.BLK	PP072480.D	30 May 2025 17:20	YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP053025

Review By	yogesh	Review On	5/30/2025 1:38:04 PM
Supervise By		Supervise On	
SubDirectory	PP053025	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2150-07	PP072481.D	30 May 2025 17:37	YP\AJ	Ok
23	Q2150-08	PP072482.D	30 May 2025 17:53	YP\AJ	Ok,NS
24	Q2150-09	PP072483.D	30 May 2025 18:09	YP\AJ	Ok
25	Q2150-10	PP072484.D	30 May 2025 18:26	YP\AJ	Ok
26	Q2151-01	PP072485.D	30 May 2025 18:42	YP\AJ	Ok
27	Q2152-01	PP072486.D	30 May 2025 18:58	YP\AJ	Ok,NS
28	Q2153-01	PP072487.D	30 May 2025 19:15	YP\AJ	Ok,NS
29	Q2155-01	PP072488.D	30 May 2025 19:31	YP\AJ	Ok,NS
30	Q2155-02	PP072489.D	30 May 2025 19:48	YP\AJ	Ok,NS
31	Q2157-01	PP072490.D	30 May 2025 20:04	YP\AJ	Ok
32	AR1660CCC500	PP072491.D	30 May 2025 21:09	YP\AJ	Ok
33	AR1242CCC500	PP072492.D	30 May 2025 21:42	YP\AJ	Ok
34	AR1248CCC500	PP072493.D	30 May 2025 21:59	YP\AJ	Ok
35	AR1254CCC500	PP072494.D	30 May 2025 22:15	YP\AJ	Ok
36	I.BLK	PP072495.D	30 May 2025 22:31	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072162.D	19 May 2025 09:09		YP\AJ	Ok
2	I.BLK	I.BLK	PP072163.D	19 May 2025 09:26		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP072164.D	19 May 2025 09:42		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP072165.D	19 May 2025 09:59		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP072166.D	19 May 2025 10:15		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP072167.D	19 May 2025 10:32		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP072168.D	19 May 2025 10:48		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP072169.D	19 May 2025 11:05		YP\AJ	Ok
9	AR1232ICC1000	AR1232ICC1000	PP072170.D	19 May 2025 11:21		YP\AJ	Ok
10	AR1232ICC750	AR1232ICC750	PP072171.D	19 May 2025 11:38		YP\AJ	Ok
11	AR1232ICC500	AR1232ICC500	PP072172.D	19 May 2025 11:54		YP\AJ	Ok
12	AR1232ICC250	AR1232ICC250	PP072173.D	19 May 2025 12:10		YP\AJ	Ok
13	AR1232ICC050	AR1232ICC050	PP072174.D	19 May 2025 12:27		YP\AJ	Ok,M
14	AR1242ICC1000	AR1242ICC1000	PP072175.D	19 May 2025 12:43		YP\AJ	Ok
15	AR1242ICC750	AR1242ICC750	PP072176.D	19 May 2025 12:59		YP\AJ	Ok
16	AR1242ICC500	AR1242ICC500	PP072177.D	19 May 2025 13:15		YP\AJ	Ok
17	AR1242ICC250	AR1242ICC250	PP072178.D	19 May 2025 13:32		YP\AJ	Ok
18	AR1242ICC050	AR1242ICC050	PP072179.D	19 May 2025 13:48		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC1000	AR1248ICC1000	PP072180.D	19 May 2025 14:04		YPIAJ	Ok
20	AR1248ICC750	AR1248ICC750	PP072181.D	19 May 2025 14:20		YPIAJ	Ok
21	AR1248ICC500	AR1248ICC500	PP072182.D	19 May 2025 14:37		YPIAJ	Ok
22	AR1248ICC250	AR1248ICC250	PP072183.D	19 May 2025 14:53		YPIAJ	Ok
23	AR1248ICC050	AR1248ICC050	PP072184.D	19 May 2025 15:25		YPIAJ	Ok,M
24	AR1254ICC1000	AR1254ICC1000	PP072185.D	19 May 2025 15:42		YPIAJ	Ok
25	AR1254ICC750	AR1254ICC750	PP072186.D	19 May 2025 15:58		YPIAJ	Ok
26	AR1254ICC500	AR1254ICC500	PP072187.D	19 May 2025 16:14		YPIAJ	Ok
27	AR1254ICC250	AR1254ICC250	PP072188.D	19 May 2025 16:31		YPIAJ	Ok
28	AR1254ICC050	AR1254ICC050	PP072189.D	19 May 2025 16:47		YPIAJ	Ok,M
29	AR1262ICC500	AR1262ICC500	PP072190.D	19 May 2025 17:03		YPIAJ	Ok
30	AR1268ICC1000	AR1268ICC1000	PP072191.D	19 May 2025 17:20		YPIAJ	Ok
31	AR1268ICC750	AR1268ICC750	PP072192.D	19 May 2025 17:36		YPIAJ	Ok
32	AR1268ICC500	AR1268ICC500	PP072193.D	19 May 2025 17:53		YPIAJ	Ok
33	AR1268ICC250	AR1268ICC250	PP072194.D	19 May 2025 18:09		YPIAJ	Ok
34	AR1268ICC050	AR1268ICC050	PP072195.D	19 May 2025 18:25		YPIAJ	Ok,M
35	PP051925ICV500	ICVPP051925	PP072196.D	19 May 2025 18:42		YPIAJ	Ok
36	AR1232ICV500	ICVPP051925AR1232	PP072197.D	19 May 2025 19:14		YPIAJ	Ok
37	AR1242ICV500	ICVPP051925AR1242	PP072198.D	19 May 2025 19:47		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

38	AR1248ICV500	ICVPP051925AR1248	PP072199.D	19 May 2025 20:19		YPIAJ	Ok
39	AR1254ICV500	ICVPP051925AR1254	PP072200.D	19 May 2025 20:52		YPIAJ	Ok
40	AR1268ICV500	ICVPP051925AR1268	PP072201.D	19 May 2025 21:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP053025

Review By	yogesh	Review On	5/30/2025 1:38:04 PM
Supervise By		Supervise On	
SubDirectory	PP053025	HP Acquire Method	HP Processing Method PP051925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072460.D	30 May 2025 10:05		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP072461.D	30 May 2025 10:21		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP072462.D	30 May 2025 10:40		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP072463.D	30 May 2025 10:57		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP072464.D	30 May 2025 11:13		YPIAJ	Ok
6	I.BLK	I.BLK	PP072465.D	30 May 2025 11:29		YPIAJ	Ok
7	PB168208BL	PB168208BL	PP072466.D	30 May 2025 12:59		YPIAJ	Ok
8	PB168208BS	PB168208BS	PP072467.D	30 May 2025 13:15		YPIAJ	Ok
9	Q2150-01	TP-44	PP072468.D	30 May 2025 13:32		YPIAJ	Ok,NS
10	Q2150-02	TP-42	PP072469.D	30 May 2025 13:48		YPIAJ	Ok
11	Q2150-02MS	TP-42MS	PP072470.D	30 May 2025 14:04	AR1016 recovery fail	YPIAJ	Ok,NS
12	Q2150-02MSD	TP-42MSD	PP072471.D	30 May 2025 14:20	AR1016 recovery fail	YPIAJ	Ok,NS
13	Q2150-03	TP-39	PP072472.D	30 May 2025 14:37		YPIAJ	Ok
14	Q2150-04	TP-48	PP072473.D	30 May 2025 14:53		YPIAJ	Ok
15	Q2150-05	TP-47	PP072474.D	30 May 2025 15:10		YPIAJ	Ok
16	Q2150-06	TP-50	PP072475.D	30 May 2025 15:26		YPIAJ	Ok,NS
17	AR1660CCC500	AR1660CCC500	PP072476.D	30 May 2025 16:15		YPIAJ	Ok
18	AR1242CCC500	AR1242CCC500	PP072477.D	30 May 2025 16:31		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP053025

Review By	yogesh	Review On	5/30/2025 1:38:04 PM
Supervise By		Supervise On	
SubDirectory	PP053025	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP072478.D	30 May 2025 16:48		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP072479.D	30 May 2025 17:04		YPIAJ	Ok
21	I.BLK	I.BLK	PP072480.D	30 May 2025 17:20		YPIAJ	Ok
22	Q2150-07	TP-51	PP072481.D	30 May 2025 17:37		YPIAJ	Ok
23	Q2150-08	TP-52	PP072482.D	30 May 2025 17:53		YPIAJ	Ok,NS
24	Q2150-09	TP-54	PP072483.D	30 May 2025 18:09		YPIAJ	Ok
25	Q2150-10	TP-53	PP072484.D	30 May 2025 18:26		YPIAJ	Ok
26	Q2151-01	WC-1	PP072485.D	30 May 2025 18:42		YPIAJ	Ok
27	Q2152-01	OK-02-05292025	PP072486.D	30 May 2025 18:58		YPIAJ	Ok,NS
28	Q2153-01	TR-04-0592025	PP072487.D	30 May 2025 19:15		YPIAJ	Ok,NS
29	Q2155-01	446	PP072488.D	30 May 2025 19:31	AR1260 Hit	YPIAJ	Ok,NS
30	Q2155-02	447-SOLID	PP072489.D	30 May 2025 19:48		YPIAJ	Ok,NS
31	Q2157-01	291	PP072490.D	30 May 2025 20:04		YPIAJ	Ok
32	AR1660CCC500	AR1660CCC500	PP072491.D	30 May 2025 21:09		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP072492.D	30 May 2025 21:42		YPIAJ	Ok
34	AR1248CCC500	AR1248CCC500	PP072493.D	30 May 2025 21:59		YPIAJ	Ok
35	AR1254CCC500	AR1254CCC500	PP072494.D	30 May 2025 22:15		YPIAJ	Ok
36	I.BLK	I.BLK	PP072495.D	30 May 2025 22:31		YPIAJ	Ok

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/30/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 05/29/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 05/30/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135945

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2150-01	TP-44	1	1.15	10.30	11.45	9.28	78.9	
Q2150-02	TP-42	2	1.13	10.45	11.58	10.14	86.2	
Q2150-03	TP-39	3	1.14	10.46	11.6	9.89	83.7	
Q2150-04	TP-48	4	1.18	10.44	11.62	9.6	80.7	
Q2150-05	TP-47	5	1.17	10.18	11.35	9.42	81.0	
Q2150-06	TP-50	6	1.18	10.70	11.88	10.38	86.0	
Q2150-07	TP-51	7	1.17	10.74	11.91	10.16	83.7	
Q2150-08	TP-52	8	1.16	10.61	11.77	10.25	85.7	
Q2150-09	TP-54	9	1.16	10.50	11.66	9.94	83.6	
Q2150-10	TP-53	10	1.12	10.56	11.68	10.13	85.3	
Q2151-01	WC-1	11	1.15	10.02	11.17	9.71	85.4	
Q2151-02	WC-1-EPH	12	1.17	10.23	11.4	9.99	86.2	
Q2151-03	WC-1-VOC	13	1.19	10.41	11.6	9.98	84.4	
Q2152-01	OK-02-05292025	31	1.18	10.25	11.43	10.2	88.0	
Q2152-02	OK-02-05292025-E2	32	1.13	10.47	11.6	10.54	89.9	
Q2153-01	TR-04-0592025	33	1.15	10.40	11.55	10.54	90.3	
Q2153-02	TR-04-0592025-E2	34	1.17	10.39	11.56	10.6	90.8	
Q2154-01	BC271336-1-1	14	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-02	BC271336-1-2	15	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-03	BC271336-2-1	16	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-04	BC271336-2-2	17	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-05	BC278406-1-1	18	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-06	BC278406-1-2	19	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-07	BC278406-2-1	20	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-08	BC278406-2-2	21	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-09	BC22682C-1-1	22	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-10	BC22682C-1-2	23	1.00	1.00	2.00	2.00	100.0	oilc
Q2154-11	GHA409H-1-1	24	1.00	1.00	2.00	2.00	100.0	oilc



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 5/30/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 05/29/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 05/30/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135945

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2154-12	GHA409H-1-2	25	1.00	1.00	2.00	2.00	100.0	oilc
Q2155-01	446	26	1.18	10.02	11.2	10.42	92.2	
Q2155-02	447-SOLID	27	1.13	10.35	11.48	6.13	48.3	
Q2156-01	KZZ331H-1-1	28	1.00	1.00	2.00	2.00	100.0	oilc
Q2156-02	KZZ331H-1-2	29	1.00	1.00	2.00	2.00	100.0	oilc
Q2157-01	291	30	1.00	1.00	2.00	2.00	100.0	debris

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

W 135945

WorkList Name : %1-052925

WorkList ID : 189801

Department : Wet-Chemistry

Date : 05-29-2025 08:14:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2150-01	TP-44	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/27/2025	Chemtech -SO
Q2150-02	TP-42	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/27/2025	Chemtech -SO
Q2150-03	TP-39	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/27/2025	Chemtech -SO
Q2150-04	TP-48	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/27/2025	Chemtech -SO
Q2150-05	TP-47	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/27/2025	Chemtech -SO
Q2150-06	TP-50	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/27/2025	Chemtech -SO
Q2150-07	TP-51	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/27/2025	Chemtech -SO
Q2150-08	TP-52	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/27/2025	Chemtech -SO
Q2150-09	TP-54	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/28/2025	Chemtech -SO
Q2150-10	TP-53	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/28/2025	Chemtech -SO
Q2151-01	WC-1	Solid	Percent Solids	Cool 4 deg C	CAMP02	L41	05/28/2025	Chemtech -SO
Q2151-02	WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/29/2025	Chemtech -SO
Q2151-03	WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/29/2025	Chemtech -SO
Q2152-01	OK-02-05292025	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	05/29/2025	Chemtech -SO
Q2152-02	OK-02-05292025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L21	05/29/2025	Chemtech -SO
Q2153-01	TR-04-0592025	Solid	Percent Solids	Cool 4 deg C	PSEG05	L21	05/29/2025	Chemtech -SO
Q2153-02	TR-04-0592025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	L31	05/29/2025	Chemtech -SO
Q2154-01	BC271336-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG05	L31	05/29/2025	Chemtech -SO
Q2154-02	BC271336-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-03	BC271336-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-04	BC271336-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO

Date/Time 05/29/25 16:15

Raw Sample Received by: R CWC1

Raw Sample Relinquished by: R.C. (Sm)

Date/Time 05/29/25

Raw Sample Received by: R.C. (Sm)

Raw Sample Relinquished by: R.C. (Sm)

WORKLIST(Hardcopy Internal Chain)

135945

WorkList Name : %1-052925

WorkList ID : 189801

Department : Wet-Chemistry

Date : 05-29-2025 08:14:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2154-05	BC278406-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-06	BC278406-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-07	BC278406-2-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-08	BC278406-2-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-09	BC22682C-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-10	BC22682C-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-11	GHA409H-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2154-12	GHA409H-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2155-01	446	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2155-02	447-SOLID	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2156-01	KZZ331H-1-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2156-02	KZZ331H-1-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO
Q2157-01	291	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	05/29/2025	Chemtech -SO

Date/Time 05/29/25 16:15
 Raw Sample Received by: EWG
 Raw Sample Relinquished by: L.C. (son)

Date/Time 05/29/25 17:30
 Raw Sample Received by: L.C. (son)
 Raw Sample Relinquished by: SC 2000

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: Acid Cleanup

Extraction Start Date : 05/30/2025

Matrix : Solid

Extraction Start Time : 08:50

Weigh By: EH

Extraction By: RJ

Extraction End Date : 05/30/2025

Balance check: RJ

Filter By: RJ

Extraction End Time : 11:50

Balance ID: EX-SC-2

pH Meter ID: N/A

Concentration By: EH

pH Strip Lot#: N/A

Hood ID: 3,7

Supervisor By : RUPESH

Extraction Method: ☐ Separatory Funnel

☐ Continous Liquid/Liquid

☐ Sonication

☐ Waste Dilution

☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24461
Surrogate	1.0ML	200 PPB	PP24597
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2613
Baked Na2SO4	N/A	EP2614
Sand	N/A	E2865
Hexane	N/A	E3938
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS723, Q2157-01 used Limited volume as sample is Oily Debris.

KD Bath ID: N/A

Envap ID: NEVAP-02

KD Bath Temperature: N/A


Envap Temperature: 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
5/30/25	RS (Ext Lab)	Y-P. prest HP
11:55	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 05/30/2025

Sample ID	Client Sample ID	Test	g mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168208BL	ABLK208	PCB	30.02	N/A	ritesh	Evelyn	10			U1-1
PB168208BS	ALCS208	PCB	30.03	N/A	ritesh	Evelyn	10			2
Q2150-01	TP-44	PCB	30.05	N/A	ritesh	Evelyn	10	E		3
Q2150-02	TP-42	PCB	30.03	N/A	ritesh	Evelyn	10	E		4
Q2150-02MS	TP-42MS	PCB	30.01	N/A	ritesh	Evelyn	10	E		5
Q2150-02MS D	TP-42MSD	PCB	30.02	N/A	ritesh	Evelyn	10	E		6
Q2150-03	TP-39	PCB	30.08	N/A	ritesh	Evelyn	10	E		U2-1
Q2150-04	TP-48	PCB	30.05	N/A	ritesh	Evelyn	10	E		2
Q2150-05	TP-47	PCB	30.04	N/A	ritesh	Evelyn	10	E		3
Q2150-06	TP-50	PCB	30.02	N/A	ritesh	Evelyn	10	E		4
Q2150-07	TP-51	PCB	30.07	N/A	ritesh	Evelyn	10	E		5
Q2150-08	TP-52	PCB	30.03	N/A	ritesh	Evelyn	10	E		6
Q2150-09	TP-54	PCB	30.09	N/A	ritesh	Evelyn	10	E		U3-1
Q2150-10	TP-53	PCB	30.05	N/A	ritesh	Evelyn	10	E		2
Q2151-01	WC-1	PCB	30.02	N/A	ritesh	Evelyn	10	E		3
Q2152-01	OK-02-05292025	PCB	30.03	N/A	ritesh	Evelyn	10	E		4
Q2153-01	TR-04-0592025	PCB	30.06	N/A	ritesh	Evelyn	10	E		5
Q2155-01	446	PCB	30.02	N/A	ritesh	Evelyn	10	E	Small Particles	6
Q2155-02	447-SOLID	PCB	30.01	N/A	ritesh	Evelyn	10	E	Small Particles	U7-1
Q2157-01	291	PCB	5.08	N/A	ritesh	Evelyn	10	E	Oily Debris	2



16820050

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2150 WorkList ID : 189836 Department : Extraction Date : 05-30-2025 08:28:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2150-01	TP-44	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/27/2025	8082A
Q2150-02	TP-42	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/27/2025	8082A
Q2150-03	TP-39	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/27/2025	8082A
Q2150-04	TP-48	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/27/2025	8082A
Q2150-05	TP-47	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/27/2025	8082A
Q2150-06	TP-50	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/27/2025	8082A
Q2150-07	TP-51	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/27/2025	8082A
Q2150-08	TP-52	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/27/2025	8082A
Q2150-09	TP-54	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/28/2025	8082A
Q2150-10	TP-53	Solid	PCB	Cool 4 deg C	CAMP02	L41	05/28/2025	8082A
Q2151-01	WC-1	Solid	PCB	Cool 4 deg C	PSEG03	L41	05/28/2025	8082A
Q2152-01	OK-02-05292025	Solid	PCB	Cool 4 deg C	PSEG05	L21	05/29/2025	8082A
Q2153-01	TR-04-0592025	Solid	PCB	Cool 4 deg C	PSEG05	L31	05/29/2025	8082A
Q2155-01	446	Solid	PCB	Cool 4 deg C	PSEG03	L31	05/29/2025	8082A
Q2155-02	447-SOLID	Solid	PCB	Cool 4 deg C	PSEG03	L31	05/29/2025	8082A
Q2157-01	291	Solid	PCB	Cool 4 deg C	PSEG03	L31	05/29/2025	8082A

Date/Time 05/30/25 8:45
Raw Sample Received by: RJ (EPA-106)
Raw Sample Relinquished by: RJ (EPA-106)

Date/Time 05/30/25 9:10
Raw Sample Received by: RJ (EPA-106)
Raw Sample Relinquished by: RJ (EPA-106)

Prep Standard - Chemical Standard Summary

Order ID : Q2150

Test : PCB

Prepbatch ID : PB168208,

Sequence ID/Qc Batch ID: pp053025,

Standard ID :

EP2610,EP2613,EP2614,PP24329,PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369,PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387,PP24461,PP24597,

Chemical ID :

E2865,E3551,E3804,E3877,E3917,E3932,E3933,E3938,M6157,P 11522,P12699,P12702,P12931,P12936,P12949,P12955,P12957,P13356,P13357,P13373,P13381,P13589,P13591,P13697,P13702,P13830,P13878,P13883,W 3112,W3177,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2613	05/09/2025	11/05/2025	RUPESHKUMAR SHAH	None	None	Riteshkumar Patel 05/09/2025
<u>FROM</u> 4000.00000ml of E3932 + 4000.00000ml of E3933 = Final Quantity: 8000.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2614	05/19/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/19/2025
FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025
<u>FROM</u> 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml								

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24331	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24332	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24333	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24334	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24332 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24335	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13702 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24336	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24337	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24335 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24338	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24339	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24337 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24340	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13878 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24341	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24342	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24340 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24343	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24344	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24342 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24345	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12931 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24346	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.75000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24347	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24345 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24348	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24349	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24347 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24350	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12936 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24351	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24352	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24350 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24353	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24354	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24352 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24355	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13830 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24356	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24357	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24355 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24358	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24359	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24357 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24360	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13883 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24361	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24362	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24360 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24363	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24364	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24362 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24365	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13381 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24366	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24367	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24365 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24368	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24369	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24367 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24370	03/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12949 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24371	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 98.50000ml of W3177 + 0.50000ml of PP24329 + 1.00000ml of PP24370 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24372	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24371 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24373	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13373 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1886	AR1221 500 PPB ICV	PP24374	03/18/2025	08/12/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of E3877 + 0.50000ml of W3177 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24375	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12699 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24376	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24375 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24377	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13589 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24378	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24377 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24379	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13591 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24380	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24379 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24382	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24384	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12702 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24385	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24384 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24386	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P11522 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24387	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24386 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24461	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel
								04/16/2025

FROM 0.50000ml of P12955 + 99.50000ml of E3917 = Final Quantity: 100.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24597	05/20/2025	11/05/2025	Abdul Mirza	None	None	Yogesh Patel 05/22/2025
<u>FROM</u>	1.00000ml of P13357 + 999.00000ml of E3932 = Final Quantity: 1000.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	0000243821	06/30/2025	04/30/2020 / RAJESH	04/28/2020 / RAJESH	E2865

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	11/05/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3917

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3932

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	11/05/2025	05/05/2025 / RUPESH	04/23/2025 / RUPESH	E3933

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	04/30/2026	/	05/14/2025 / RUPESH	E3938

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	91867 / Aroclor 1232 100 ug/mL	020823	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12699

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12702

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	a0203672	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12931

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	a0202803	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12936

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	09/18/2025	03/18/2025 / yogesh	12/20/2023 / Yogesh	P12949

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	10/11/2025	04/11/2025 / Abdul	12/20/2023 / Yogesh	P12955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	11/20/2025	05/20/2025 / Abdul	04/22/2024 / Abdul	P13357

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0207475	09/18/2025	03/18/2025 / yogesh	05/03/2024 / Abdul	P13381

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13589

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13591

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0215270	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0217391	09/18/2025	03/18/2025 / yogesh	12/09/2024 / Ankita	P13830

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0220950	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13883

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Sand
Purified
Washed and Ignited



Material No.: 3382-05
Batch No.: 0000243821
Manufactured Date: 2018/04/09
Retest Date: 2025/04/07
Revision No: 1


Certificate of Analysis

Test	Specification	Result
Substances Soluble in HCl	$\leq 0.16\%$	0.01

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC

E 2865


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
------	---------------	--------

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

Harout Sahagian **E3877**

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

Avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

RS

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3932

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantorsTM**



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E3933



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors**



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E3938



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



MG157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantors**™



Material No.: 9673-33

Batch No.: 2411262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak

Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED WEIGHT REPORT

Part Number: **91867**
 Lot Number: **020823**
 Description: **WP 037 - Aroclor 1232**
PCB Technical Mixture
 Expiration Date: **020833**
 Recommended Storage: **Ambient (20 °C)**
 Nominal Concentration ($\mu\text{g/mL}$): **100**
 NIST Test ID#: **6UTB**

Solvent(
Aceton


Weight(s) shown below were combined and diluted to (mL): **100.0**
 5E-05 Balance Uncertainty
 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc ($\mu\text{g/mL}$)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
 Column ID SPB-608 30 meter X 0.53mm X 5 μm film thickness
 Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
 Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
 Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
 Rate = 8°C/min, Total run time = 35 min
 Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
 Standard injection = 1.5 μL , Range=3






110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32009 **Lot No.:** A0203672

Description: Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

p12928

→

p12932

AJ

12/27/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	—%	1,004.7 µg/mL	+/- 55.7515

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Russ Boothamer

Russ Boothamer - Operations Technician I

Date Mixed: 26-Oct-2023

Balance Serial #

B442140311

Jennifer Polino

Jennifer Polino - Operations Tech III - ARM QC

Date Passed: 06-Nov-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32010 **Lot No.:** A0202803

Description: Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P1210967
P1210977
P1210973

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	13897600	---%	1,001.7 µg/mL	+/- 55.5850

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente
Laith Clemente - Operations Technician I

Date Mixed: 03-Oct-2023

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Solvent(s):
Hexane

Lot#
273615

Lot Number:
022023

Formulated By:
Benson Chan

DATE
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Reviewed By:
Pedro L. Rentas

DATE
022023

Expiration Date:
022033

Expanded
Uncertainty
(+/-) (µg/mL)

4.0

Recommended Storage:
Ambient (20 °C)

Actual
Conc (µg/mL)

1002.8

Nominal Concentration (µg/mL):
1000

OSHA PEL (TWA)

LD50

NIST Test ID#:
6UTB

Actual
Weight(g)

0.20060

Weight(s) shown below were combined and diluted to (mL):
200.0

Conc (µg/mL)

1000

5E-05 Balance Uncertainty

Purity

0.2

0.010 Flask Uncertainty


(%)

100

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Expanded Uncertainty (+/-) (µg/mL)	Actual Conc (µg/mL)	Actual Weight(g)	Actual Weight(g)	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	4.0	1002.8	0.20060	1002.8	12674-11-2	N/A
2. Aroclor 1260	21	020491JC	1000	100	4.0	1003.9	0.20081	1003.9	11086-82-5	0.5mg/m3 orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number: 20084
Lot Number: 022023
Description: CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260
022033
Expiration Date: Ambient (20 °C)
Recommended Storage: 1000
Nominal Concentration (µg/mL): 6UTB
NIST Test ID#: 200.0

Solvent(s): Hexane
Lot#: 273615

Formulated By: Benson Chan
Reviewed By: Pedro L. Rentas

022023
DATE

022023
DATE


5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Weight(s) shown below were combined and diluted to (mL): 200.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





CERTIFIED WEIGHT REPORT

Part Number:
Lot Number:
Description:

99139
121823
Aroclor 1254

Solvent(s):
Iso-octane

Lot#
82227

Expiration Date:
Recommended Storage:
Nominal Concentration (µg/mL):
NIST Test ID#:

121833
Ambient (20 °C)
100
6UTB

Volume(s) shown below were combined and diluted to (mL):
Note: Aroclor 1254 is a mix of isomers.

5E-05 Balance Uncertainty
0.003 Flask Uncertainty

20.0 Dilution Factor

Compound	Part Number	Lot Number
----------	-------------	------------

Initial Volume (mL)	Initial Concentration (µg/mL)	Final Concentration (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
---------------------	-------------------------------	-----------------------------	------------------------------------	------	----------------	------

Formulated By: Anthony Mahoney
Reviewed By: Pedro L. Rentas


P12956 / Y.P.
P12957 / Z.P.

1. Aroclor 1254	79100	121823	0.10	2.00	0.017	1003.3	100.1	1.8	11097-69-1	0.5mg/m3 (skin)	orl-rat 1295mg/kg
-----------------	-------	--------	------	------	-------	--------	-------	-----	------------	-----------------	-------------------

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 250 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = Etdaq Channel 1
Standard injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
----------	------	---------------





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1221 Standard
Product Number: PP-292-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006783205
Lot Issue Date: 20-Feb-2024
Expiration Date: 31-Mar-2032

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1221	100.3 ±	0.5 µg/mL	011104-28-2	NT01017

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13342
AJ
05106124
P13343

Page: 1 of 2

CSD-QA-015.2

ISO 17025

Cert No. AT-1937


250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0207475
Description : Aroclor® 1268 Standard
Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : May 31, 2030 **Storage:** 25°C nominal
Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	10947000	---%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13380
↓
P 13381
↓
SAZU
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 06-Feb-2024

Balance Serial # B442140311

Dillon Murphy
Dillon Murphy - Operations Technician I


Date Passed: 09-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13380 } (2)
↓
P13381
1

[Signature]
05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 95T215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:



Monica Bourgeois
QMS Representative



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0215270

Description : Aroclor® 1221 Standard
Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13902 } Y.P.
P13703 } 10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	14969200	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 16-Aug-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 20-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024

Balance Serial #

C322230531

Brittany Federhik

Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32409 **Lot No.:** A0220950

Description: Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,002.0 µg/mL	+/- 55.6035

Solvent: Hexane

CAS # 110-54-3

Purity 99%

P13882

↓

AJ
01/28/25

P13883

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckal - Mix Technician

Date Mixed: 09-Jan-2025 **Balance Serial #** C322230531



Britiany Federinko - Operations Tech I

Date Passed: 14-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: CDM SMITH
ADDRESS: 110 FIELDREST AVE #8 6TH FLOOR
CITY EDISON STATE: NJ ZIP: 08837
ATTENTION: MARCIE ENCINAS
PHONE: 732-590-4679 FAX: 732-225-7851

CLIENT PROJECT INFORMATION

PROJECT NAME: SOUTH RIVER WM REPAIRMENT
PROJECT NO. 302781 LOCATION: SOUTH RIVER, NJ
PROJECT MANAGER: MARCIE ENCINAS
e-mail: ENCINAS.MA@CDMSMITH.COM
PHONE: 732-590-4679 FAX: 732-225-7851

CLIENT BILLING INFORMATION

BILL TO: CDM SMITH PO#:
ADDRESS: 110 FIELDREST AVE #8 6TH FLOOR
CITY EDISON STATE: NJ ZIP: 08837
ATTENTION: MARCIE ENCINAS PHONE: 732-590-4679

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☒ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1. TCL VOC 2. TCL SVOC 3. METALS 4. PCB 5. PESTICIDE 6. HERBICIDE 7. DRO/GRO 8. 9.

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										COMMENTS ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	TP-44	SOIL		X	5/27/25	0820	6	X	X	X	X	X	X	X			E
2.	TP-42	SOIL		X	5/27/25	0915	6	X	X	X	X	X	X	X			E
3.	TP-39	SOIL		X	5/27/25	0950	6	X	X	X	X	X	X	X			E
4.	TP-48	SOIL		X	5/27/25	1220	6	X	X	X	X	X	X	X			E
5.	TP-47	SOIL		X	5/27/25	1120	6	X	X	X	X	X	X	X			E
6.	TP-50	SOIL		X	5/27/25	1330	6	X	X	X	X	X	X	X			E
7.	TP-51	SOIL		X	5/27/25	1505	6	X	X	X	X	X	X	X			E
8.	TP-52	SOIL		X	5/28/25	0850	6	X	X	X	X	X	X	X			E
9.	TP-54	SOIL		X	5/28/25	1145	6	X	X	X	X	X	X	X			E
10.	TP-53	SOIL		X	5/28/25	1425	6	X	X	X	X	X	X	X			F

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. [Signature]	DATE/TIME: 5/28/25 1850	RECEIVED BY: 1. [Signature]	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP 3.5 °C
RELINQUISHED BY SAMPLER: 2. [Signature]	DATE/TIME:	RECEIVED BY: 2. [Signature]	Comments:
RELINQUISHED BY SAMPLER: 3. [Signature]	DATE/TIME: 1645 5-28-25	RECEIVED BY: 3. [Signature]	Page ____ of

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2150	CAMP02	Order Date : 5/28/2025 3:53:00 PM	Project Mgr :
Client Name : CDM Smith		Project Name : South River WM Replacem	Report Type : Level 1
Client Contact : Marcie Ann Encinas		Receive DateTime : 5/28/2025 12:00:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : CDM Smith		Purchase Order : 18:45	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2150-01	TP-44	Solid	05/27/2025	08:20					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q2150-02	TP-42	Solid	05/27/2025	09:15					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q2150-03	TP-39	Solid	05/27/2025	09:50					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q2150-04	TP-48	Solid	05/27/2025	12:20					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q2150-05	TP-47	Solid	05/27/2025	11:20					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q2150-06	TP-50	Solid	05/27/2025	13:30					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q2150-07	TP-51	Solid	05/27/2025	15:05					
					VOC-TCLVOA-10		8260D		10 Bus. Days
Q2150-08	TP-52	Solid	05/28/2025	08:50					


LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2150	CAMP02	Order Date : 5/28/2025 3:53:00 PM	Project Mgr :
Client Name : CDM Smith		Project Name : South River WM Replacem	Report Type : Level 1
Client Contact : Marcie Ann Encinas		Receive DateTime : 5/28/2025 12:00:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : CDM Smith		Purchase Order : 16:45	Hard Copy Date :
Invoice Contact : Marcie Ann Encinas			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2150-09	TP-54	Solid	05/28/2025	11:45	VOC-TCLVOA-10		8260D		10 Bus. Days
Q2150-10	TP-53	Solid	05/28/2025	14:25	VOC-TCLVOA-10		8260D		10 Bus. Days
					VOC-TCLVOA-10		8260D		10 Bus. Days

Relinquished By : 

Date / Time : 5/29/25 0950

Received By : 

Date / Time : 05/29/25 0950

Storage Area : VOA Refridgerator Room

12 ft # 6
ft 2