

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Instrument ID: ECD_L

Calibration Date(s): 05/21/2025 05/21/2025

Calibration Times: 11:35 12:29

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL095735.D</u>	CF 075 =	<u>PL095736.D</u>		
CF 050 =		<u>PL095737.D</u>	CF 025 =	<u>PL095738.D</u>	CF 005 =	<u>PL095739.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3029170000	2897950000	2890660000	2805810000	3029440000	2930610000	3
4,4'-DDE	3753620000	3658270000	3631690000	3489800000	3806130000	3667900000	3
4,4'-DDT	2712830000	2654360000	2650610000	2590070000	2917840000	2705140000	5
Aldrin	4347110000	4211420000	4211170000	4079190000	4539770000	4277730000	4
alpha-BHC	5055140000	4887150000	4799900000	4583430000	4914010000	4847930000	4
alpha-Chlordane	3936090000	3857140000	3869930000	3816090000	4261610000	3948170000	5
beta-BHC	1911140000	1892640000	1930210000	1931530000	2191310000	1971370000	6
Decachlorobiphenyl	2296450000	2273040000	2304350000	2354860000	2552990000	2356340000	5
delta-BHC	4510360000	4418490000	4383980000	4233750000	4621480000	4433610000	3
Dieldrin	3903950000	3813000000	3813820000	3708570000	4056580000	3859180000	3
Endosulfan I	3625480000	3571730000	3593970000	3550090000	3982900000	3664830000	5
Endosulfan II	3313830000	3256160000	3325550000	3384990000	3949720000	3446050000	8
Endosulfan sulfate	2925470000	2895950000	2926550000	2906380000	3280250000	2986920000	6
Endrin	3186950000	3188700000	3185310000	3116310000	3456510000	3226750000	4
Endrin aldehyde	2356490000	2323760000	2365700000	2353750000	2688400000	2417620000	6
Endrin ketone	3179290000	3121440000	3135900000	3068650000	3329290000	3166910000	3
gamma-BHC (Lindane)	4588190000	4447360000	4400770000	4245410000	4677520000	4471850000	4
gamma-Chlordane	3946420000	3852300000	3845310000	3754890000	4061530000	3892090000	3
Heptachlor	3804040000	3718580000	3721520000	3622900000	4058170000	3785040000	4
Heptachlor epoxide	3727710000	3692000000	3731080000	3682110000	4251420000	3816860000	6
Methoxychlor	1252650000	1237700000	1261470000	1251380000	1373430000	1275330000	4
Tetrachloro-m-xylene	3099670000	3048860000	3065820000	3062420000	3499570000	3155270000	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Instrument ID: ECD_L

Calibration Date(s): 05/21/2025 05/21/2025

Calibration Times: 11:35 12:29

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL095735.D</u>	CF 075 =	<u>PL095736.D</u>		
CF 050 =		<u>PL095737.D</u>	CF 025 =	<u>PL095738.D</u>	CF 005 =	<u>PL095739.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	4445580000	4366880000	4404380000	4229840000	4477970000	4384930000	2
4,4'-DDE	5328200000	5286620000	5342960000	5185270000	5667730000	5362160000	3
4,4'-DDT	4834370000	4768870000	4805750000	4636730000	4867260000	4782600000	2
Aldrin	5413270000	5337420000	5359250000	5094120000	5331460000	5307100000	2
alpha-BHC	6120380000	5984780000	5956690000	5567970000	5656920000	5857350000	4
alpha-Chlordane	5212180000	5117720000	5169640000	5059150000	5455120000	5202760000	3
beta-BHC	2433740000	2431120000	2471290000	2436550000	2615060000	2477550000	3
Decachlorobiphenyl	4307130000	4293010000	4424120000	4407270000	4440750000	4374460000	2
delta-BHC	5855570000	5710500000	5700250000	5411210000	5559290000	5647370000	3
Dieldrin	5312990000	5256750000	5296490000	5103800000	5530570000	5300120000	3
Endosulfan I	4730980000	4708290000	4767730000	4668460000	4949420000	4764970000	2
Endosulfan II	4687260000	4644550000	4721900000	4610060000	5095140000	4751780000	4
Endosulfan sulfate	4396050000	4376160000	4475180000	4410280000	4816080000	4494750000	4
Endrin	4781360000	4775420000	4841120000	4733510000	5249090000	4876100000	4
Endrin aldehyde	3353000000	3342650000	3418100000	3365440000	3792270000	3454290000	6
Endrin ketone	5025560000	5055710000	5163390000	5096070000	5528620000	5173870000	4
gamma-BHC (Lindane)	5738840000	5640150000	5652210000	5354130000	5620800000	5601220000	3
gamma-Chlordane	5300820000	5202550000	5232880000	5068800000	5458400000	5252690000	3
Heptachlor	5660130000	5568750000	5611780000	5421260000	5783370000	5609060000	2
Heptachlor epoxide	4854250000	4848530000	4943810000	4810760000	5229780000	4937430000	3
Methoxychlor	2503630000	2513740000	2592490000	2627630000	2843040000	2616110000	5
Tetrachloro-m-xylene	3908270000	3796990000	3846980000	3798040000	4219390000	3913940000	5