

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 05/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL095839.D Time Analyzed: 12:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.742	6.634	6.834	52.630	50.000	5.3
4,4'-DDE	6.232	6.124	6.324	48.800	50.000	-2.4
4,4'-DDT	7.058	6.950	7.150	43.650	50.000	-12.7
Aldrin	5.304	5.197	5.397	54.070	50.000	8.1
alpha-BHC	4.031	3.925	4.125	56.630	50.000	13.3
alpha-Chlordane	6.061	5.954	6.154	52.450	50.000	4.9
beta-BHC	4.550	4.444	4.644	54.950	50.000	9.9
Decachlorobiphenyl	9.114	9.003	9.203	46.650	50.000	-6.7
delta-BHC	4.798	4.692	4.892	55.280	50.000	10.6
Dieldrin	6.382	6.275	6.475	51.990	50.000	4.0
Endosulfan I	6.109	6.002	6.202	52.340	50.000	4.7
Endosulfan II	6.823	6.715	6.915	49.290	50.000	-1.4
Endosulfan sulfate	7.186	7.078	7.278	49.150	50.000	-1.7
Endrin	6.610	6.502	6.702	45.300	50.000	-9.4
Endrin aldehyde	6.952	6.844	7.044	48.220	50.000	-3.6
Endrin ketone	7.667	7.560	7.760	50.720	50.000	1.4
gamma-BHC (Lindane)	4.362	4.257	4.457	56.780	50.000	13.6
gamma-Chlordane	5.980	5.873	6.073	53.770	50.000	7.5
Heptachlor	4.961	4.855	5.055	50.260	50.000	0.5
Heptachlor epoxide	5.724	5.618	5.818	50.040	50.000	0.1
Methoxychlor	7.531	7.423	7.623	43.480	50.000	-13.0
Tetrachloro-m-xylene	3.578	3.473	3.673	56.560	50.000	13.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL01 Date Analyzed: 05/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL095839.D Time Analyzed: 12:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.923	5.826	6.026	55.200	50.000	10.4
4,4'-DDE	5.371	5.274	5.474	47.740	50.000	-4.5
4,4'-DDT	6.178	6.080	6.280	41.880	50.000	-16.2
Aldrin	4.363	4.268	4.468	56.630	50.000	13.3
alpha-BHC	3.391	3.296	3.496	57.910	50.000	15.8
alpha-Chlordane	5.184	5.087	5.287	54.420	50.000	8.8
beta-BHC	4.021	3.926	4.126	55.690	50.000	11.4
Decachlorobiphenyl	8.062	7.964	8.164	43.500	50.000	-13.0
delta-BHC	4.257	4.161	4.361	55.940	50.000	11.9
Dieldrin	5.504	5.408	5.608	53.480	50.000	7.0
Endosulfan I	5.239	5.143	5.343	54.970	50.000	9.9
Endosulfan II	6.071	5.973	6.173	53.480	50.000	7.0
Endosulfan sulfate	6.472	6.375	6.575	51.060	50.000	2.1
Endrin	5.780	5.683	5.883	43.470	50.000	-13.1
Endrin aldehyde	6.249	6.151	6.351	51.400	50.000	2.8
Endrin ketone	6.978	6.880	7.080	50.750	50.000	1.5
gamma-BHC (Lindane)	3.725	3.630	3.830	56.550	50.000	13.1
gamma-Chlordane	5.119	5.023	5.223	55.410	50.000	10.8
Heptachlor	4.078	3.983	4.183	52.900	50.000	5.8
Heptachlor epoxide	4.866	4.771	4.971	57.090	50.000	14.2
Methoxychlor	6.748	6.651	6.851	41.270	50.000	-17.5
Tetrachloro-m-xylene	2.881	2.786	2.986	56.350	50.000	12.7