

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 05/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL095859.D Time Analyzed: 18:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.735	6.634	6.834	50.550	50.000	1.1
4,4'-DDE	6.225	6.124	6.324	51.080	50.000	2.2
4,4'-DDT	7.050	6.950	7.150	53.550	50.000	7.1
Aldrin	5.297	5.197	5.397	51.670	50.000	3.3
alpha-BHC	4.025	3.925	4.125	54.220	50.000	8.4
alpha-Chlordane	6.055	5.954	6.154	50.700	50.000	1.4
beta-BHC	4.544	4.444	4.644	52.730	50.000	5.5
Decachlorobiphenyl	9.103	9.003	9.203	51.680	50.000	3.4
delta-BHC	4.792	4.692	4.892	48.340	50.000	-3.3
Dieldrin	6.375	6.275	6.475	51.390	50.000	2.8
Endosulfan I	6.102	6.002	6.202	50.990	50.000	2.0
Endosulfan II	6.816	6.715	6.915	54.940	50.000	9.9
Endosulfan sulfate	7.179	7.078	7.278	51.290	50.000	2.6
Endrin	6.603	6.502	6.702	51.420	50.000	2.8
Endrin aldehyde	6.945	6.844	7.044	51.330	50.000	2.7
Endrin ketone	7.660	7.560	7.760	51.090	50.000	2.2
gamma-BHC (Lindane)	4.356	4.257	4.457	54.370	50.000	8.7
gamma-Chlordane	5.973	5.873	6.073	50.760	50.000	1.5
Heptachlor	4.955	4.855	5.055	52.490	50.000	5.0
Heptachlor epoxide	5.719	5.618	5.818	49.240	50.000	-1.5
Methoxychlor	7.523	7.423	7.623	54.700	50.000	9.4
Tetrachloro-m-xylene	3.572	3.473	3.673	53.520	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/21/2025 05/21/2025

Client Sample No.: CCAL02 Date Analyzed: 05/30/2025

Lab Sample No.: PSTDCCC050 Data File : PL095859.D Time Analyzed: 18:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.925	5.826	6.026	49.660	50.000	-0.7
4,4'-DDE	5.373	5.274	5.474	50.560	50.000	1.1
4,4'-DDT	6.179	6.080	6.280	48.750	50.000	-2.5
Aldrin	4.366	4.268	4.468	53.760	50.000	7.5
alpha-BHC	3.395	3.296	3.496	54.870	50.000	9.7
alpha-Chlordane	5.186	5.087	5.287	51.750	50.000	3.5
beta-BHC	4.024	3.926	4.126	52.720	50.000	5.4
Decachlorobiphenyl	8.063	7.964	8.164	46.490	50.000	-7.0
delta-BHC	4.260	4.161	4.361	53.340	50.000	6.7
Dieldrin	5.507	5.408	5.608	50.790	50.000	1.6
Endosulfan I	5.242	5.143	5.343	51.710	50.000	3.4
Endosulfan II	6.072	5.973	6.173	49.930	50.000	-0.1
Endosulfan sulfate	6.474	6.375	6.575	48.620	50.000	-2.8
Endrin	5.782	5.683	5.883	51.010	50.000	2.0
Endrin aldehyde	6.250	6.151	6.351	49.370	50.000	-1.3
Endrin ketone	6.979	6.880	7.080	45.180	50.000	-9.6
gamma-BHC (Lindane)	3.728	3.630	3.830	53.950	50.000	7.9
gamma-Chlordane	5.122	5.023	5.223	51.880	50.000	3.8
Heptachlor	4.081	3.983	4.183	53.780	50.000	7.6
Heptachlor epoxide	4.869	4.771	4.971	52.620	50.000	5.2
Methoxychlor	6.750	6.651	6.851	47.570	50.000	-4.9
Tetrachloro-m-xylene	2.885	2.786	2.986	53.470	50.000	6.9