

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG No.: Q2150

Instrument ID: BNA_F Calibration Date/Time: 05/30/2025 13:20

Lab File ID: BF142566.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.187	1.142		-3.8	
Phenol-d6	1.429	1.395		-2.4	
Nitrobenzene-d5	0.367	0.349		-4.9	
2-Fluorobiphenyl	1.490	1.348		-9.5	
2,4,6-Tribromophenol	0.222	0.215		-3.2	
Terphenyl-d14	1.464	1.377		-5.9	

All other compounds must meet a minimum RRF of 0.010.