

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG No.: Q2150

Instrument ID: BNA_F Calibration Date/Time: 06/02/2025 11:30

Lab File ID: BF142590.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.187	1.153		-2.9	
Phenol-d6	1.429	1.385		-3.1	
Nitrobenzene-d5	0.367	0.347		-5.4	
2-Fluorobiphenyl	1.490	1.341		-10.0	
2,4,6-Tribromophenol	0.222	0.212		-4.5	
Terphenyl-d14	1.464	1.327		-9.4	

All other compounds must meet a minimum RRF of 0.010.