

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060225\
 Data File : BF142590.D
 Acq On : 02 Jun 2025 11:30
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 11:53:54 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 20 16:26:47 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.475	0.464	2.3	117	0.00
3	Pyridine	1.210	1.210	0.0	120	0.00
4	n-Nitrosodimethylamine	0.628	0.595	5.3	111	-0.02
5 S	2-Fluorophenol	1.187	1.153	2.9	118	0.00
6	Aniline	1.921	1.892	1.5	119	0.00
7 S	Phenol-d6	1.429	1.385	3.1	117	-0.01
8	2-Chlorophenol	1.293	1.260	2.6	116	0.00
9	Benzaldehyde	0.859	0.802	6.6	114	0.00
10 C	Phenol	1.608	1.558	3.1	118	-0.01
11	bis(2-Chloroethyl)ether	1.157	1.124	2.9	117	0.00
12	1,3-Dichlorobenzene	1.443	1.385	4.0	115	0.00
13 C	1,4-Dichlorobenzene	1.453	1.412	2.8	116	0.00
14	1,2-Dichlorobenzene	1.391	1.341	3.6	117	0.00
15	Benzyl Alcohol	1.056	1.016	3.8	115	0.00
16	2,2'-oxybis(1-Chloropropane	1.944	1.872	3.7	116	0.00
17	2-Methylphenol	1.010	0.984	2.6	117	0.00
18	Hexachloroethane	0.507	0.483	4.7	114	0.00
19 P	n-Nitroso-di-n-propylamine	0.886	0.829	6.4	114	-0.02
20	3+4-Methylphenols	1.293	1.242	3.9	115	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
22	Acetophenone	0.443	0.417	5.9	114	0.00
23 S	Nitrobenzene-d5	0.367	0.347	5.4	114	-0.01
24	Nitrobenzene	0.331	0.314	5.1	114	-0.01
25	Isophorone	0.613	0.575	6.2	114	-0.01
26 C	2-Nitrophenol	0.177	0.176	0.6	118	0.00
27	2,4-Dimethylphenol	0.312	0.299	4.2	116	0.00
28	bis(2-Chloroethoxy)methane	0.383	0.365	4.7	117	-0.01
29 C	2,4-Dichlorophenol	0.283	0.275	2.8	117	0.00
30	1,2,4-Trichlorobenzene	0.308	0.296	3.9	117	0.00
31	Naphthalene	0.985	0.939	4.7	117	0.00
32	Benzoic acid	0.190	0.181	4.7	113	-0.04
33	4-Chloroaniline	0.399	0.389	2.5	117	0.00
34 C	Hexachlorobutadiene	0.192	0.175	8.9	110	0.00
35	Caprolactam	0.080	0.084	-5.0	124	-0.02
36 C	4-Chloro-3-methylphenol	0.291	0.275	5.5	114	-0.01
37	2-Methylnaphthalene	0.620	0.581	6.3	115	0.00
38	1-Methylnaphthalene	0.641	0.600	6.4	114	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
40	1,2,4,5-Tetrachlorobenzene	0.574	0.542	5.6	115	0.00
41 P	Hexachlorocyclopentadiene	0.387	0.340	12.1	101	0.00
42 S	2,4,6-Tribromophenol	0.222	0.212	4.5	117	0.00
43 C	2,4,6-Trichlorophenol	0.387	0.366	5.4	115	0.00
44	2,4,5-Trichlorophenol	0.408	0.396	2.9	117	0.00
45 S	2-Fluorobiphenyl	1.490	1.341	10.0	113	-0.01
46	1,1'-Biphenyl	1.552	1.455	6.2	115	0.00
47	2-Chloronaphthalene	1.146	1.080	5.8	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060225\
 Data File : BF142590.D
 Acq On : 02 Jun 2025 11:30
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 02 11:53:54 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 20 16:26:47 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.331	0.325	1.8	117	0.00
49	Acenaphthylene	1.936	1.830	5.5	116	0.00
50	Dimethylphthalate	1.320	1.263	4.3	118	-0.01
51	2,6-Dinitrotoluene	0.285	0.279	2.1	116	0.00
52 C	Acenaphthene	1.182	1.118	5.4	117	0.00
53	3-Nitroaniline	0.311	0.315	-1.3	123	0.00
54 P	2,4-Dinitrophenol	0.147	0.143	2.7	112	0.00
55	Dibenzofuran	1.701	1.606	5.6	117	0.00
56 P	4-Nitrophenol	0.230	0.227	1.3	117	0.00
57	2,4-Dinitrotoluene	0.372	0.374	-0.5	120	0.00
58	Fluorene	1.314	1.257	4.3	119	0.00
59	2,3,4,6-Tetrachlorophenol	0.341	0.334	2.1	117	0.00
60	Diethylphthalate	1.289	1.226	4.9	116	-0.01
61	4-Chlorophenyl-phenylether	0.646	0.612	5.3	118	0.00
62	4-Nitroaniline	0.282	0.295	-4.6	125	-0.01
63	Azobenzene	1.158	1.107	4.4	117	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
65	4,6-Dinitro-2-methylphenol	0.112	0.113	-0.9	119	0.00
66 c	n-Nitrosodiphenylamine	0.684	0.641	6.3	120	0.00
67	4-Bromophenyl-phenylether	0.236	0.220	6.8	120	0.00
68	Hexachlorobenzene	0.262	0.246	6.1	119	0.00
69	Atrazine	0.182	0.184	-1.1	126	0.00
70 C	Pentachlorophenol	0.148	0.139	6.1	115	0.00
71	Phenanthrene	1.069	1.016	5.0	122	0.00
72	Anthracene	1.091	1.045	4.2	122	0.00
73	Carbazole	0.933	0.918	1.6	125	0.00
74	Di-n-butylphthalate	1.021	1.018	0.3	126	0.00
75 C	Fluoranthene	0.999	0.991	0.8	128	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	136	0.00
77	Benzidine	0.744	0.809	-8.7	138	0.00
78	Pyrene	1.866	1.760	5.7	129	0.00
79 S	Terphenyl-d14	1.464	1.327	9.4	127	0.00
80	Butylbenzylphthalate	0.516	0.546	-5.8	141	0.00
81	Benzo(a)anthracene	1.336	1.272	4.8	136	0.00
82	3,3'-Dichlorobenzidine	0.405	0.425	-4.9	142	0.00
83	Chrysene	1.200	1.136	5.3	135	0.00
84	Bis(2-ethylhexyl)phthalate	0.660	0.694	-5.2	141	0.00
85 c	Di-n-octyl phthalate	1.290	1.195	7.4	129	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	126	0.00
87	Indeno(1,2,3-cd)pyrene	1.501	1.368	8.9	119	0.00
88	Benzo(b)fluoranthene	1.184	1.126	4.9	126	0.00
89	Benzo(k)fluoranthene	1.109	1.039	6.3	126	0.00
90 C	Benzo(a)pyrene	1.116	1.071	4.0	125	0.00
91	Dibenzo(a,h)anthracene	1.218	1.126	7.6	120	-0.01
92	Benzo(g,h,i)perylene	1.218	1.105	9.3	118	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF060225\
Data File : BF142590.D
Acq On : 02 Jun 2025 11:30
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jun 02 11:53:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF052025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 20 16:26:47 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0