

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/02/2025

Lab File ID: BF142590.D Time Analyzed: 11:30

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	112830	6.898	441878	8.19	241388	9.94
UPPER LIMIT	225660	7.398	883756	8.686	482776	10.439
LOWER LIMIT	56415	6.398	220939	7.686	120694	9.439
EPA SAMPLE NO.						
01 TP-44DL	127289	6.90	470532	8.18	232054	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.