

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/02/2025

Lab File ID: BF142590.D Time Analyzed: 11:30

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	415282	11.427	230387	14.074	238851	15.568
UPPER LIMIT	830564	11.927	460774	14.574	477702	16.068
LOWER LIMIT	207641	10.927	115194	13.574	119426	15.068
EPA SAMPLE NO.						
01 TP-44DL	325882	11.43	250994	14.07	278926	15.57

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.