

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG No.: Q2150

Instrument ID: MSVOA_Y Calibration Date(s): 05/27/2025 05/27/2025

Heated Purge: (Y/N) Y Calibration Time(s): 08:50 10:44

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF005 = VY022421.D		RRF010 = VY022422.D		RRF020 = VY022423.D		RRF050 = VY022424.D		RRF100 = VY022425.D		RRF150 = VY022426.D	
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD				
Dichlorodifluoromethane	0.544	0.509	0.526	0.423	0.430	0.429	0.477	11.6				
Chloromethane	1.364	1.331	1.300	1.196	0.936	0.912	1.173	17.1				
Vinyl Chloride	1.535	1.493	1.522	1.453	1.292	1.257	1.425	8.5				
Bromomethane	1.261	1.281	1.397	1.275	1.213	1.345	1.295	5.1				
Chloroethane	0.930	0.881	0.949	1.015	0.907	0.907	0.932	5				
Trichlorofluoromethane	1.243	1.174	1.222	1.197	1.187	1.193	1.203	2.1				
1,1,2-Trichlorotrifluoroethane	0.552	0.517	0.541	0.520	0.523	0.512	0.528	2.9				
1,1-Dichloroethene	0.552	0.512	0.533	0.511	0.520	0.515	0.524	3				
Acetone	0.094	0.090	0.081	0.085	0.086	0.083	0.087	5.6				
Carbon Disulfide	1.672	1.675	1.707	1.653	1.672	1.650	1.672	1.2				
Methyl tert-butyl Ether	1.353	1.359	1.357	1.386	1.445	1.445	1.391	3.1				
Methyl Acetate	0.288	0.319	0.296	0.308	0.351	0.383	0.324	11.2				
Methylene Chloride	0.778	0.669	0.578	0.549	0.550	0.543	0.611	15.5				
trans-1,2-Dichloroethene	0.595	0.556	0.555	0.561	0.582	0.582	0.572	2.9				
1,1-Dichloroethane	1.000	0.986	1.035	1.027	1.062	1.056	1.028	2.9				
Cyclohexane	1.220	1.044	0.997	0.939	0.947	0.931	1.013	10.9				
2-Butanone	0.137	0.144	0.136	0.142	0.152	0.154	0.144	5.2				
Carbon Tetrachloride	0.473	0.456	0.471	0.481	0.510	0.512	0.484	4.7				
cis-1,2-Dichloroethene	0.655	0.647	0.650	0.652	0.684	0.685	0.662	2.6				
Bromochloromethane	0.400	0.426	0.435	0.439	0.439	0.437	0.429	3.5				
Chloroform	0.966	0.970	1.023	1.026	1.052	1.038	1.013	3.5				
1,1,1-Trichloroethane	0.900	0.894	0.916	0.913	0.952	0.940	0.919	2.5				
Methylcyclohexane	0.611	0.602	0.603	0.614	0.638	0.636	0.618	2.6				
Benzene	1.308	1.312	1.360	1.391	1.463	1.461	1.383	5				
1,2-Dichloroethane	0.348	0.348	0.360	0.369	0.383	0.384	0.365	4.5				
Trichloroethene	0.334	0.325	0.344	0.343	0.357	0.347	0.342	3.3				
1,2-Dichloropropane	0.321	0.307	0.318	0.323	0.342	0.339	0.325	4				
Bromodichloromethane	0.452	0.449	0.449	0.464	0.493	0.494	0.467	4.5				
4-Methyl-2-Pentanone	0.185	0.197	0.198	0.218	0.237	0.242	0.213	10.8				
Toluene	0.782	0.805	0.842	0.868	0.929	0.957	0.864	7.9				

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG No.: Q2150
 Instrument ID: MSVOA_Y Calibration Date(s): 05/27/2025 05/27/2025
 Heated Purge: (Y/N) Y Calibration Time(s): 08:50 10:44
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY022421.D		RRF010 = VY022422.D		RRF020 = VY022423.D			
		RRF050 = VY022424.D		RRF100 = VY022425.D		RRF150 = VY022426.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.400	0.414	0.416	0.434	0.472	0.473	0.435	7.1	
cis-1,3-Dichloropropene	0.477	0.480	0.504	0.508	0.549	0.547	0.511	6.1	
1,1,2-Trichloroethane	0.219	0.216	0.226	0.229	0.243	0.243	0.229	5.1	
2-Hexanone	0.121	0.125	0.129	0.145	0.157	0.159	0.140	12	
Dibromochloromethane	0.258	0.274	0.284	0.302	0.320	0.321	0.293	8.7	
1,2-Dibromoethane	0.199	0.189	0.207	0.215	0.223	0.225	0.210	6.8	
Tetrachloroethene	0.433	0.417	0.433	0.429	0.429	0.413	0.426	2.1	
Chlorobenzene	1.020	1.000	1.051	1.067	1.132	1.127	1.066	5.1	
Ethyl Benzene	1.823	1.828	1.920	1.980	2.142	2.185	1.980	7.8	
m/p-Xylenes	0.698	0.660	0.723	0.756	0.825	0.841	0.751	9.5	
o-Xylene	0.654	0.637	0.673	0.709	0.768	0.792	0.705	8.9	
Styrene	1.037	1.020	1.117	1.197	1.317	1.359	1.175	12.1	
Bromoform	0.169	0.176	0.184	0.192	0.206	0.210	0.189	8.6	
Isopropylbenzene	3.863	3.741	3.946	3.871	4.105	4.086	3.935	3.6	
1,1,2,2-Tetrachloroethane	0.623	0.574	0.604	0.615	0.652	0.653	0.620	4.8	
1,3-Dichlorobenzene	1.587	1.558	1.674	1.695	1.836	1.874	1.704	7.5	
1,4-Dichlorobenzene	1.596	1.527	1.617	1.642	1.707	1.687	1.629	4	
1,2-Dichlorobenzene	1.383	1.332	1.431	1.416	1.495	1.482	1.423	4.3	
1,2-Dibromo-3-Chloropropane	0.103	0.093	0.096	0.093	0.093	0.096	0.096	3.8	
1,2,4-Trichlorobenzene	0.714	0.724	0.771	0.828	0.842	0.856	0.789	7.8	
1,2,3-Trichlorobenzene	0.575	0.554	0.644	0.684	0.703	0.722	0.647	10.7	
1,2-Dichloroethane-d4	0.507	0.525	0.517	0.532	0.541	0.538	0.527	2.5	
Dibromofluoromethane	0.287	0.281	0.283	0.290	0.309	0.308	0.293	4.3	
Toluene-d8	1.159	1.133	1.168	1.176	1.274	1.251	1.194	4.7	
4-Bromofluorobenzene	0.377	0.331	0.345	0.355	0.382	0.376	0.361	5.7	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.