

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
 Data File : VY022491.D
 Acq On : 02 Jun 2025 11:46
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
 Supervised By :Mahesh Dadoda 06/03/2025

Quant Time: Jun 03 02:52:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 02:50:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	201089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	353755	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	289394	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	127807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	10508	4.802	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	9.600%#		
35) Dibromofluoromethane	7.646	113	9323	4.477	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	8.960%#		
50) Toluene-d8	10.109	98	37729	4.482	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	8.960%#		
62) 4-Bromofluorobenzene	12.408	95	11988	4.697	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	9.400%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	8717	4.961	ug/l	99
3) Chloromethane	2.068	50	26535	5.119	ug/l	98
4) Vinyl Chloride	2.208	62	27572	4.544	ug/l	98
5) Bromomethane	2.598	94	27457	4.784	ug/l	98
6) Chloroethane	2.739	64	19566	4.710	ug/l	95
7) Trichlorofluoromethane	3.062	101	22825	4.428	ug/l	92
8) Diethyl Ether	3.464	74	5184	4.553	ug/l	92
9) 1,1,2-Trichlorotrifluo...	3.824	101	10259	4.744	ug/l	97
10) Methyl Iodide	4.007	142	9174	3.956	ug/l #	95
11) Tert butyl alcohol	4.891	59	3374	21.587	ug/l #	88
12) 1,1-Dichloroethene	3.799	96	9440	4.573	ug/l	97
13) Acrolein	3.671	56	5656	29.646	ug/l	95
14) Allyl chloride	4.391	41	15096	4.693	ug/l	98
15) Acrylonitrile	5.080	53	10667	22.026	ug/l	97
16) Acetone	3.885	43	13023	30.108	ug/l #	87
17) Carbon Disulfide	4.110	76	30222	4.586	ug/l	99
18) Methyl Acetate	4.397	43	5630	4.327	ug/l	96
19) Methyl tert-butyl Ether	5.122	73	26276	4.562	ug/l	97
20) Methylene Chloride	4.622	84	17312	7.168	ug/l	89
21) trans-1,2-Dichloroethene	5.116	96	10489	4.549	ug/l	86
22) Diisopropyl ether	6.037	45	32497	4.544	ug/l	95
23) Vinyl Acetate	5.970	43	91478	21.559	ug/l	98
24) 1,1-Dichloroethane	5.921	63	19460	4.603	ug/l	97
25) 2-Butanone	6.909	43	14423	22.605	ug/l	95
26) 2,2-Dichloropropane	6.896	77	16805	4.471	ug/l	99
27) cis-1,2-Dichloroethene	6.896	96	11906	4.467	ug/l	98
28) Bromochloromethane	7.250	49	9453	5.175	ug/l	100
29) Tetrahydrofuran	7.274	42	9377	22.781	ug/l	95
30) Chloroform	7.427	83	19302	4.620	ug/l	87
31) Cyclohexane	7.707	56	23482	5.519	ug/l #	90
32) 1,1,1-Trichloroethane	7.628	97	16476	4.413	ug/l	98
36) 1,1-Dichloropropene	7.841	75	14700	4.367	ug/l	99
37) Ethyl Acetate	6.994	43	6409m	4.326	ug/l	
38) Carbon Tetrachloride	7.823	117	14691	4.233	ug/l	95
39) Methylcyclohexane	9.109	83	20076	4.415	ug/l	95
40) Benzene	8.085	78	42926	4.326	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
 Data File : VY022491.D
 Acq On : 02 Jun 2025 11:46
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 06/03/2025
 Supervised By : Mahesh Dadoda 06/03/2025

Quant Time: Jun 03 02:52:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 02:50:07 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.232	41	4192	4.394	ug/l #	89
42) 1,2-Dichloroethane	8.164	62	11332	4.198	ug/l	99
43) Isopropyl Acetate	8.207	43	13330	4.117	ug/l	95
44) Trichloroethene	8.872	130	10404	4.292	ug/l	79
45) 1,2-Dichloropropane	9.146	63	9678	4.091	ug/l	100
46) Dibromomethane	9.244	93	5596	4.223	ug/l	95
47) Bromodichloromethane	9.433	83	13448	3.999	ug/l	96
48) Methyl methacrylate	9.225	41	6483	4.374	ug/l	91
49) 1,4-Dioxane	9.231	88	1356m	85.767	ug/l	
51) 4-Methyl-2-Pentanone	10.000	43	30969	19.273	ug/l	98
52) Toluene	10.170	92	25893	4.150	ug/l	99
53) t-1,3-Dichloropropene	10.396	75	12388	3.982	ug/l	91
54) cis-1,3-Dichloropropene	9.859	75	14966	4.103	ug/l	100
55) 1,1,2-Trichloroethane	10.579	97	7318	4.385	ug/l	98
56) Ethyl methacrylate	10.445	69	9092	3.751	ug/l	94
57) 1,3-Dichloropropane	10.719	76	12648	4.255	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.719	63	24214	22.409	ug/l	97
59) 2-Hexanone	10.768	43	21243	19.832	ug/l	94
60) Dibromochloromethane	10.914	129	8781	4.152	ug/l	94
61) 1,2-Dibromoethane	11.018	107	6146	3.991	ug/l	100
64) Tetrachloroethene	10.652	164	10863	4.417	ug/l	95
65) Chlorobenzene	11.444	112	27467	4.377	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.518	131	8422	3.959	ug/l	94
67) Ethyl Benzene	11.524	91	47883	4.098	ug/l	96
68) m/p-Xylenes	11.627	106	35645	8.060	ug/l	99
69) o-Xylene	11.956	106	16866	4.075	ug/l	95
70) Styrene	11.975	104	26313	3.840	ug/l	99
71) Bromoform	12.133	173	4649	4.159	ug/l #	97
73) Isopropylbenzene	12.255	105	44350	4.307	ug/l	100
74) N-amyl acetate	12.072	43	9632	3.699	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.505	83	7289	4.359	ug/l	96
76) 1,2,3-Trichloropropane	12.560	75	7478m	4.432	ug/l	
77) Bromobenzene	12.536	156	9241	4.234	ug/l	95
78) n-propylbenzene	12.597	91	56543	4.511	ug/l	98
79) 2-Chlorotoluene	12.682	91	30387	4.591	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	35122	4.362	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	2596	4.411	ug/l	96
82) 4-Chlorotoluene	12.780	91	30501	4.447	ug/l	99
83) tert-Butylbenzene	12.999	119	31232	4.252	ug/l	96
84) 1,2,4-Trimethylbenzene	13.042	105	34752	4.314	ug/l	100
85) sec-Butylbenzene	13.176	105	49546	4.449	ug/l	99
86) p-Isopropyltoluene	13.292	119	38822	4.240	ug/l	98
87) 1,3-Dichlorobenzene	13.292	146	19350	4.431	ug/l	97
88) 1,4-Dichlorobenzene	13.371	146	19268	4.529	ug/l	91
89) n-Butylbenzene	13.615	91	37774	4.333	ug/l	100
90) Hexachloroethane	13.883	117	8072	4.506	ug/l	98
91) 1,2-Dichlorobenzene	13.664	146	15922	4.244	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.273	75	949	3.754	ug/l	96
93) 1,2,4-Trichlorobenzene	14.919	180	8529	4.159	ug/l	95
94) Hexachlorobutadiene	15.023	225	5627	5.024	ug/l	98
95) Naphthalene	15.145	128	15167	3.859	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	7952	4.563	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
Data File : VY022491.D
Acq On : 02 Jun 2025 11:46
Operator : SY/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025

Quant Time: Jun 03 02:52:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 02:50:07 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
 Data File : VY022491.D
 Acq On : 02 Jun 2025 11:46
 Operator : SY/MD
 Sample : VSTDIC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005

Quant Time: Jun 03 02:52:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 02:50:07 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 06/03/2025
 Supervised By : Mahesh Dadoda 06/03/2025





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.713 min Scan# 9

Delta R.T. 0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion:168 Resp: 201089

Ion Ratio Lower Upper

168 100





99 59.3 44.3 66.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025

Supervised By :Mahesh Dadoda 06/03/2025



#2

Dichlorodifluoromethane

Concen: 4.961 ug/l

RT: 1.867 min Scan# 24

Delta R.T. -0.000 min

Lab File: VY022491.D




Acq: 02 Jun 2025 11:46


Tgt Ion: 85 Resp: 8717

Ion Ratio Lower Upper

85 100


87 32.9 16.7 50.0





#3
Chloromethane
Concen: 5.119 ug/l
RT: 2.068 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 50 Resp: 2653
Ion Ratio Lower Upper
50 100
52 31.1 26.0 39.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025




#4
Vinyl Chloride
Concen: 4.544 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46



Tgt Ion: 62 Resp: 27572
Ion Ratio Lower Upper
62 100
64 32.9 25.3 37.9









#5
Bromomethane
Concen: 4.784 ug/l
RT: 2.598 min Scan# 144
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 94 Resp: 2745
Ion Ratio Lower Upper
94 100
96 90.5 74.0 111.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#6
Chloroethane
Concen: 4.710 ug/l
RT: 2.739 min Scan# 167
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46


Tgt Ion: 64 Resp: 19566
Ion Ratio Lower Upper
64 100
66 32.1 23.8 35.6





#7
 Trichlorofluoromethane
 Concen: 4.428 ug/l
 RT: 3.062 min Scan# 219
 Delta R.T. 0.006 min
 Lab File: VY022491.D
 Acq: 02 Jun 2025 11:46




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005




Tgt Ion:101 Resp: 2282
 Ion Ratio Lower Upper
 101 100
 103 59.0 52.2 78.2

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 06/03/2025
 Supervised By :Mahesh Dadoda 06/03/2025




#8
 Diethyl Ether
 Concen: 4.553 ug/l
 RT: 3.464 min Scan# 286
 Delta R.T. -0.000 min
 Lab File: VY022491.D
 Acq: 02 Jun 2025 11:46



Tgt Ion: 74 Resp: 5184
 Ion Ratio Lower Upper
 74 100
 45 98.9 45.9 137.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 4.744 ug/l
RT: 3.824 min Scan# 344
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion:101 Resp: 10259
Ion Ratio Lower Upper
101 100
85 48.8 39.3 58.9
151 67.9 57.8 86.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#10
Methyl Iodide
Concen: 3.956 ug/l
RT: 4.007 min Scan# 375
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion:142 Resp: 9174
Ion Ratio Lower Upper
142 100
127 47.3 37.1 55.7
141 7.8 11.4 17.2#





#11

Tert butyl alcohol

Concen: 21.587 ug/l

RT: 4.891 min Scan# 519

Delta R.T. 0.006 min

Lab File: VY022491.D


Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005



Tgt Ion: 59 Resp: 3374

Ion Ratio Lower Upper

59 100




57 5.6 8.0 12.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025

Supervised By :Mahesh Dadoda 06/03/2025



#12

1,1-Dichloroethene

Concen: 4.573 ug/l

RT: 3.799 min Scan# 341

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46




Tgt Ion: 96 Resp: 9440


Ion Ratio Lower Upper

96 100

61 164.6 134.6 201.8

98 66.2 50.2 75.4









#13
Acrolein
Concen: 29.646 ug/l
RT: 3.671 min Scan# 318
Delta R.T. 0.012 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 56 Resp: 5650
Ion Ratio Lower Upper
56 100
55 75.3 57.0 85.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#14
Allyl chloride
Concen: 4.693 ug/l
RT: 4.391 min Scan# 438
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion: 41 Resp: 15096
Ion Ratio Lower Upper
41 100
39 84.0 69.1 103.7
76 40.8 33.5 50.3









#15
Acrylonitrile
Concen: 22.026 ug/l
RT: 5.080 min Scan# 51
Delta R.T. 0.012 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 53 Resp: 1066
Ion Ratio Lower Upper
53 100
52 83.8 65.7 98.5
51 41.4 30.1 45.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#16
Acetone
Concen: 30.108 ug/l
RT: 3.885 min Scan# 355
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion: 43 Resp: 13023
Ion Ratio Lower Upper
43 100
58 37.3 24.0 36.0#





#17

Carbon Disulfide

Concen: 4.586 ug/l

RT: 4.110 min Scan# 392

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 3022

Ion Ratio Lower Upper

76 100





78 9.5 7.4 11.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025

Supervised By :Mahesh Dadoda 06/03/2025



#18

Methyl Acetate

Concen: 4.327 ug/l

RT: 4.397 min Scan# 439

Delta R.T. -0.000 min

Lab File: VY022491.D




Acq: 02 Jun 2025 11:46


Tgt Ion: 43 Resp: 5630

Ion Ratio Lower Upper

43 100

74 22.8 19.8 29.6





#19

Methyl tert-butyl Ether

Concen: 4.562 ug/l

RT: 5.122 min Scan# 51

Delta R.T. -0.006 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC005

Tgt Ion: 73 Resp: 26270

Ion Ratio Lower Upper

73 100





57 20.2 17.4 26.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025

Supervised By :Mahesh Dadoda 06/03/2025



#20

Methylene Chloride

Concen: 7.168 ug/l

RT: 4.622 min Scan# 476

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Tgt Ion: 84 Resp: 17312




Ion Ratio Lower Upper


84 100

49 112.6 101.9 152.9

51 33.2 29.8 44.8


86 55.4 51.3 76.9





#21
trans-1,2-Dichloroethene
Concen: 4.549 ug/l
RT: 5.116 min Scan# 511
Delta R.T. -0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 96 Resp: 10489
Ion Ratio Lower Upper
96 100
61 119.8 112.6 169.0
98 57.7 50.6 75.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025




#22
Diisopropyl ether
Concen: 4.544 ug/l
RT: 6.037 min Scan# 708
Delta R.T. 0.012 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Tgt Ion: 45 Resp: 32497
Ion Ratio Lower Upper
45 100
43 55.1 48.3 72.5
87 29.3 23.0 34.6
59 9.6 9.4 14.0





#23
 Vinyl Acetate
 Concen: 21.559 ug/l
 RT: 5.970 min Scan# 697
 Delta R.T. -0.000 min
 Lab File: VY022491.D
 Acq: 02 Jun 2025 11:46




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005




Tgt Ion: 43 Resp: 91478
 Ion Ratio Lower Upper
 43 100
 86 10.1 8.6 12.8

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 06/03/2025
 Supervised By :Mahesh Dadoda 06/03/2025



#24
 1,1-Dichloroethane
 Concen: 4.603 ug/l
 RT: 5.921 min Scan# 689
 Delta R.T. -0.000 min
 Lab File: VY022491.D
 Acq: 02 Jun 2025 11:46



Tgt Ion: 63 Resp: 19460
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.4 10.1
 100 5.0 2.1 6.3









#25
2-Butanone
Concen: 22.605 ug/l
RT: 6.909 min Scan# 851
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 1442
Ion Ratio Lower Upper
43 100
72 26.2 22.9 34.3




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#26
2,2-Dichloropropane
Concen: 4.471 ug/l
RT: 6.896 min Scan# 849
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion: 77 Resp: 16805
Ion Ratio Lower Upper
77 100
97 23.2 11.4 34.2





#27

cis-1,2-Dichloroethene

Concen: 4.467 ug/l

RT: 6.896 min Scan# 849

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion: 96	Resp: 11900
Ion Ratio	Lower Upper
96 100	
61 146.0	0.0 290.0
98 67.8	0.0 128.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#28

Bromochloromethane

Concen: 5.175 ug/l




RT: 7.250 min Scan# 907


Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Tgt Ion: 49	Resp: 9453
Ion Ratio	Lower Upper
49 100	
129 1.8	0.0 3.6
130 69.5	55.7 83.5





#29

Tetrahydrofuran

Concen: 22.781 ug/l

RT: 7.274 min Scan# 910

Delta R.T. 0.006 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC005

Tgt Ion:	42	Resp:	937
Ion	Ratio	Lower	Upper
42	100		
72	41.4	37.4	56.2
71	41.5	34.2	51.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#30

Chloroform

Concen: 4.620 ug/l




RT: 7.427 min Scan# 936


Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46


Tgt Ion:	83	Resp:	19302
Ion	Ratio	Lower	Upper
83	100		
85	57.4	54.2	81.4





#31
Cyclohexane
Concen: 5.519 ug/l
RT: 7.707 min Scan# 982
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion: 56 Resp: 2348
Ion Ratio Lower Upper
56 100
69 51.3 27.0 40.6
84 88.1 72.6 109.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#32
1,1,1-Trichloroethane
Concen: 4.413 ug/l
RT: 7.628 min Scan# 969
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion: 97 Resp: 16476
Ion Ratio Lower Upper
97 100
99 65.1 50.9 76.3
61 48.9 38.5 57.7





#33

1,2-Dichloroethane-d4

Concen: 4.802 ug/l

RT: 8.067 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 65 Resp: 10508

Ion Ratio Lower Upper

65 100





67 56.3 0.0 103.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025

Supervised By :Mahesh Dadoda 06/03/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.622 min Scan# 1132

Delta R.T. 0.006 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46




Tgt Ion:114 Resp: 353755


Ion Ratio Lower Upper

114 100

63 20.6 0.0 40.8

88 14.3 0.0 27.8









#35
Dibromofluoromethane
Concen: 4.477 ug/l
RT: 7.646 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 113 Resp: 932
Ion Ratio Lower Upper
113 100
111 101.4 81.1 121.7
192 17.8 14.2 21.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#36
1,1-Dichloropropene
Concen: 4.367 ug/l
RT: 7.841 min Scan# 1004
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46


Tgt Ion: 75 Resp: 14700
Ion Ratio Lower Upper
75 100
110 33.1 16.9 50.7
77 32.0 24.8 37.2





#37
Ethyl Acetate
Concen: 4.326 ug/l m
RT: 6.994 min Scan# 865
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005



Tgt Ion: 43 Resp: 6409
Ion Ratio Lower Upper
43 100
61 17.3 11.5 17.3
70 12.8 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#38
Carbon Tetrachloride
Concen: 4.233 ug/l
RT: 7.823 min Scan# 1001
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion:117 Resp: 14691
Ion Ratio Lower Upper
117 100
119 87.1 75.0 112.6
121 31.7 25.0 37.4





#39

Methylcyclohexane

Concen: 4.415 ug/l

RT: 9.109 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY022491.D


Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005






Tgt Ion:	83	Resp:	20070
Ion	Ratio	Lower	Upper
83	100		
55	76.4	58.1	87.1
98	44.4	39.0	58.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#40

Benzene

Concen: 4.326 ug/l




RT: 8.085 min Scan# 1044


Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46


Tgt Ion:	78	Resp:	42926
Ion	Ratio	Lower	Upper
78	100		
77	26.2	19.3	28.9





#41
Methacrylonitrile
Concen: 4.394 ug/l
RT: 7.232 min Scan# 904
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion: 41 Resp: 4192
Ion Ratio Lower Upper
41 100
39 40.1 46.6 69.8
67 74.0 63.3 94.9
52 33.6 27.5 41.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025




#42
1,2-Dichloroethane
Concen: 4.198 ug/l
RT: 8.164 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46



Tgt Ion: 62 Resp: 11332
Ion Ratio Lower Upper
62 100
98 9.2 0.0 19.2









#43
Isopropyl Acetate
Concen: 4.117 ug/l
RT: 8.207 min Scan# 1064
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 43 Resp: 13330
Ion Ratio Lower Upper
43 100
61 16.2 15.6 23.4
87 13.6 10.2 15.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#44
Trichloroethene
Concen: 4.292 ug/l
RT: 8.872 min Scan# 1173
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion:130 Resp: 10404
Ion Ratio Lower Upper
130 100
95 122.3 0.0 203.0





#45

1,2-Dichloropropane

Concen: 4.091 ug/l

RT: 9.146 min Scan# 1218

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion: 63 Resp: 9678

Ion Ratio Lower Upper

63 100





65 29.7 24.0 36.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025

Supervised By :Mahesh Dadoda 06/03/2025



#46

Dibromomethane

Concen: 4.223 ug/l

RT: 9.244 min Scan# 1234

Delta R.T. 0.012 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46




Tgt Ion: 93 Resp: 5596


Ion Ratio Lower Upper

93 100

95 79.9 65.6 98.4

174 81.3 70.4 105.6









#47
Bromodichloromethane
Concen: 3.999 ug/l
RT: 9.433 min Scan# 1265
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 1344
Ion Ratio Lower Upper
83 100
85 64.6 49.1 73.7
127 7.0 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#48
Methyl methacrylate
Concen: 4.374 ug/l
RT: 9.225 min Scan# 1231
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion: 41 Resp: 6483
Ion Ratio Lower Upper
41 100
69 80.0 67.8 101.8
39 57.3 55.0 82.4





#49

1,4-Dioxane

Concen: 85.767 ug/l m

RT: 9.231 min Scan# 1

Delta R.T. -0.012 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	88	Resp:	1350
Ion Ratio	Lower	Upper	
88	100		
43	53.1	25.4	38.2
58	65.5	54.9	82.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#50

Toluene-d8

Concen: 4.482 ug/l




RT: 10.109 min Scan# 1376


Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46


Tgt Ion:	98	Resp:	37729
Ion Ratio	Lower	Upper	
98	100		
100	63.0	51.4	77.0





#51
4-Methyl-2-Pentanone
Concen: 19.273 ug/l
RT: 10.000 min Scan# 1358
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion: 43 Resp: 30969
Ion Ratio Lower Upper
43 100
58 36.7 30.4 45.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#52
Toluene
Concen: 4.150 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion: 92 Resp: 25893
Ion Ratio Lower Upper
92 100
91 171.8 139.0 208.4









#53
t-1,3-Dichloropropene
Concen: 3.982 ug/l
RT: 10.396 min Scan# 1423
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument : MSVOA_Y
ClientSampleId : VSTDIC005

Tgt Ion: 75 Resp: 1238
Ion Ratio Lower Upper
75 100
77 38.1 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#54
cis-1,3-Dichloropropene
Concen: 4.103 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46


Tgt Ion: 75 Resp: 14966
Ion Ratio Lower Upper
75 100
77 31.5 25.3 37.9
39 53.0 42.7 64.1





#55
1,1,2-Trichloroethane
Concen: 4.385 ug/l
RT: 10.579 min Scan# 1452
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion	Ratio	Lower	Upper
97	100		
83	85.3	68.1	102.1
85	60.1	48.2	72.2
99	57.6	49.5	74.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#56
Ethyl methacrylate
Concen: 3.751 ug/l
RT: 10.445 min Scan# 1431
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion	Ratio	Lower	Upper
69	100		
41	73.4	52.2	78.2
39	46.3	36.7	55.1









#57
1,3-Dichloropropane
Concen: 4.255 ug/l
RT: 10.719 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument : MSVOA_Y
ClientSampleId : VSTDICC005

Tgt Ion: 76 Resp: 12648
Ion Ratio Lower Upper
76 100
78 33.6 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#58
2-Chloroethyl Vinyl ether
Concen: 22.409 ug/l
RT: 9.719 min Scan# 1312
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46


Tgt Ion: 63 Resp: 24214
Ion Ratio Lower Upper
63 100
106 28.3 21.6 32.4





#59
2-Hexanone
Concen: 19.832 ug/l
RT: 10.768 min Scan# 1484
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion: 43 Resp: 2124
Ion Ratio Lower Upper
43 100
58 50.4 27.3 81.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#60
Dibromochloromethane
Concen: 4.152 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion:129 Resp: 8781
Ion Ratio Lower Upper
129 100
127 73.3 39.1 117.2









#61
1,2-Dibromoethane
Concen: 3.991 ug/l
RT: 11.018 min Scan# 1525
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:107 Resp: 6140
Ion Ratio Lower Upper
107 100
109 96.8 77.7 116.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#62
4-Bromofluorobenzene
Concen: 4.697 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46


Tgt Ion: 95 Resp: 11988
Ion Ratio Lower Upper
95 100
174 81.7 0.0 170.0
176 80.2 0.0 166.2





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 1591
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005




Tgt Ion:117 Resp: 289394
Ion Ratio Lower Upper
117 100
82 55.7 44.6 66.8
119 31.2 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#64
Tetrachloroethene
Concen: 4.417 ug/l
RT: 10.652 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46



Tgt Ion:164 Resp: 10863
Ion Ratio Lower Upper
164 100
166 131.2 104.0 156.0
129 96.1 74.5 111.7
131 102.0 72.3 108.5









#65
Chlorobenzene
Concen: 4.377 ug/l
RT: 11.444 min Scan# 1146
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion:112 Resp: 2746
Ion Ratio Lower Upper
112 100
114 30.9 25.0 37.6




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 3.959 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion:131 Resp: 8422
Ion Ratio Lower Upper
131 100
133 96.8 47.3 141.8
119 74.3 33.1 99.3









#67
Ethyl Benzene
Concen: 4.098 ug/l
RT: 11.524 min Scan# 1608
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 4788
Ion Ratio Lower Upper
91 100
106 32.7 24.2 36.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#68
m/p-Xylenes
Concen: 8.060 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion:106 Resp: 35645
Ion Ratio Lower Upper
106 100
91 200.7 159.4 239.0









#69
o-Xylene
Concen: 4.075 ug/l
RT: 11.956 min Scan# 1680
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion:106 Resp: 1686
Ion Ratio Lower Upper
106 100
91 219.4 105.8 317.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#70
Styrene
Concen: 3.840 ug/l
RT: 11.975 min Scan# 1682
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46


Tgt Ion:104 Resp: 26313
Ion Ratio Lower Upper
104 100
78 48.8 39.5 59.3
103 55.8 44.1 66.1





#71
Bromoform
Concen: 4.159 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:173 Resp: 4649
Ion Ratio Lower Upper
173 100
175 49.9 23.8 71.5
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.346 min Scan# 1907
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46



Tgt Ion:152 Resp: 127807
Ion Ratio Lower Upper
152 100
115 59.2 28.9 86.7
150 160.5 0.0 349.6





#73

Isopropylbenzene

Concen: 4.307 ug/l

RT: 12.255 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC005

Tgt Ion:105 Resp: 44350

Ion Ratio Lower Upper

105 100





120 25.8 13.0 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025

Supervised By :Mahesh Dadoda 06/03/2025



#74

N-ethyl acetate

Concen: 3.699 ug/l

RT: 12.072 min Scan# 1698

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Tgt Ion: 43 Resp: 9632




Ion Ratio Lower Upper


43 100

70 45.9 34.2 51.2

55 25.1 18.2 27.4

61 23.9 19.0 28.6









#75
1,1,2,2-Tetrachloroethane
Concen: 4.359 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 7289
Ion Ratio Lower Upper
83 100
131 8.1 4.9 14.7
85 62.4 32.6 97.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#76
1,2,3-Trichloropropane
Concen: 4.432 ug/l m
RT: 12.560 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion: 75 Resp: 7478
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 4.234 ug/l

RT: 12.536 min Scan# 1774

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	156	Resp:	924
Ion Ratio	Lower	Upper	
156	100		
77	196.5	94.0	282.1
158	98.9	48.2	144.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#78

n-propylbenzene

Concen: 4.511 ug/l




RT: 12.597 min Scan# 1784


Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Tgt Ion:	91	Resp:	56543
Ion Ratio	Lower	Upper	
91	100		
120	22.5	10.8	32.4









#79
2-Chlorotoluene
Concen: 4.591 ug/l
RT: 12.682 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005

Tgt Ion: 91 Resp: 3038
Ion Ratio Lower Upper
91 100
126 34.6 17.0 51.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#80
1,3,5-Trimethylbenzene
Concen: 4.362 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46


Tgt Ion:105 Resp: 35122
Ion Ratio Lower Upper
105 100
120 48.3 24.1 72.4





#81
trans-1,4-Dichloro-2-butene
Concen: 4.411 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion: 75 Resp: 2590
Ion Ratio Lower Upper
75 100
53 99.3 78.5 117.7
89 51.7 35.7 53.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#82
4-Chlorotoluene
Concen: 4.447 ug/l
RT: 12.780 min Scan# 1814
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion: 91 Resp: 30501
Ion Ratio Lower Upper
91 100
126 34.6 17.0 50.9





#83

tert-Butylbenzene

Concen: 4.252 ug/l

RT: 12.999 min Scan# 1850

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	119	Resp:	3123
Ion Ratio	Lower	Upper	
119	100		
91	67.0	32.1	96.3
134	23.9	13.0	38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#84

1,2,4-Trimethylbenzene

Concen: 4.314 ug/l




RT: 13.042 min Scan# 1857


Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Tgt Ion:	105	Resp:	34752
Ion Ratio	Lower	Upper	
105	100		
120	45.2	22.7	68.1









#85
 sec-Butylbenzene
 Concen: 4.449 ug/l
 RT: 13.176 min Scan# 1879
 Delta R.T. -0.000 min
 Lab File: VY022491.D
 Acq: 02 Jun 2025 11:46

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC005

Tgt Ion:105 Resp: 49540
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.7 28.9




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
 Supervised By :Mahesh Dadoda 06/03/2025



#86
 p-Isopropyltoluene
 Concen: 4.240 ug/l
 RT: 13.292 min Scan# 1898
 Delta R.T. -0.000 min
 Lab File: VY022491.D
 Acq: 02 Jun 2025 11:46


Tgt Ion:119 Resp: 38822
 Ion Ratio Lower Upper
 119 100
 134 25.0 13.1 39.1
 91 24.2 11.6 34.8





#87
1,3-Dichlorobenzene
Concen: 4.431 ug/l
RT: 13.292 min Scan# 1898
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion:146 Resp: 19350
Ion Ratio Lower Upper
146 100
111 38.7 20.1 60.2
148 61.9 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#88
1,4-Dichlorobenzene
Concen: 4.529 ug/l
RT: 13.371 min Scan# 1911
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion:146 Resp: 19268
Ion Ratio Lower Upper
146 100
111 44.9 19.4 58.2
148 69.5 31.6 94.8









#89
n-Butylbenzene
Concen: 4.333 ug/l
RT: 13.615 min Scan# 1951
Delta R.T. -0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Tgt Ion: 91 Resp: 37774
Ion Ratio Lower Upper
91 100
92 52.0 25.9 77.8
134 23.8 12.0 36.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#90
Hexachloroethane
Concen: 4.506 ug/l
RT: 13.883 min Scan# 1995
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46


Tgt Ion:117 Resp: 8072
Ion Ratio Lower Upper
117 100
201 69.9 34.0 102.0





#91
1,2-Dichlorobenzene
Concen: 4.244 ug/l
RT: 13.664 min Scan# 1958
Delta R.T. 0.006 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005




Tgt Ion:146 Resp: 1592
Ion Ratio Lower Upper
146 100
111 42.6 20.2 60.6
148 64.0 31.4 94.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 3.754 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Tgt Ion: 75 Resp: 949
Ion Ratio Lower Upper
75 100
155 87.0 39.9 119.6
157 109.5 54.0 162.0





#93
1,2,4-Trichlorobenzene
Concen: 4.159 ug/l
RT: 14.919 min Scan# 2165
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC005



Tgt Ion:180 Resp: 8529
Ion Ratio Lower Upper
180 100
182 101.9 48.4 145.3
145 32.6 15.5 46.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#94
Hexachlorobutadiene
Concen: 5.024 ug/l
RT: 15.023 min Scan# 2182
Delta R.T. -0.000 min
Lab File: VY022491.D
Acq: 02 Jun 2025 11:46

Tgt Ion:225 Resp: 5627
Ion Ratio Lower Upper
225 100
223 62.4 32.5 97.4
227 65.6 32.3 96.9





#95

Naphthalene

Concen: 3.859 ug/l

RT: 15.145 min Scan# 21

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC005

Tgt Ion:	128	Resp:	1516
Ion Ratio	Lower	Upper	
128	100		
127	13.7	10.4	15.6
129	11.4	8.9	13.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 06/03/2025
Supervised By :Mahesh Dadoda 06/03/2025



#96

1,2,3-Trichlorobenzene

Concen: 4.563 ug/l

RT: 15.328 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY022491.D

Acq: 02 Jun 2025 11:46

Tgt Ion:	180	Resp:	7952
Ion Ratio	Lower	Upper	
180	100		
182	94.7	47.1	141.3
145	32.5	15.2	45.6

