

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052925\
 Data File : VY022456.D
 Acq On : 29 May 2025 09:50
 Operator : SY/MD
 Sample : VY0529SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0529SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh

Dadoda

05/30/2025

Supervised By :Semsettin

Yesilyurt

05/30/2025

Quant Time: May 30 05:30:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 28 10:34:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	205776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	359832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	305847	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	140548	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.067	65	119518	55.142	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.280%	
35) Dibromofluoromethane	7.640	113	110555	52.463	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.920%	
50) Toluene-d8	10.109	98	443079	51.578	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 103.160%	
62) 4-Bromofluorobenzene	12.408	95	129686	49.908	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.820%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.867	85	45950	23.405	ug/l	96
3) Chloromethane	2.068	50	127635	26.430	ug/l	99
4) Vinyl Chloride	2.208	62	143639	24.484	ug/l	98
5) Bromomethane	2.598	94	153498	28.792	ug/l	96
6) Chloroethane	2.739	64	103463	26.988	ug/l	99
7) Trichlorofluoromethane	3.050	101	125182	25.290	ug/l	97
8) Diethyl Ether	3.458	74	27089	23.735	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.818	101	46248	21.297	ug/l	100
10) Methyl Iodide	4.007	142	52695	21.509	ug/l	98
11) Tert butyl alcohol	4.897	59	18285	119.684	ug/l	99
12) 1,1-Dichloroethene	3.793	96	47557	22.059	ug/l	96
13) Acrolein	3.659	56	23667	103.387	ug/l	99
14) Allyl chloride	4.391	41	72904	22.480	ug/l	98
15) Acrylonitrile	5.068	53	53746	116.830	ug/l	100
16) Acetone	3.885	43	41379	116.179	ug/l	96
17) Carbon Disulfide	4.110	76	148882	21.639	ug/l	99
18) Methyl Acetate	4.391	43	33619	25.198	ug/l	99
19) Methyl tert-butyl Ether	5.116	73	131747	23.017	ug/l	99
20) Methylene Chloride	4.629	84	53789	21.387	ug/l	97
21) trans-1,2-Dichloroethene	5.122	96	51482	21.873	ug/l	97
22) Diisopropyl ether	6.025	45	161945	22.626	ug/l	98
23) Vinyl Acetate	5.970	43	456630	111.108	ug/l	98
24) 1,1-Dichloroethane	5.921	63	96288	22.766	ug/l	99
25) 2-Butanone	6.903	43	68030	114.690	ug/l	99
26) 2,2-Dichloropropane	6.890	77	81729	21.341	ug/l	100
27) cis-1,2-Dichloroethene	6.896	96	59568	21.861	ug/l	99
28) Bromochloromethane	7.256	49	38824	21.983	ug/l	100
29) Tetrahydrofuran	7.274	42	46439	117.391	ug/l	98
30) Chloroform	7.427	83	94926	22.777	ug/l	98
31) Cyclohexane	7.707	56	86206	20.677	ug/l #	93
32) 1,1,1-Trichloroethane	7.622	97	81565	21.565	ug/l	98
36) 1,1-Dichloropropene	7.841	75	69669	20.835	ug/l	100
37) Ethyl Acetate	6.994	43	31627	22.255	ug/l	96
38) Carbon Tetrachloride	7.823	117	71355	20.484	ug/l	94
39) Methylcyclohexane	9.116	83	88788	19.975	ug/l	97
40) Benzene	8.085	78	212449	21.352	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052925\
 Data File : VY022456.D
 Acq On : 29 May 2025 09:50
 Operator : SY/MD
 Sample : VY0529SBS01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0529SBS01

Manual Integrations
 APPROVED

Reviewed By :Mahesh

Dadoda

05/30/2025

Supervised By :Semsettin

Yesilyurt

05/30/2025

Quant Time: May 30 05:30:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 28 10:34:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.232	41	17862	20.709	ug/l	89
42) 1,2-Dichloroethane	8.165	62	57566	21.906	ug/l	100
43) Isopropyl Acetate	8.201	43	68235	21.973	ug/l #	87
44) Trichloroethene	8.872	130	51499	20.937	ug/l	98
45) 1,2-Dichloropropane	9.146	63	51267	21.930	ug/l	96
46) Dibromomethane	9.238	93	28152	21.556	ug/l	98
47) Bromodichloromethane	9.427	83	72157	21.486	ug/l	99
48) Methyl methacrylate	9.225	41	31618	21.934	ug/l	100
49) 1,4-Dioxane	9.238	88	6628	459.490	ug/l	99
51) 4-Methyl-2-Pentanone	10.000	43	169643	110.718	ug/l	100
52) Toluene	10.176	92	127807	20.565	ug/l	99
53) t-1,3-Dichloropropene	10.402	75	66189	21.159	ug/l	96
54) cis-1,3-Dichloropropene	9.859	75	77958	21.209	ug/l	97
55) 1,1,2-Trichloroethane	10.573	97	35077	21.244	ug/l	95
56) Ethyl methacrylate	10.445	69	50615	21.361	ug/l	97
57) 1,3-Dichloropropane	10.719	76	62576	21.506	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	124329	105.282	ug/l	98
59) 2-Hexanone	10.768	43	112377	111.915	ug/l	100
60) Dibromochloromethane	10.914	129	44484	21.090	ug/l	98
61) 1,2-Dibromoethane	11.018	107	32567	21.565	ug/l	100
64) Tetrachloroethene	10.652	164	54451	20.904	ug/l	97
65) Chlorobenzene	11.444	112	136170	20.880	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.518	131	45682	20.674	ug/l	98
67) Ethyl Benzene	11.524	91	245127	20.243	ug/l	99
68) m/p-Xylenes	11.633	106	183812	40.037	ug/l	100
69) o-Xylene	11.957	106	87779	20.343	ug/l	100
70) Styrene	11.975	104	146742	20.424	ug/l	100
71) Bromoform	12.133	173	24149	20.865	ug/l #	97
73) Isopropylbenzene	12.255	105	226043	20.435	ug/l	99
74) N-amyl acetate	12.072	43	58739	21.701	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.511	83	37969	21.786	ug/l	99
76) 1,2,3-Trichloropropane	12.560	75	26369m	18.730	ug/l	
77) Bromobenzene	12.536	156	48852	20.522	ug/l	96
78) n-propylbenzene	12.597	91	274435	20.673	ug/l	98
79) 2-Chlorotoluene	12.682	91	148607	20.680	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	179270	20.742	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	13230	21.252	ug/l	99
82) 4-Chlorotoluene	12.780	91	152410	20.594	ug/l	99
83) tert-Butylbenzene	12.999	119	157690	20.194	ug/l	98
84) 1,2,4-Trimethylbenzene	13.042	105	178877	20.566	ug/l	100
85) sec-Butylbenzene	13.176	105	238715	20.613	ug/l	100
86) p-Isopropyltoluene	13.292	119	196000	20.268	ug/l	100
87) 1,3-Dichlorobenzene	13.292	146	97966	20.454	ug/l	99
88) 1,4-Dichlorobenzene	13.371	146	97065	21.194	ug/l	99
89) n-Butylbenzene	13.621	91	188466	20.785	ug/l	100
90) Hexachloroethane	13.883	117	40139	20.858	ug/l	96
91) 1,2-Dichlorobenzene	13.657	146	84578	21.142	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.279	75	6003	22.291	ug/l	94
93) 1,2,4-Trichlorobenzene	14.919	180	48125	21.699	ug/l	97
94) Hexachlorobutadiene	15.023	225	24884	21.380	ug/l	97
95) Naphthalene	15.145	128	91603	22.342	ug/l	99
96) 1,2,3-Trichlorobenzene	15.334	180	39958	21.968	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052925\
Data File : VY022456.D
Acq On : 29 May 2025 09:50
Operator : SY/MD
Sample : VY0529SBSD01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Quant Time: May 30 05:30:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
Quant Title : SW846 8260
QLast Update : Wed May 28 10:34:06 2025
Response via : Initial Calibration

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
05/30/2025

Supervised By :Semsettin
Yesilyurt
05/30/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY052925\
 Data File : VY022456.D
 Acq On : 29 May 2025 09:50
 Operator : SY/MD
 Sample : VY0529SBSD01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0529SBSD01

Quant Time: May 30 05:30:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y052725S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 28 10:34:06 2025
 Response via : Initial Calibration


Manual Integrations APPROVED


Reviewed By :Mahesh
 Dadoda

05/30/2025

Supervised By :Semsetin
 Yesilyurt


05/30/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 983
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01






Tgt Ion:168 Resp: 205770
Ion Ratio Lower Upper
168 100
99 57.8 44.7 67.1

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
05/30/2025


Supervised By :Semsettin
Yesilyurt
05/30/2025



#2
Dichlorodifluoromethane
Concen: 23.405 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 85 Resp: 45950
Ion Ratio Lower Upper
85 100
87 30.5 16.4 49.2





#3
Chloromethane
Concen: 26.430 ug/l
RT: 2.068 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion: 50 Resp: 127639
Ion Ratio Lower Upper
50 100
52 32.5 25.5 38.3


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#4
Vinyl Chloride
Concen: 24.484 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 62 Resp: 143639
Ion Ratio Lower Upper
62 100
64 31.4 26.0 39.0





#5
Bromomethane
Concen: 28.792 ug/l
RT: 2.598 min Scan# 144
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion: 94 Resp: 153498
Ion Ratio Lower Upper
94 100
96 91.5 76.0 114.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#6
Chloroethane
Concen: 26.988 ug/l
RT: 2.739 min Scan# 167
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Tgt Ion: 64 Resp: 103463
Ion Ratio Lower Upper
64 100
66 32.8 25.9 38.9







#7
Trichlorofluoromethane
Concen: 25.290 ug/l
RT: 3.050 min Scan# 219
Delta R.T. -0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01

Tgt Ion:101 Resp: 12518
Ion Ratio Lower Upper
101 100
103 66.0 50.9 76.3



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#8
Diethyl Ether
Concen: 23.735 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Tgt Ion: 74 Resp: 27089
Ion Ratio Lower Upper
74 100
45 89.9 44.9 134.5







#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.297 ug/l
RT: 3.818 min Scan# 344
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01

Tgt Ion:101 Resp: 46248
Ion Ratio Lower Upper
101 100
85 49.0 39.2 58.8
151 73.2 58.8 88.2



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#10
Methyl Iodide
Concen: 21.509 ug/l
RT: 4.007 min Scan# 375
Delta R.T. -0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Tgt Ion:142 Resp: 52695
Ion Ratio Lower Upper
142 100
127 46.4 36.2 54.4
141 15.5 11.6 17.4







#11
Tert butyl alcohol
Concen: 119.684 ug/l
RT: 4.897 min Scan# 519
Delta R.T. 0.012 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01

Tgt Ion: 59 Resp: 1828
Ion Ratio Lower Upper
59 100
57 10.9 9.1 13.7



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#12
1,1-Dichloroethene
Concen: 22.059 ug/l
RT: 3.793 min Scan# 340
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 96 Resp: 47557
Ion Ratio Lower Upper
96 100
61 165.0 134.1 201.1
98 55.9 50.2 75.2





#13
Acrolein
Concen: 103.387 ug/l
RT: 3.659 min Scan# 318
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
Client Sample Id :
VY0529SBSD01



Tgt Ion: 56 Resp: 2366
Ion Ratio Lower Upper
56 100
55 69.6 56.5 84.7


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#14
Allyl chloride
Concen: 22.480 ug/l
RT: 4.391 min Scan# 438
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion: 41 Resp: 72904
Ion Ratio Lower Upper
41 100
39 86.0 66.6 100.0
76 41.2 33.7 50.5





#15
Acrylonitrile
Concen: 116.830 ug/l
RT: 5.068 min Scan# 549
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion: 53 Resp: 53740
Ion Ratio Lower Upper
53 100
52 81.2 65.1 97.7
51 37.6 30.0 45.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#16
Acetone
Concen: 116.179 ug/l
RT: 3.885 min Scan# 355
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 43 Resp: 41379
Ion Ratio Lower Upper
43 100
58 27.6 20.6 30.8





#17
Carbon Disulfide
Concen: 21.639 ug/l
RT: 4.110 min Scan# 392
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion: 76 Resp: 14888
Ion Ratio Lower Upper
76 100
78 8.6 7.2 10.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#18
Methyl Acetate
Concen: 25.198 ug/l
RT: 4.391 min Scan# 438
Delta R.T. -0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion: 43 Resp: 33619
Ion Ratio Lower Upper
43 100
74 25.9 21.2 31.8





#19
Methyl tert-butyl Ether
Concen: 23.017 ug/l
RT: 5.116 min Scan# 517
Delta R.T. -0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion: 73 Resp: 13174
Ion Ratio Lower Upper
73 100
57 21.5 17.7 26.5


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#20
Methylene Chloride
Concen: 21.387 ug/l
RT: 4.629 min Scan# 477
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 84 Resp: 53789
Ion Ratio Lower Upper
84 100
49 123.2 97.8 146.8
51 36.7 29.5 44.3
86 58.7 51.8 77.6





#21
trans-1,2-Dichloroethene
Concen: 21.873 ug/l
RT: 5.122 min Scan# 5148
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion: 96 Resp: 5148
Ion Ratio Lower Upper
96 100
61 137.4 110.4 165.6
98 66.8 48.6 73.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#22
Diisopropyl ether
Concen: 22.626 ug/l
RT: 6.025 min Scan# 706
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 45 Resp: 161945
Ion Ratio Lower Upper
45 100
43 62.2 49.8 74.8
87 26.7 23.0 34.6
59 11.0 9.7 14.5





#23
 Vinyl Acetate
 Concen: 111.108 ug/l
 RT: 5.970 min Scan# 697
 Delta R.T. 0.000 min
 Lab File: VY022456.D
 Acq: 29 May 2025 09:50



Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0529SBS01



Tgt Ion: 43 Resp: 456630
 Ion Ratio Lower Upper
 43 100
 86 10.4 8.8 13.2


Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda




05/30/2025
 Supervised By :Semsettin
 Yesilyurt



05/30/2025




#24
 1,1-Dichloroethane
 Concen: 22.766 ug/l
 RT: 5.921 min Scan# 689
 Delta R.T. 0.000 min
 Lab File: VY022456.D
 Acq: 29 May 2025 09:50



Tgt Ion: 63 Resp: 96288
 Ion Ratio Lower Upper
 63 100
 98 6.4 3.5 10.5
 100 4.7 2.1 6.5







#25
2-Butanone
Concen: 114.690 ug/l
RT: 6.903 min Scan# 851
Delta R.T. -0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01

Tgt Ion: 43 Resp: 68030
Ion Ratio Lower Upper
43 100
72 28.0 22.8 34.2



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#26
2,2-Dichloropropane
Concen: 21.341 ug/l
RT: 6.890 min Scan# 848
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 77 Resp: 81729
Ion Ratio Lower Upper
77 100
97 23.0 11.6 34.7





#27
 cis-1,2-Dichloroethene
 Concen: 21.861 ug/l
 RT: 6.896 min Scan# 849
 Delta R.T. 0.000 min
 Lab File: VY022456.D
 Acq: 29 May 2025 09:50



Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0529SBS01



Tgt Ion: 96 Resp: 5956
 Ion Ratio Lower Upper
 96 100
 61 148.0 0.0 292.0
 98 64.8 0.0 131.0


Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda




05/30/2025
 Supervised By :Semsettin
 Yesilyurt



05/30/2025




#28
 Bromochloromethane
 Concen: 21.983 ug/l
 RT: 7.256 min Scan# 908
 Delta R.T. 0.006 min
 Lab File: VY022456.D
 Acq: 29 May 2025 09:50



Tgt Ion: 49 Resp: 38824
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 4.0
 130 70.9 56.9 85.3





#29

Tetrahydrofuran

Concen: 117.391 ug/l

RT: 7.274 min Scan# 911

Delta R.T. 0.006 min

Lab File: VY022456.D


Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y

ClientSampleId :


VY0529SBSD01



Tgt Ion:	42	Resp:	46439
Ion	Ratio	Lower	Upper
42	100		
72	48.1	37.7	56.5
71	42.3	34.7	52.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda



05/30/2025
Supervised By :Semsettin
Yesilyurt

05/30/2025



#30

Chloroform

Concen: 22.777 ug/l




RT: 7.427 min Scan# 936


Delta R.T. 0.000 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50


Tgt Ion:	83	Resp:	94926
Ion	Ratio	Lower	Upper
83	100		
85	63.8	50.1	75.1





#31
Cyclohexane
Concen: 20.677 ug/l
RT: 7.707 min Scan# 981
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion: 56 Resp: 86200
Ion Ratio Lower Upper
56 100
69 36.9 24.6 36.8
84 92.7 70.4 105.6


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#32
1,1,1-Trichloroethane
Concen: 21.565 ug/l
RT: 7.622 min Scan# 968
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Tgt Ion: 97 Resp: 81565
Ion Ratio Lower Upper
97 100
99 65.8 51.3 76.9
61 48.6 37.3 55.9





#33

1,2-Dichloroethane-d4

Concen: 55.142 ug/l

RT: 8.067 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VY022456.D


Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y

ClientSampleId :

VY0529SBSD01



Tgt Ion: 65 Resp: 119518

Ion Ratio Lower Upper



65 100

67 52.7 0.0 104.4

Manual Integrations

APPROVED


Reviewed By :Mahesh
Dadoda



05/30/2025

Supervised By :Semsettin
Yesilyurt

05/30/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.622 min Scan# 1132

Delta R.T. 0.006 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50




Tgt Ion:114 Resp: 359832


Ion Ratio Lower Upper

114 100

63 20.7 0.0 42.6


88 14.1 0.0 29.4





#35
Dibromofluoromethane
Concen: 52.463 ug/l
RT: 7.640 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion: 113 Resp: 11055
Ion Ratio Lower Upper
113 100
111 102.2 82.4 123.6
192 17.8 15.2 22.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#36
1,1-Dichloropropene
Concen: 20.835 ug/l
RT: 7.841 min Scan# 1004
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Tgt Ion: 75 Resp: 69669
Ion Ratio Lower Upper
75 100
110 33.8 17.1 51.2
77 31.5 25.1 37.7





#37

Ethyl Acetate

Concen: 22.255 ug/l

RT: 6.994 min Scan# 80

Delta R.T. 0.006 min

Lab File: VY022456.D


Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y

ClientSampleId :



VY0529SBS01



Tgt Ion:	43	Resp:	3162
Ion Ratio	Lower	Upper	
43	100		
61	12.2	11.7	17.5
70	9.5	7.8	11.6


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda



05/30/2025
Supervised By :Semsettin
Yesilyurt

05/30/2025



#38

Carbon Tetrachloride

Concen: 20.484 ug/l




RT: 7.823 min Scan# 1001


Delta R.T. 0.000 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50

Tgt Ion:	117	Resp:	71355
Ion Ratio	Lower	Upper	
117	100		
119	88.9	76.3	114.5
121	29.8	26.1	39.1





#39

Methylcyclohexane

Concen: 19.975 ug/l

RT: 9.116 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y



ClientSampleId :

VY0529SBSD01

Tgt Ion: 83	Resp: 88788
Ion Ratio	Lower Upper
83 100	
55 71.9	56.2 84.2
98 50.7	38.4 57.6



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda



05/30/2025
Supervised By :Semsettin
Yesilyurt

05/30/2025



#40

Benzene

Concen: 21.352 ug/l




RT: 8.085 min Scan# 1044


Delta R.T. 0.000 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50


Tgt Ion: 78	Resp: 212449
Ion Ratio	Lower Upper
78 100	
77 22.7	19.4 29.2





#41
Methacrylonitrile
Concen: 20.709 ug/l
RT: 7.232 min Scan# 904
Delta R.T. -0.018 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion: 41 Resp: 17862
Ion Ratio Lower Upper
41 100
39 55.1 40.4 60.6
67 77.5 53.0 79.4
52 34.1 22.7 34.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#42
1,2-Dichloroethane
Concen: 21.906 ug/l
RT: 8.165 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion: 62 Resp: 57566
Ion Ratio Lower Upper
62 100
98 9.1 0.0 18.4





#43
Isopropyl Acetate
Concen: 21.973 ug/l
RT: 8.201 min Scan# 1063
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion: 43 Resp: 68235
Ion Ratio Lower Upper
43 100
61 28.6 15.4 23.0
87 12.5 10.6 15.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#44
Trichloroethene
Concen: 20.937 ug/l
RT: 8.872 min Scan# 1173
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Tgt Ion:130 Resp: 51499
Ion Ratio Lower Upper
130 100
95 99.2 0.0 194.8





#45

1,2-Dichloropropane

Concen: 21.930 ug/l

RT: 9.146 min Scan# 1218

Delta R.T. 0.000 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y

ClientSampleId :

VY0529SBSD01

Tgt Ion: 63 Resp: 5126

Ion Ratio Lower Upper



63 100

65 33.4 25.1 37.7

Manual Integrations

APPROVED



Reviewed By :Mahesh
Dadoda



05/30/2025

Supervised By :Semsettin
Yesilyurt

05/30/2025



#46

Dibromomethane

Concen: 21.556 ug/l

RT: 9.238 min Scan# 1233

Delta R.T. 0.006 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50




Tgt Ion: 93 Resp: 28152


Ion Ratio Lower Upper

93 100

95 80.8 66.2 99.2

174 90.5 71.1 106.7





#47

Bromodichloromethane

Concen: 21.486 ug/l

RT: 9.427 min Scan# 1264

Delta R.T. 0.000 min

Lab File: VY022456.D


Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y

ClientSampleId :


VY0529SBS01



Tgt Ion: 83	Resp: 7215
Ion Ratio	Lower Upper
83 100	
85 62.8	49.8 74.6
127 7.8	6.6 9.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda



05/30/2025
Supervised By :Semsettin
Yesilyurt

05/30/2025



#48

Methyl methacrylate

Concen: 21.934 ug/l




RT: 9.225 min Scan# 1231


Delta R.T. 0.000 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50

Tgt Ion: 41	Resp: 31618
Ion Ratio	Lower Upper
41 100	
69 88.2	70.3 105.5
39 68.8	55.4 83.0





#49

1,4-Dioxane

Concen: 459.490 ug/l

RT: 9.238 min Scan# 1233

Delta R.T. 0.000 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y

ClientSampleId :



VY0529SBSD01

Tgt Ion: 88 Resp: 6628

Ion	Ratio	Lower	Upper
88	100		
43	34.7	26.9	40.3
58	68.9	55.8	83.6



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda



05/30/2025
Supervised By :Semsettin
Yesilyurt

05/30/2025



#50

Toluene-d8

Concen: 51.578 ug/l

RT: 10.109 min Scan# 1376




Delta R.T. 0.000 min


Lab File: VY022456.D

Acq: 29 May 2025 09:50

Tgt Ion: 98 Resp: 443079

Ion	Ratio	Lower	Upper
98	100		
100	64.6	52.2	78.2







#51
4-Methyl-2-Pentanone
Concen: 110.718 ug/l
RT: 10.000 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01

Tgt Ion: 43 Resp: 16964
Ion Ratio Lower Upper
43 100
58 38.8 31.2 46.8



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#52
Toluene
Concen: 20.565 ug/l
RT: 10.176 min Scan# 1387
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 92 Resp: 127807
Ion Ratio Lower Upper
92 100
91 176.3 139.8 209.6





#53
t-1,3-Dichloropropene
Concen: 21.159 ug/l
RT: 10.402 min Scan# 1424
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
Client Sample Id :
VY0529SBSD01



Tgt Ion: 75 Resp: 66189
Ion Ratio Lower Upper
75 100
77 30.4 26.2 39.2


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#54
cis-1,3-Dichloropropene
Concen: 21.209 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion: 75 Resp: 77958
Ion Ratio Lower Upper
75 100
77 34.1 25.4 38.2
39 54.6 42.5 63.7





#55
1,1,2-Trichloroethane
Concen: 21.244 ug/l
RT: 10.573 min Scan# 1452
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion	Ratio	Lower	Upper
97	100		
83	95.7	70.2	105.4
85	54.2	46.1	69.1
99	62.3	49.9	74.9


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#56
Ethyl methacrylate
Concen: 21.361 ug/l
RT: 10.445 min Scan# 1431
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion	Ratio	Lower	Upper
69	100		
41	66.4	51.2	76.8
39	45.8	35.2	52.8





#57
1,3-Dichloropropane
Concen: 21.506 ug/l
RT: 10.719 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion: 76 Resp: 62570
Ion Ratio Lower Upper
76 100
78 33.4 25.8 38.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#58
2-Chloroethyl Vinyl ether
Concen: 105.282 ug/l
RT: 9.713 min Scan# 1311
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 63 Resp: 124329
Ion Ratio Lower Upper
63 100
106 26.9 22.2 33.2





#59
2-Hexanone
Concen: 111.915 ug/l
RT: 10.768 min Scan# 1483
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion: 43 Resp: 11237
Ion Ratio Lower Upper
43 100
58 54.1 27.1 81.2


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#60
Dibromochloromethane
Concen: 21.090 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion:129 Resp: 44484
Ion Ratio Lower Upper
129 100
127 78.0 38.3 114.8





#61
1,2-Dibromoethane
Concen: 21.565 ug/l
RT: 11.018 min Scan# 1525
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion:107 Resp: 3256
Ion Ratio Lower Upper
107 100
109 92.2 74.0 111.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#62
4-Bromofluorobenzene
Concen: 49.908 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion: 95 Resp: 129686
Ion Ratio Lower Upper
95 100
174 85.9 0.0 171.6
176 82.7 0.0 164.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 11591
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion:117 Resp: 30584
Ion Ratio Lower Upper
117 100
82 56.2 43.3 64.9
119 31.7 24.5 36.7


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#64
Tetrachloroethene
Concen: 20.904 ug/l
RT: 10.652 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion:164 Resp: 54451
Ion Ratio Lower Upper
164 100
166 124.5 102.0 153.0
129 91.5 71.3 106.9
131 90.9 69.4 104.0





#65
Chlorobenzene
Concen: 20.880 ug/l
RT: 11.444 min Scan# 1595
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion:112 Resp: 136170
Ion Ratio Lower Upper
112 100
114 31.5 26.1 39.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 20.674 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Tgt Ion:131 Resp: 45682
Ion Ratio Lower Upper
131 100
133 95.0 46.6 139.8
119 65.5 33.3 99.8







#67
Ethyl Benzene
Concen: 20.243 ug/l
RT: 11.524 min Scan# 1608
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01

Tgt Ion: 91 Resp: 24512
Ion Ratio Lower Upper
91 100
106 29.8 24.4 36.6



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#68
m/p-Xylenes
Concen: 40.037 ug/l
RT: 11.633 min Scan# 1626
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Tgt Ion:106 Resp: 183812
Ion Ratio Lower Upper
106 100
91 199.6 159.8 239.8







#69
o-Xylene
Concen: 20.343 ug/l
RT: 11.957 min Scan# 1679
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01

Tgt Ion:106 Resp: 87779
Ion Ratio Lower Upper
106 100
91 213.1 106.3 318.9



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#70
Styrene
Concen: 20.424 ug/l
RT: 11.975 min Scan# 1682
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion:104 Resp: 146742
Ion Ratio Lower Upper
104 100
78 48.8 38.8 58.2
103 54.5 43.4 65.2





#71
Bromoform
Concen: 20.865 ug/l
RT: 12.133 min Scan# 1708
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion:173 Resp: 24149
Ion Ratio Lower Upper
173 100
175 46.8 24.3 73.0
254 0.0 0.1 0.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.347 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion:152 Resp: 140548
Ion Ratio Lower Upper
152 100
115 58.3 28.4 85.3
150 164.8 0.0 345.8





#73
Isopropylbenzene
Concen: 20.435 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion:105 Resp: 22604
Ion Ratio Lower Upper
105 100
120 25.7 13.2 39.5


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#74
N-ethyl acetate
Concen: 21.701 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion: 43 Resp: 58739
Ion Ratio Lower Upper
43 100
70 41.5 33.0 49.4
55 22.2 17.5 26.3
61 23.1 19.0 28.6





#75
1,1,2,2-Tetrachloroethane
Concen: 21.786 ug/l
RT: 12.511 min Scan# 1770
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion: 83 Resp: 37969
Ion Ratio Lower Upper
83 100
131 8.7 4.8 14.4
85 63.3 32.0 96.2


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#76
1,2,3-Trichloropropane
Concen: 18.730 ug/l m
RT: 12.560 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Tgt Ion: 75 Resp: 26369
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0







#77
Bromobenzene
Concen: 20.522 ug/l
RT: 12.536 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01

Tgt Ion: 156 Resp: 48852
Ion Ratio Lower Upper
156 100
77 190.4 91.4 274.2
158 96.8 48.1 144.3



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#78
n-propylbenzene
Concen: 20.673 ug/l
RT: 12.597 min Scan# 1784
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 91 Resp: 274435
Ion Ratio Lower Upper
91 100
120 21.6 11.2 33.5





#79
2-Chlorotoluene
Concen: 20.680 ug/l
RT: 12.682 min Scan# 1798
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion: 91 Resp: 148607
Ion Ratio Lower Upper
91 100
126 34.5 17.4 52.3


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#80
1,3,5-Trimethylbenzene
Concen: 20.742 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion:105 Resp: 179270
Ion Ratio Lower Upper
105 100
120 49.2 24.3 73.0





#81
trans-1,4-Dichloro-2-butene
Concen: 21.252 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBS01



Tgt Ion: 75 Resp: 13230
Ion Ratio Lower Upper
75 100
53 94.1 75.1 112.7
89 48.3 37.4 56.0


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#82
4-Chlorotoluene
Concen: 20.594 ug/l
RT: 12.780 min Scan# 1814
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50


Tgt Ion: 91 Resp: 152410
Ion Ratio Lower Upper
91 100
126 34.7 17.2 51.6





#83
tert-Butylbenzene
Concen: 20.194 ug/l
RT: 12.999 min Scan# 1850
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion:119 Resp: 157690
Ion Ratio Lower Upper
119 100
91 65.8 31.8 95.3
134 25.6 13.0 38.9


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#84
1,2,4-Trimethylbenzene
Concen: 20.566 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50




Tgt Ion:105 Resp: 178877
Ion Ratio Lower Upper
105 100
120 45.4 22.9 68.5





#85
 sec-Butylbenzene
 Concen: 20.613 ug/l
 RT: 13.176 min Scan# 1879
 Delta R.T. 0.000 min
 Lab File: VY022456.D
 Acq: 29 May 2025 09:50



Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0529SBSD01



Tgt Ion:105 Resp: 238715
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.7 29.1


Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda




05/30/2025
 Supervised By :Semsettin
 Yesilyurt



05/30/2025




#86
 p-Isopropyltoluene
 Concen: 20.268 ug/l
 RT: 13.292 min Scan# 1898
 Delta R.T. 0.000 min
 Lab File: VY022456.D
 Acq: 29 May 2025 09:50




Tgt Ion:119 Resp: 196000
 Ion Ratio Lower Upper
 119 100
 134 25.9 13.0 39.0
 91 23.1 11.6 34.8





#87
1,3-Dichlorobenzene
Concen: 20.454 ug/l
RT: 13.292 min Scan# 1898
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion:146 Resp: 97960
Ion Ratio Lower Upper
146 100
111 38.1 19.6 58.8
148 63.4 31.7 95.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt



05/30/2025




#88
1,4-Dichlorobenzene
Concen: 21.194 ug/l
RT: 13.371 min Scan# 1911
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Tgt Ion:146 Resp: 97065
Ion Ratio Lower Upper
146 100
111 38.6 18.7 56.1
148 63.4 31.4 94.3





#89

n-Butylbenzene

Concen: 20.785 ug/l

RT: 13.621 min Scan# 1951

Delta R.T. 0.006 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y



ClientSampleId :

VY0529SBS01

Tgt Ion:	91	Resp:	18846
Ion	Ratio	Lower	Upper
91	100		
92	51.5	25.9	77.6
134	24.6	12.4	37.0



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda



05/30/2025
Supervised By :Semsettin
Yesilyurt

05/30/2025



#90

Hexachloroethane

Concen: 20.858 ug/l




RT: 13.883 min Scan# 1995


Delta R.T. 0.000 min

Lab File: VY022456.D

Acq: 29 May 2025 09:50

Tgt Ion:	117	Resp:	40139
Ion	Ratio	Lower	Upper
117	100		
201	66.8	35.0	105.0







#91
1,2-Dichlorobenzene
Concen: 21.142 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01

Tgt Ion:146 Resp: 84578
Ion Ratio Lower Upper
146 100
111 39.6 20.5 61.6
148 63.4 31.9 95.9



Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda






05/30/2025
Supervised By :Semsettin
Yesilyurt


05/30/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 22.291 ug/l
RT: 14.279 min Scan# 2060
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50

Tgt Ion: 75 Resp: 6003
Ion Ratio Lower Upper
75 100
155 79.8 42.3 126.8
157 101.8 54.4 163.3





#93

1,2,4-Trichlorobenzene

Concen: 21.699 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. 0.000 min

Lab File: VY022456.D


Acq: 29 May 2025 09:50

Instrument :

MSVOA_Y

ClientSampleId :


VY0529SBS01



Tgt Ion:180	Resp:	4812
Ion Ratio	Lower	Upper
180	100	
182	90.7	46.2 138.6
145	31.3	14.2 42.8


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda



05/30/2025
Supervised By :Semsettin
Yesilyurt

05/30/2025



#94

Hexachlorobutadiene


Concen: 21.380 ug/l

RT: 15.023 min Scan# 2182



Delta R.T. 0.000 min


Lab File: VY022456.D

Acq: 29 May 2025 09:50




Tgt Ion:225	Resp:	24884
Ion Ratio	Lower	Upper
225	100	
223	63.9	32.6 97.7
227	62.9	33.0 99.0





#95
Naphthalene
Concen: 22.342 ug/l
RT: 15.145 min Scan# 2202
Delta R.T. 0.000 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Instrument :
MSVOA_Y
ClientSampleId :
VY0529SBSD01



Tgt Ion:128 Resp: 9160
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.4
129 11.0 8.5 12.7


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda




05/30/2025
Supervised By :Semsettin
Yesilyurt

05/30/2025



#96
1,2,3-Trichlorobenzene
Concen: 21.968 ug/l
RT: 15.334 min Scan# 2233
Delta R.T. 0.006 min
Lab File: VY022456.D
Acq: 29 May 2025 09:50



Tgt Ion:180 Resp: 39958
Ion Ratio Lower Upper
180 100
182 94.1 48.1 144.3
145 32.2 14.9 44.5

