

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052725

Review By	Amit Patel	Review On	5/28/2025 12:56:58 PM
Supervise By	Maresh Dadoda	Supervise On	5/28/2025 12:57:05 PM
SubDirectory	VY052725	HP Acquire Method	MSVOA_Y HP Processing Method 82y052725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134020
Initial Calibration Stds	VP134021,VP134022,VP134024,VP134026,VP134028,VP134029
CCC	VP134031
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134030
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022420.D	27 May 2025 08:20		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY022421.D	27 May 2025 08:50		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY022422.D	27 May 2025 09:14		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY022423.D	27 May 2025 09:36		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY022424.D	27 May 2025 09:59		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY022425.D	27 May 2025 10:22		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY022426.D	27 May 2025 10:44		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022427.D	27 May 2025 11:20		SY/MD	Ok
9	VSTDICV050	ICVVY052725	VY022428.D	27 May 2025 11:56		SY/MD	Ok,M
10	VY0527SBL01	VY0527SBL01	VY022429.D	27 May 2025 13:12		SY/MD	Ok
11	VY0527SBS01	VY0527SBS01	VY022430.D	27 May 2025 13:35		SY/MD	Ok,M
12	VY0527SBSD01	VY0527SBSD01	VY022431.D	27 May 2025 13:57		SY/MD	Ok,M
13	Q2118-11RE	BP-VPB-182-GW-500-5	VY022432.D	27 May 2025 14:27	vial-B Surrogate Fail	SY/MD	Confirms
14	Q2113-01	HD-02-05222025	VY022433.D	27 May 2025 14:51	vial-B not purged	SY/MD	Not Ok
15	Q2128-03	TP03-MHM-VOC	VY022434.D	27 May 2025 15:14	vial-A	SY/MD	Ok
16	Q2127-11RE	COMP-2RE	VY022435.D	27 May 2025 15:38	vial-B Internal Standard Fail	SY/MD	Confirms
17	Q2130-01	TP-3	VY022436.D	27 May 2025 16:01	vial-A	SY/MD	Ok
18	VIBLK	VIBLK	VY022437.D	27 May 2025 16:24		SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052725

Review By	Amit Patel	Review On	5/28/2025 12:56:58 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/28/2025 12:57:05 PM		
SubDirectory	VY052725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y052725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134020
Initial Calibration Stds	VP134021,VP134022,VP134024,VP134026,VP134028,VP134029
CCC	VP134031
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134030
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VSTDCCC050	VSTDCCC050EC	VY022438.D	27 May 2025 17:10		SY/MD	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052925

Review By	Mahesh Dadoda	Review On	5/30/2025 12:21:56 PM
Supervise By	Semsettin Yesilyurt	Supervise On	5/30/2025 12:28:04 PM
SubDirectory	VY052925	HP Acquire Method	MSVOA_Y HP Processing Method 82y052725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134049
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134050,VP134051 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022452.D	29 May 2025 07:54		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022453.D	29 May 2025 08:23	CCAL failed low for comp. #13	SY/MD	Ok,M
3	VY0529SBL01	VY0529SBL01	VY022454.D	29 May 2025 08:56		SY/MD	Ok
4	VY0529SBS01	VY0529SBS01	VY022455.D	29 May 2025 09:27		SY/MD	Ok,M
5	VY0529SBSD01	VY0529SBSD01	VY022456.D	29 May 2025 09:50		SY/MD	Ok,M
6	Q2126-03	MDL-SOIL-03-QT2-202	VY022457.D	29 May 2025 10:53	CCAL failed low	SY/MD	ReRun
7	Q2126-03	MDL-SOIL-03-QT2-202	VY022458.D	29 May 2025 11:38	CCAL failed low	SY/MD	ReRun
8	Q2147-02	VOC	VY022459.D	29 May 2025 12:02	vial-A Internal Standard Fail;	SY/MD	Ok
9	Q2150-01	TP-44	VY022460.D	29 May 2025 12:25	vial-A	SY/MD	Ok
10	Q2150-02	TP-42	VY022461.D	29 May 2025 12:49	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q2150-03	TP-39	VY022462.D	29 May 2025 13:12	vial-A	SY/MD	Ok
12	Q2150-04	TP-48	VY022463.D	29 May 2025 13:36	vial-A	SY/MD	Ok
13	Q2150-05	TP-47	VY022464.D	29 May 2025 13:59	vial-A Internal Standard Fail; Surrogate fail	SY/MD	ReRun
14	Q2150-06	TP-50	VY022465.D	29 May 2025 14:22	vial-A	SY/MD	Ok
15	Q2150-07	TP-51	VY022466.D	29 May 2025 14:46	vial-A	SY/MD	Ok
16	Q2150-08	TP-52	VY022467.D	29 May 2025 15:09	vial-A	SY/MD	Ok
17	Q2150-09	TP-54	VY022468.D	29 May 2025 15:33	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052925

Review By	Maresh Dadoda	Review On	5/30/2025 12:21:56 PM
Supervise By	Semsettin Yesilyurt	Supervise On	5/30/2025 12:28:04 PM
SubDirectory	VY052925	HP Acquire Method	MSVOA_Y HP Processing Method 82y052725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134049
CCC	VP134050,VP134051
Internal Standard/PEM	VP133934
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q2150-10	TP-53	VY022469.D	29 May 2025 15:56	vial-A	SY/MD	Ok
19	VSTDCCC050	VSTDCCC050EC	VY022470.D	29 May 2025 17:04		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060225

Review By	Mahesh Dadoda	Review On	6/3/2025 2:38:36 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/3/2025 2:39:16 PM		
SubDirectory	VY060225	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y060225S.M

STD. NAME	STD REF.#
Tune/Reschk	VP134084
Initial Calibration Stds	VP134085,VP134086,VP134087,VP134088,VP134089,VP134090
CCC	VP134092,VP134093
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134091
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022488.D	02 Jun 2025 08:31		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022489.D	02 Jun 2025 09:08	Need ICAL	SY/MD	Not Ok
3	VY0602SBL01	VY0602SBL01	VY022490.D	02 Jun 2025 10:38		SY/MD	Not Ok
4	VSTDICC005	VSTDICC005	VY022491.D	02 Jun 2025 11:46		SY/MD	Ok,M
5	VSTDICC010	VSTDICC010	VY022492.D	02 Jun 2025 12:09		SY/MD	Ok,M
6	VSTDICC020	VSTDICC020	VY022493.D	02 Jun 2025 12:32		SY/MD	Ok,M
7	VSTDICCC050	VSTDICCC050	VY022494.D	02 Jun 2025 12:54		SY/MD	Ok,M
8	VSTDICC100	VSTDICC100	VY022495.D	02 Jun 2025 13:17		SY/MD	Ok,M
9	VSTDICC150	VSTDICC150	VY022496.D	02 Jun 2025 13:39		SY/MD	Ok,M
10	VIBLK	VIBLK	VY022497.D	02 Jun 2025 14:03		SY/MD	Ok
11	VSTDICV050	ICVVY060225	VY022498.D	02 Jun 2025 14:59	Icv Fail For DOD For Com.# 18	SY/MD	Ok,M
12	VY0602SBL02	VY0602SBL02	VY022499.D	02 Jun 2025 16:00		SY/MD	Ok
13	VY0602SBS01	VY0602SBS01	VY022500.D	02 Jun 2025 16:24		SY/MD	Ok,M
14	VY0602SBSD01	VY0602SBSD01	VY022501.D	02 Jun 2025 16:47		SY/MD	Ok,M
15	Q2160-02	TP04-MHG-VOC	VY022502.D	02 Jun 2025 17:10	vial A	SY/MD	Ok
16	Q2152-01	OK-02-05292025	VY022503.D	02 Jun 2025 17:34	Not Purged,vial B	SY/MD	Not Ok
17	Q2153-01RE	TR-04-0592025RE	VY022504.D	02 Jun 2025 17:57	Surrogate Fail;Internal Standard Fail, vial B	SY/MD	Confirms

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY060225

Review By	Mahesh Dadoda	Review On	6/3/2025 2:38:36 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/3/2025 2:39:16 PM		
SubDirectory	VY060225	HP Acquire Method	MSVOA_Y	HP Processing Method	82Y060225S.M
STD. NAME		STD REF.#			
Tune/Reschk		VP134084			
Initial Calibration Stds		VP134085,VP134086,VP134087,VP134088,VP134089,VP134090			
CCC		VP134092,VP134093			
Internal Standard/PEM		VP133934			
ICV/I.BLK		VP134091			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2147-02	VOC	VY022505.D	02 Jun 2025 18:21	Not purge,vial B	SY/MD	Not Ok
19	Q2150-02	TP-42	VY022506.D	02 Jun 2025 18:44	vial B	SY/MD	Ok
20	Q2150-05RE	TP-47RE	VY022507.D	02 Jun 2025 19:08	Internal Standard Fail,vial B	SY/MD	Confirms
21	Q2161-01	B27-SOIL-SAMPLE	VY022508.D	02 Jun 2025 19:31	vial-B Not purge	SY/MD	Not Ok
22	VSTDCCC050	VSTDCCC050EC	VY022509.D	02 Jun 2025 19:54		SY/MD	Ok,M

M : Manual Integration