

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150
Lab File ID: VY022453.D Date Analyzed: 05/29/2025
Instrument ID: MSVOA_Y Time Analyzed: 08:23
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	217073	7.72	367403	8.62	311885	11.42
UPPER LIMIT	434146	8.219	734806	9.122	623770	11.92
LOWER LIMIT	108537	7.219	183702	8.122	155943	10.92
EPA SAMPLE NO.						
TP-44	284770	7.71	525823	8.62	402756	11.42
TP-42	51659 *	7.71	80393 *	8.62	54066 *	11.42
TP-39	305036	7.71	550366	8.62	413432	11.42
TP-48	278545	7.72	518246	8.62	419041	11.42
TP-47	18571 *	7.72	25910 *	8.62	13686 *	11.42
TP-50	284911	7.71	538977	8.62	400662	11.42
TP-51	266889	7.71	485834	8.62	362765	11.42
TP-52	286667	7.71	522350	8.62	404746	11.42
TP-54	283443	7.71	517546	8.62	406523	11.41
TP-53	273493	7.71	503324	8.62	399677	11.41
VY0529SBL01	320837	7.72	584950	8.62	455837	11.43
VY0529SBS01	217849	7.72	376540	8.62	319754	11.42
VY0529SBSD01	205776	7.71	359832	8.62	305847	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.