

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150  
 Lab File ID: VY022494.D Date Analyzed: 06/02/2025  
 Instrument ID: MSVOA\_Y Time Analyzed: 12:54  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #  |
|----------------|---------------|-------|---------------|-------|---------------|-------|
| 12 HOUR STD    | 209963        | 7.71  | 350947        | 8.62  | 298187        | 11.42 |
| UPPER LIMIT    | 419926        | 8.213 | 701894        | 9.116 | 596374        | 11.92 |
| LOWER LIMIT    | 104982        | 7.213 | 175474        | 8.116 | 149094        | 10.92 |
| EPA SAMPLE NO. |               |       |               |       |               |       |
| TP-42          | 279093        | 7.71  | 514728        | 8.62  | 411552        | 11.42 |
| TP-47RE        | 60813 *       | 7.71  | 105433 *      | 8.62  | 78237 *       | 11.42 |
| VY0602SBL02    | 309051        | 7.71  | 557091        | 8.62  | 442919        | 11.41 |
| VY0602SBS01    | 201633        | 7.71  | 342745        | 8.62  | 291387        | 11.41 |
| VY0602SBSD01   | 196193        | 7.71  | 338031        | 8.62  | 287481        | 11.41 |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.