

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150
 Lab File ID: VY022494.D Date Analyzed: 06/02/2025
 Instrument ID: MSVOA_Y Time Analyzed: 12:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	139554	13.347				
UPPER LIMIT	279108	13.847				
LOWER LIMIT	69777	12.847				
EPA SAMPLE NO.						
TP-42	142944	13.35				
TP-47RE	24479 *	13.35				
VY0602SBL02	162489	13.35				
VY0602SBS01	136847	13.35				
VY0602SBSD01	133746	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.