

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150
Lab File ID: VY022420.D BFB Injection Date: 05/27/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:20
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.8
75	30.0 - 60.0% of mass 95	58.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	1.1 (1.2) 1
174	50.0 - 100.0% of mass 95	88.8
175	5.0 - 9.0% of mass 174	6.7 (7.6) 1
176	95.0 - 101.0% of mass 174	85.7 (96.4) 1
177	5.0 - 9.0% of mass 176	5.5 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY022421.D	05/27/2025	08:50
VSTDIC010	VSTDIC010	VY022422.D	05/27/2025	09:14
VSTDIC020	VSTDIC020	VY022423.D	05/27/2025	09:36
VSTDIC050	VSTDIC050	VY022424.D	05/27/2025	09:59
VSTDIC100	VSTDIC100	VY022425.D	05/27/2025	10:22
VSTDIC150	VSTDIC150	VY022426.D	05/27/2025	10:44