

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052725

Review By	Amit Patel	Review On	5/28/2025 12:56:58 PM
Supervise By	Mahesh Dadoda	Supervise On	5/28/2025 12:57:05 PM
SubDirectory	VY052725	HP Acquire Method	MSVOA_Y HP Processing Method 82y052725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134020
Initial Calibration Stds	VP134021,VP134022,VP134024,VP134026,VP134028,VP134029
CCC	VP134031
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134030
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022420.D	27 May 2025 08:20		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY022421.D	27 May 2025 08:50		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY022422.D	27 May 2025 09:14		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY022423.D	27 May 2025 09:36		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY022424.D	27 May 2025 09:59		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY022425.D	27 May 2025 10:22		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY022426.D	27 May 2025 10:44		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022427.D	27 May 2025 11:20		SY/MD	Ok
9	VSTDICV050	ICVVY052725	VY022428.D	27 May 2025 11:56		SY/MD	Ok,M
10	VY0527SBL01	VY0527SBL01	VY022429.D	27 May 2025 13:12		SY/MD	Ok
11	VY0527SBS01	VY0527SBS01	VY022430.D	27 May 2025 13:35		SY/MD	Ok,M
12	VY0527SBSD01	VY0527SBSD01	VY022431.D	27 May 2025 13:57		SY/MD	Ok,M
13	Q2118-11RE	BP-VPB-182-GW-500-5	VY022432.D	27 May 2025 14:27	vial-B Surrogate Fail	SY/MD	Confirms
14	Q2113-01	HD-02-05222025	VY022433.D	27 May 2025 14:51	vial-B not purged	SY/MD	Not Ok
15	Q2128-03	TP03-MHM-VOC	VY022434.D	27 May 2025 15:14	vial-A	SY/MD	Ok
16	Q2127-11RE	COMP-2RE	VY022435.D	27 May 2025 15:38	vial-B Internal Standard Fail	SY/MD	Confirms
17	Q2130-01	TP-3	VY022436.D	27 May 2025 16:01	vial-A	SY/MD	Ok
18	VIBLK	VIBLK	VY022437.D	27 May 2025 16:24		SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052725

Review By	Amit Patel	Review On	5/28/2025 12:56:58 PM		
Supervise By	Mahesh Dadoda	Supervise On	5/28/2025 12:57:05 PM		
SubDirectory	VY052725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y052725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134020
Initial Calibration Stds	VP134021,VP134022,VP134024,VP134026,VP134028,VP134029
CCC	VP134031
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134030
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VSTDCCC050	VSTDCCC050EC	VY022438.D	27 May 2025 17:10		SY/MD	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration