

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052925

Review By	Maresh Dadoda	Review On	5/30/2025 12:21:56 PM
Supervise By	Semsettin Yesilyurt	Supervise On	5/30/2025 12:28:04 PM
SubDirectory	VY052925	HP Acquire Method	MSVOA_Y HP Processing Method 82y052725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134049
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134050,VP134051 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022452.D	29 May 2025 07:54		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022453.D	29 May 2025 08:23	CCAL failed low for comp. #13	SY/MD	Ok,M
3	VY0529SBL01	VY0529SBL01	VY022454.D	29 May 2025 08:56		SY/MD	Ok
4	VY0529SBS01	VY0529SBS01	VY022455.D	29 May 2025 09:27		SY/MD	Ok,M
5	VY0529SBSD01	VY0529SBSD01	VY022456.D	29 May 2025 09:50		SY/MD	Ok,M
6	Q2126-03	MDL-SOIL-03-QT2-202	VY022457.D	29 May 2025 10:53	CCAL failed low	SY/MD	ReRun
7	Q2126-03	MDL-SOIL-03-QT2-202	VY022458.D	29 May 2025 11:38	CCAL failed low	SY/MD	ReRun
8	Q2147-02	VOC	VY022459.D	29 May 2025 12:02	vial-A Internal Standard Fail;	SY/MD	Ok
9	Q2150-01	TP-44	VY022460.D	29 May 2025 12:25	vial-A	SY/MD	Ok
10	Q2150-02	TP-42	VY022461.D	29 May 2025 12:49	vial-A Internal Standard Fail	SY/MD	ReRun
11	Q2150-03	TP-39	VY022462.D	29 May 2025 13:12	vial-A	SY/MD	Ok
12	Q2150-04	TP-48	VY022463.D	29 May 2025 13:36	vial-A	SY/MD	Ok
13	Q2150-05	TP-47	VY022464.D	29 May 2025 13:59	vial-A Internal Standard Fail; Surrogate fail	SY/MD	ReRun
14	Q2150-06	TP-50	VY022465.D	29 May 2025 14:22	vial-A	SY/MD	Ok
15	Q2150-07	TP-51	VY022466.D	29 May 2025 14:46	vial-A	SY/MD	Ok
16	Q2150-08	TP-52	VY022467.D	29 May 2025 15:09	vial-A	SY/MD	Ok
17	Q2150-09	TP-54	VY022468.D	29 May 2025 15:33	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052925

Review By	Mahesh Dadoda	Review On	5/30/2025 12:21:56 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/30/2025 12:28:04 PM		
SubDirectory	VY052925	HP Acquire Method	MSVOA_Y	HP Processing Method	82y052725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134049
CCC	VP134050,VP134051
Internal Standard/PEM	VP133934
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q2150-10	TP-53	VY022469.D	29 May 2025 15:56	vial-A	SY/MD	Ok
19	VSTDCCC050	VSTDCCC050EC	VY022470.D	29 May 2025 17:04		SY/MD	Ok,M

M : Manual Integration