

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: SCAL01

Lab Code: CHEM

SAS No.: Q2161 SDG NO.: Q2161

Lab File ID: BP024887.D

DFTPP Injection Date: 06/10/2025

Instrument ID: BNA_P

DFTPP Injection Time: 10:12

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.5
68	Less than 2.0% of mass 69	0.6 (1.8) 1
69	Mass 69 relative abundance	35
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	46.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	30.9
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	12
442	Greater than 50% of mass 198	78.8
443	15.0 - 24.0% of mass 442	15.3 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024888.D	06/10/2025	10:53
B28-SOIL-SAMPLE	Q2161-02	BP024896.D	06/10/2025	16:27
B27-SOIL-SAMPLE	Q2161-01	BP024902.D	06/10/2025	20:32