

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: SCAL01

Lab Code: CHEM Case No.: Q2161 SAS No.: Q2161 SDG No.: Q2161

Instrument ID: MSVOA_X Calibration Date/Time: 06/02/2025 10:21

Lab File ID: VX046433.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.765	0.849		10.98	20
Chloromethane	0.742	0.817	0.1	10.11	20
Vinyl Chloride	0.691	0.734		6.22	20
Bromomethane	0.320	0.335		4.69	20
Chloroethane	0.369	0.412		11.65	20
Trichlorofluoromethane	1.021	1.136		11.26	20
1,1,2-Trichlorotrifluoroethane	0.632	0.722		14.24	20
1,1-Dichloroethene	0.593	0.650		9.61	20
Acetone	0.374	0.439		17.38	20
Carbon Disulfide	1.406	1.520		8.11	20
Methyl tert-butyl Ether	2.079	2.353		13.18	20
Methyl Acetate	0.867	1.189		37.14	20
Methylene Chloride	0.716	0.757		5.73	20
trans-1,2-Dichloroethene	0.596	0.645		8.22	20
1,1-Dichloroethane	1.219	1.379	0.1	13.13	20
Cyclohexane	1.111	1.177		5.94	20
2-Butanone	0.543	0.604		11.23	20
Carbon Tetrachloride	0.544	0.625		14.89	20
cis-1,2-Dichloroethene	0.718	0.805		12.12	20
Bromochloromethane	0.587	0.610		3.92	20
Chloroform	1.271	1.419		11.64	20
1,1,1-Trichloroethane	1.101	1.237		12.35	20
Methylcyclohexane	0.623	0.681		9.31	20
Benzene	1.417	1.575		11.15	20
1,2-Dichloroethane	0.612	0.699		14.22	20
Trichloroethene	0.341	0.385		12.9	20
1,2-Dichloropropane	0.352	0.411		16.76	20
Bromodichloromethane	0.547	0.648		18.46	20
4-Methyl-2-Pentanone	0.605	0.682		12.73	20
Toluene	0.869	0.958		10.24	20
t-1,3-Dichloropropene	0.487	0.591		21.35	20
cis-1,3-Dichloropropene	0.538	0.646		20.07	20
1,1,2-Trichloroethane	0.343	0.382		11.37	20
2-Hexanone	0.448	0.509		13.62	20
Dibromochloromethane	0.376	0.460		22.34	20
1,2-Dibromoethane	0.356	0.395		10.95	20
Tetrachloroethene	0.354	0.389		9.89	20
Chlorobenzene	1.094	1.197	0.3	9.41	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: SCAL01

Lab Code: CHEM Case No.: Q2161 SAS No.: Q2161 SDG No.: Q2161

Instrument ID: MSVOA_X Calibration Date/Time: 06/02/2025 10:21

Lab File ID: VX046433.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	2.190		13.53	20
m/p-Xylenes	0.706	0.792		12.18	20
o-Xylene	0.688	0.779		13.23	20
Styrene	1.127	1.328		17.83	20
Bromoform	0.281	0.340	0.1	21	20
Isopropylbenzene	3.893	4.366		12.15	20
1,1,2,2-Tetrachloroethane	1.364	1.417	0.3	3.89	20
1,3-Dichlorobenzene	1.649	1.823		10.55	20
1,4-Dichlorobenzene	1.684	1.827		8.49	20
1,2-Dichlorobenzene	1.655	1.818		9.85	20
1,2-Dibromo-3-Chloropropane	0.302	0.337		11.59	20
1,2,4-Trichlorobenzene	0.951	1.078		13.35	20
1,2,3-Trichlorobenzene	0.981	1.103		12.44	20
1,2-Dichloroethane-d4	0.932	0.936		0.43	20
Dibromofluoromethane	0.360	0.394		9.44	20
Toluene-d8	1.246	1.254		0.64	20
4-Bromofluorobenzene	0.478	0.507		6.07	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.