

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2161

Client: Scalamandre – Tully JV

Analytical Method: SW8260D

Datafile : VY022474.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0530SBS01	Dichlorodifluoromethane	20	23.7	ug/Kg	119			64	136	
	Chloromethane	20	29.4	ug/Kg	147			52	151	
	Vinyl chloride	20	27.2	ug/Kg	136			56	148	
	Bromomethane	20	28.7	ug/Kg	144		*	58	141	
	Chloroethane	20	27.2	ug/Kg	136		*	69	130	
	Trichlorofluoromethane	20	26.1	ug/Kg	131			69	134	
	1,1,2-Trichlorotrifluoroethane	20	22.6	ug/Kg	113			81	123	
	1,1-Dichloroethene	20	23.1	ug/Kg	116			79	121	
	Acetone	100	110	ug/Kg	110			40	171	
	Carbon disulfide	20	22.3	ug/Kg	112			59	130	
	Methyl tert-butyl Ether	20	23.0	ug/Kg	115			77	129	
	Methyl Acetate	20	25.3	ug/Kg	127			69	149	
	Methylene Chloride	20	24.1	ug/Kg	121			72	131	
	trans-1,2-Dichloroethene	20	23.0	ug/Kg	115			80	123	
	1,1-Dichloroethane	20	23.9	ug/Kg	119			82	123	
	Cyclohexane	20	22.0	ug/Kg	110			76	122	
	2-Butanone	100	110	ug/Kg	110			69	131	
	Carbon Tetrachloride	20	21.3	ug/Kg	106			76	129	
	cis-1,2-Dichloroethene	20	22.9	ug/Kg	115			82	123	
	Bromochloromethane	20	24.0	ug/Kg	120			80	127	
	Chloroform	20	24.0	ug/Kg	120			82	125	
	1,1,1-Trichloroethane	20	22.8	ug/Kg	114			80	126	
	Methylcyclohexane	20	20.7	ug/Kg	104			77	123	
	Benzene	20	21.7	ug/Kg	109			84	121	
	1,2-Dichloroethane	20	21.8	ug/Kg	109			81	126	
	Trichloroethene	20	20.9	ug/Kg	104			83	122	
	1,2-Dichloropropane	20	21.5	ug/Kg	108			83	122	
	Bromodichloromethane	20	21.8	ug/Kg	109			82	123	
	4-Methyl-2-Pentanone	100	100	ug/Kg	100			70	135	
	Toluene	20	21.5	ug/Kg	108			83	122	
	t-1,3-Dichloropropene	20	21.1	ug/Kg	106			78	124	
	cis-1,3-Dichloropropene	20	20.9	ug/Kg	104			81	122	
	1,1,2-Trichloroethane	20	20.6	ug/Kg	103			82	125	
	2-Hexanone	100	96.5	ug/Kg	97			66	138	
	Dibromochloromethane	20	21.0	ug/Kg	105			79	125	
	1,2-Dibromoethane	20	20.4	ug/Kg	102			80	125	
	Tetrachloroethene	20	20.5	ug/Kg	103			83	125	
	Chlorobenzene	20	21.3	ug/Kg	106			84	122	
	Ethyl Benzene	20	20.8	ug/Kg	104			82	124	
	m/p-Xylenes	40	41.3	ug/Kg	103			83	124	
	o-Xylene	20	21.2	ug/Kg	106			83	123	
	Styrene	20	20.6	ug/Kg	103			82	124	
	Bromoform	20	20.6	ug/Kg	103			75	127	
	Isopropylbenzene	20	21.3	ug/Kg	106			82	124	
	1,1,2,2-Tetrachloroethane	20	21.3	ug/Kg	106			77	127	
	1,3-Dichlorobenzene	20	21.0	ug/Kg	105			83	122	
	1,4-Dichlorobenzene	20	21.3	ug/Kg	106			84	121	
	1,2-Dichlorobenzene	20	21.5	ug/Kg	108			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2161

Client: Scalamandre – Tully JV

Analytical Method: SW8260D

Datafile : VY022474.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0530SBS01	1,2-Dibromo-3-Chloropropane	20	20.0	ug/Kg	100			66	134	
	1,2,4-Trichlorobenzene	20	20.2	ug/Kg	101			78	127	
	1,2,3-Trichlorobenzene	20	20.8	ug/Kg	104			70	137	