

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCAL01
 Lab Code: CHEM Case No.: Q2161 SAS No.: Q2161 SDG NO.: Q2161
 Lab File ID: VY022472.D Date Analyzed: 05/30/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:20
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	141076	13.347				
UPPER LIMIT	282152	13.847				
LOWER LIMIT	70538	12.847				
EPA SAMPLE NO.						
B28-SOIL-SAMPLE	2932 *	13.35				
VY0530SBL01	145190	13.35				
VY0530SBS01	126345	13.35				
VY0530SBSD01	129296	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.