

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: SCAL01  
 Lab Code: CHEM Case No.: Q2161 SAS No.: Q2161 SDG No.: Q2161  
 Instrument ID: MSVOA\_Y Calibration Date(s): 05/27/2025 05/27/2025  
 Heated Purge: (Y/N) Y Calibration Time(s): 08:50 10:44  
 GC Column: RXI-624 ID: 0.25 (mm)

|                                |                     |        |        |                     |        |                     |       |       |  |
|--------------------------------|---------------------|--------|--------|---------------------|--------|---------------------|-------|-------|--|
| LAB FILE ID:                   | RRF005 = VY022421.D |        |        | RRF010 = VY022422.D |        | RRF020 = VY022423.D |       |       |  |
|                                | RRF050 = VY022424.D |        |        | RRF100 = VY022425.D |        | RRF150 = VY022426.D |       |       |  |
| COMPOUND                       | RRF005              | RRF010 | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |  |
| Dichlorodifluoromethane        | 0.544               | 0.509  | 0.526  | 0.423               | 0.430  | 0.429               | 0.477 | 11.6  |  |
| Chloromethane                  | 1.364               | 1.331  | 1.300  | 1.196               | 0.936  | 0.912               | 1.173 | 17.1  |  |
| Vinyl Chloride                 | 1.535               | 1.493  | 1.522  | 1.453               | 1.292  | 1.257               | 1.425 | 8.5   |  |
| Bromomethane                   | 1.261               | 1.281  | 1.397  | 1.275               | 1.213  | 1.345               | 1.295 | 5.1   |  |
| Chloroethane                   | 0.930               | 0.881  | 0.949  | 1.015               | 0.907  | 0.907               | 0.932 | 5     |  |
| Trichlorofluoromethane         | 1.243               | 1.174  | 1.222  | 1.197               | 1.187  | 1.193               | 1.203 | 2.1   |  |
| 1,1,2-Trichlorotrifluoroethane | 0.552               | 0.517  | 0.541  | 0.520               | 0.523  | 0.512               | 0.528 | 2.9   |  |
| 1,1-Dichloroethene             | 0.552               | 0.512  | 0.533  | 0.511               | 0.520  | 0.515               | 0.524 | 3     |  |
| Acetone                        | 0.094               | 0.090  | 0.081  | 0.085               | 0.086  | 0.083               | 0.087 | 5.6   |  |
| Carbon Disulfide               | 1.672               | 1.675  | 1.707  | 1.653               | 1.672  | 1.650               | 1.672 | 1.2   |  |
| Methyl tert-butyl Ether        | 1.353               | 1.359  | 1.357  | 1.386               | 1.445  | 1.445               | 1.391 | 3.1   |  |
| Methyl Acetate                 | 0.288               | 0.319  | 0.296  | 0.308               | 0.351  | 0.383               | 0.324 | 11.2  |  |
| Methylene Chloride             | 0.778               | 0.669  | 0.578  | 0.549               | 0.550  | 0.543               | 0.611 | 15.5  |  |
| trans-1,2-Dichloroethene       | 0.595               | 0.556  | 0.555  | 0.561               | 0.582  | 0.582               | 0.572 | 2.9   |  |
| 1,1-Dichloroethane             | 1.000               | 0.986  | 1.035  | 1.027               | 1.062  | 1.056               | 1.028 | 2.9   |  |
| Cyclohexane                    | 1.220               | 1.044  | 0.997  | 0.939               | 0.947  | 0.931               | 1.013 | 10.9  |  |
| 2-Butanone                     | 0.137               | 0.144  | 0.136  | 0.142               | 0.152  | 0.154               | 0.144 | 5.2   |  |
| Carbon Tetrachloride           | 0.473               | 0.456  | 0.471  | 0.481               | 0.510  | 0.512               | 0.484 | 4.7   |  |
| cis-1,2-Dichloroethene         | 0.655               | 0.647  | 0.650  | 0.652               | 0.684  | 0.685               | 0.662 | 2.6   |  |
| Bromochloromethane             | 0.400               | 0.426  | 0.435  | 0.439               | 0.439  | 0.437               | 0.429 | 3.5   |  |
| Chloroform                     | 0.966               | 0.970  | 1.023  | 1.026               | 1.052  | 1.038               | 1.013 | 3.5   |  |
| 1,1,1-Trichloroethane          | 0.900               | 0.894  | 0.916  | 0.913               | 0.952  | 0.940               | 0.919 | 2.5   |  |
| Methylcyclohexane              | 0.611               | 0.602  | 0.603  | 0.614               | 0.638  | 0.636               | 0.618 | 2.6   |  |
| Benzene                        | 1.308               | 1.312  | 1.360  | 1.391               | 1.463  | 1.461               | 1.383 | 5     |  |
| 1,2-Dichloroethane             | 0.348               | 0.348  | 0.360  | 0.369               | 0.383  | 0.384               | 0.365 | 4.5   |  |
| Trichloroethene                | 0.334               | 0.325  | 0.344  | 0.343               | 0.357  | 0.347               | 0.342 | 3.3   |  |
| 1,2-Dichloropropane            | 0.321               | 0.307  | 0.318  | 0.323               | 0.342  | 0.339               | 0.325 | 4     |  |
| Bromodichloromethane           | 0.452               | 0.449  | 0.449  | 0.464               | 0.493  | 0.494               | 0.467 | 4.5   |  |
| 4-Methyl-2-Pentanone           | 0.185               | 0.197  | 0.198  | 0.218               | 0.237  | 0.242               | 0.213 | 10.8  |  |
| Toluene                        | 0.782               | 0.805  | 0.842  | 0.868               | 0.929  | 0.957               | 0.864 | 7.9   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: SCAL01

Lab Code: CHEM Case No.: Q2161 SAS No.: Q2161 SDG No.: Q2161

Instrument ID: MSVOA\_Y Calibration Date(s): 05/27/2025 05/27/2025

Heated Purge: (Y/N) Y Calibration Time(s): 08:50 10:44

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                |        | RRF005 = VY022421.D |        | RRF010 = VY022422.D |        | RRF020 = VY022423.D |       |       |  |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                             |        | RRF050 = VY022424.D |        | RRF100 = VY022425.D |        | RRF150 = VY022426.D |       |       |  |
| COMPOUND                    | RRF005 | RRF010              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |  |
| t-1,3-Dichloropropene       | 0.400  | 0.414               | 0.416  | 0.434               | 0.472  | 0.473               | 0.435 | 7.1   |  |
| cis-1,3-Dichloropropene     | 0.477  | 0.480               | 0.504  | 0.508               | 0.549  | 0.547               | 0.511 | 6.1   |  |
| 1,1,2-Trichloroethane       | 0.219  | 0.216               | 0.226  | 0.229               | 0.243  | 0.243               | 0.229 | 5.1   |  |
| 2-Hexanone                  | 0.121  | 0.125               | 0.129  | 0.145               | 0.157  | 0.159               | 0.140 | 12    |  |
| Dibromochloromethane        | 0.258  | 0.274               | 0.284  | 0.302               | 0.320  | 0.321               | 0.293 | 8.7   |  |
| 1,2-Dibromoethane           | 0.199  | 0.189               | 0.207  | 0.215               | 0.223  | 0.225               | 0.210 | 6.8   |  |
| Tetrachloroethene           | 0.433  | 0.417               | 0.433  | 0.429               | 0.429  | 0.413               | 0.426 | 2.1   |  |
| Chlorobenzene               | 1.020  | 1.000               | 1.051  | 1.067               | 1.132  | 1.127               | 1.066 | 5.1   |  |
| Ethyl Benzene               | 1.823  | 1.828               | 1.920  | 1.980               | 2.142  | 2.185               | 1.980 | 7.8   |  |
| m/p-Xylenes                 | 0.698  | 0.660               | 0.723  | 0.756               | 0.825  | 0.841               | 0.751 | 9.5   |  |
| o-Xylene                    | 0.654  | 0.637               | 0.673  | 0.709               | 0.768  | 0.792               | 0.705 | 8.9   |  |
| Styrene                     | 1.037  | 1.020               | 1.117  | 1.197               | 1.317  | 1.359               | 1.175 | 12.1  |  |
| Bromoform                   | 0.169  | 0.176               | 0.184  | 0.192               | 0.206  | 0.210               | 0.189 | 8.6   |  |
| Isopropylbenzene            | 3.863  | 3.741               | 3.946  | 3.871               | 4.105  | 4.086               | 3.935 | 3.6   |  |
| 1,1,2,2-Tetrachloroethane   | 0.623  | 0.574               | 0.604  | 0.615               | 0.652  | 0.653               | 0.620 | 4.8   |  |
| 1,3-Dichlorobenzene         | 1.587  | 1.558               | 1.674  | 1.695               | 1.836  | 1.874               | 1.704 | 7.5   |  |
| 1,4-Dichlorobenzene         | 1.596  | 1.527               | 1.617  | 1.642               | 1.707  | 1.687               | 1.629 | 4     |  |
| 1,2-Dichlorobenzene         | 1.383  | 1.332               | 1.431  | 1.416               | 1.495  | 1.482               | 1.423 | 4.3   |  |
| 1,2-Dibromo-3-Chloropropane | 0.103  | 0.093               | 0.096  | 0.093               | 0.093  | 0.096               | 0.096 | 3.8   |  |
| 1,2,4-Trichlorobenzene      | 0.714  | 0.724               | 0.771  | 0.828               | 0.842  | 0.856               | 0.789 | 7.8   |  |
| 1,2,3-Trichlorobenzene      | 0.575  | 0.554               | 0.644  | 0.684               | 0.703  | 0.722               | 0.647 | 10.7  |  |
| 1,2-Dichloroethane-d4       | 0.507  | 0.525               | 0.517  | 0.532               | 0.541  | 0.538               | 0.527 | 2.5   |  |
| Dibromofluoromethane        | 0.287  | 0.281               | 0.283  | 0.290               | 0.309  | 0.308               | 0.293 | 4.3   |  |
| Toluene-d8                  | 1.159  | 1.133               | 1.168  | 1.176               | 1.274  | 1.251               | 1.194 | 4.7   |  |
| 4-Bromofluorobenzene        | 0.377  | 0.331               | 0.345  | 0.355               | 0.382  | 0.376               | 0.361 | 5.7   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.