

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037150.D
 Acq On : 03 Jun 2025 15:53
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060325

Quant Time: Jun 04 01:54:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	116	0.00
3	n-Nitrosodimethylamine	0.400	0.390	2.5	106	0.00
4 S	2-Fluorophenol	0.400	0.361	9.8	102	0.00
5 S	Phenol-d6	0.400	0.358	10.5	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.421	-5.2	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	113	0.00
9	Naphthalene	0.400	0.394	1.5	115	0.01
10	Hexachlorobutadiene	0.400	0.402	-0.5	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	111	0.00
12	2-Methylnaphthalene	0.400	0.352	12.0	106	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	115	0.01
14 S	2,4,6-Tribromophenol	0.400	0.329	17.8	105	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	121	0.00
16	Acenaphthylene	0.400	0.404	-1.0	129	0.00
17	Acenaphthene	0.400	0.373	6.8	118	0.00
18	Fluorene	0.400	0.373	6.8	118	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.462	-15.5	128	0.00
21	4-Bromophenyl-phenylether	0.400	0.372	7.0	116	0.00
22	Hexachlorobenzene	0.400	0.389	2.8	119	0.00
23	Atrazine	0.400	0.374	6.5	126	0.00
24	Pentachlorophenol	0.400	0.436	-9.0	118	0.00
25	Phenanthrene	0.400	0.384	4.0	121	0.00
26	Anthracene	0.400	0.380	5.0	126	0.00
27 SURR	Fluoranthene-d10	0.400	0.375	6.3	114	0.00
28	Fluoranthene	0.400	0.361	9.8	118	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
30	Pyrene	0.400	0.389	2.8	118	0.00
31 S	Terphenyl-d14	0.400	0.415	-3.7	124	0.00
32	Benzo(a)anthracene	0.400	0.383	4.3	122	0.00
33	Chrysene	0.400	0.385	3.8	120	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.371	7.3	125	0.00
35 I	Perylene-d12	0.400	0.400	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.385	3.8	107	0.00
37	Benzo(b)fluoranthene	0.400	0.378	5.5	113	0.00
38	Benzo(k)fluoranthene	0.400	0.399	0.3	118	0.00
39 C	Benzo(a)pyrene	0.400	0.409	-2.2	119	0.00
40	Dibenzo(a,h)anthracene	0.400	0.387	3.3	108	0.00
41	Benzo(g,h,i)perylene	0.400	0.368	8.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN060325