

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037156.D
 Acq On : 03 Jun 2025 20:49
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 04 02:17:21 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	136	0.00
2	1,4-Dioxane	0.533	0.474	11.1	126	0.00
3	n-Nitrosodimethylamine	1.071	0.970	9.4	124	0.00
4 S	2-Fluorophenol	0.989	0.965	2.4	139	0.00
5 S	Phenol-d6	1.199	1.176	1.9	142	0.00
6	bis(2-Chloroethyl)ether	1.144	1.189	-3.9	143	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	143	0.00
8 S	Nitrobenzene-d5	0.422	0.417	1.2	142	0.00
9	Naphthalene	1.154	1.128	2.3	144	0.00
10	Hexachlorobutadiene	0.251	0.245	2.4	134	0.00
11 SURR	2-Methylnaphthalene-d10	0.557	0.552	0.9	140	0.00
12	2-Methylnaphthalene	0.740	0.689	6.9	142	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	137	0.01
14 S	2,4,6-Tribromophenol	0.161	0.134	16.8	126	0.00
15 S	2-Fluorobiphenyl	1.705	1.640	3.8	138	0.00
16	Acenaphthylene	1.961	1.749	10.8	136	0.00
17	Acenaphthene	1.273	1.140	10.4	135	0.00
18	Fluorene	1.674	1.442	13.9	130	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	128	0.00
20	4,6-Dinitro-2-methylphenol	0.077	0.062	19.5	159#	-0.01
21	4-Bromophenyl-phenylether	0.262	0.237	9.5	124	0.00
22	Hexachlorobenzene	0.283	0.276	2.5	131	0.00
23	Atrazine	0.216	0.178	17.6	122	0.00
24	Pentachlorophenol	0.124	0.111	10.5	154#	-0.01
25	Phenanthrene	1.296	1.175	9.3	126	0.00
26	Anthracene	1.183	1.023	13.5	126	0.00
27 SURR	Fluoranthene-d10	1.016	0.920	9.4	121	0.00
28	Fluoranthene	1.432	1.218	14.9	122	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
30	Pyrene	1.953	1.803	7.7	122	0.00
31 S	Terphenyl-d14	0.942	0.890	5.5	123	0.00
32	Benzo(a)anthracene	1.448	1.351	6.7	129	0.00
33	Chrysene	1.612	1.458	9.6	122	0.00
34	Bis(2-ethylhexyl)phthalate	0.914	0.774	15.3	124	0.00
35 I	Perylene-d12	1.000	1.000	0.0	122	0.00
36	Indeno(1,2,3-cd)pyrene	1.591	1.502	5.6	122	0.00
37	Benzo(b)fluoranthene	1.615	1.472	8.9	126	0.00
38	Benzo(k)fluoranthene	1.648	1.577	4.3	132	0.00
39 C	Benzo(a)pyrene	1.352	1.263	6.6	126	0.00
40	Dibenzo(a,h)anthracene	1.227	1.171	4.6	123	0.00
41	Benzo(g,h,i)perylene	1.410	1.318	6.5	119	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4