

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	05/27/25
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	05/29/25
Client Sample ID:	BP-VPB-182-GW-580-582	SDG No.:	Q2162
Lab Sample ID:	Q2162-03	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037158.D	1	06/02/25 08:55	06/03/25 22:01	PB168238

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	2.00	U	0.66	2.00	2.00	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.29		30 - 150		72%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.087	*	30 - 150		22%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.22		55 - 111		56%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.18	*	53 - 106		44%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.12	*	58 - 132		29%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2100		7.589			
1146-65-2	Naphthalene-d8	5800		10.372			
15067-26-2	Acenaphthene-d10	4260		14.235			
1517-22-2	Phenanthrene-d10	7770		16.984			
1719-03-5	Chrysene-d12	5560		21.18			
1520-96-3	Perylene-d12	5160		23.371			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products