

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN060325

Review By	Rahul	Review On	6/4/2025 11:44:25 AM
Supervise By	Jagrut	Supervise On	6/5/2025 10:56:16 AM
SubDirectory	BN060325	HP Acquire Method	BNA_N, 8270_HP Processing Method bn060325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6779 SP6740,1ul/100ul sample SP6768

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037142.D	03 Jun 2025 10:21		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN037143.D	03 Jun 2025 11:39	Compound #20,24 removed from 0.1 PPM	RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN037144.D	03 Jun 2025 12:15		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN037145.D	03 Jun 2025 12:51	Compound #20,24 kept on LR.	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN037146.D	03 Jun 2025 13:26		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN037147.D	03 Jun 2025 14:02		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN037148.D	03 Jun 2025 14:38	Method is good for DOD and NONDOD.	RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN037149.D	03 Jun 2025 15:14		RC/JU	Ok
9	SSTDICV0.4	ICVBN060325	BN037150.D	03 Jun 2025 15:53		RC/JU	Ok
10	PB168238BL	PB168238BL	BN037151.D	03 Jun 2025 17:05	Not Used	RC/JU	Not Ok
11	Q2181-01	38072-062624	BN037152.D	03 Jun 2025 17:41	Need 50X Dilution	RC/JU	Dilution
12	Q2181-01DL	38072-062624DL	BN037153.D	03 Jun 2025 18:18		RC/JU	Ok
13	SSTDCCC0.4	SSTDCCC0.4EC	BN037154.D	03 Jun 2025 18:54		RC/JU	Ok
14	DFTPP	DFTPP	BN037155.D	03 Jun 2025 20:10		RC/JU	Ok
15	SSTDCCC0.4	SSTDCCC0.4	BN037156.D	03 Jun 2025 20:49		RC/JU	Ok
16	PB168238BL	PB168238BL	BN037157.D	03 Jun 2025 21:25	Not Used	RC/JU	Not Ok
17	Q2162-03	BP-VPB-182-GW-580-5	BN037158.D	03 Jun 2025 22:01		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN060325

Review By	Rahul	Review On	6/4/2025 11:44:25 AM
Supervise By	Jagrut	Supervise On	6/5/2025 10:56:16 AM
SubDirectory	BN060325	HP Acquire Method	BNA_N, 8270_HP Processing Method bn060325
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2162-07	BP-VPB-182-GW-620-6	BN037159.D	03 Jun 2025 22:37		RC/JU	Ok
19	Q2162-09	BP-VPB-182-DUP-2025	BN037160.D	03 Jun 2025 23:13		RC/JU	Ok
20	Q2162-10	BP-VPB-182-EB-20250	BN037161.D	03 Jun 2025 23:49		RC/JU	Ok
21	PB168238BS	PB168238BS	BN037162.D	04 Jun 2025 00:25	Recovery Fail for 1,4 Dioxane from low side	RC/JU	Not Ok
22	PB168238BSD	PB168238BSD	BN037163.D	04 Jun 2025 01:01	Recovery Fail for 1,4 Dioxane from low side	RC/JU	Not Ok
23	SSTDCCC0.4	SSTDCCC0.4EC	BN037164.D	04 Jun 2025 02:13		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN060425

Review By	Rahul	Review On	6/4/2025 4:55:38 PM
Supervise By	Jagrut	Supervise On	6/5/2025 10:47:25 AM
SubDirectory	BN060425	HP Acquire Method	BNA_N, 8270_HP Processing Method bn060325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775
CCC	SP6779
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6768
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037165.D	04 Jun 2025 09:04		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037166.D	04 Jun 2025 10:27		RC/JU	Ok
3	PB168238BL	PB168238BL	BN037167.D	04 Jun 2025 11:03		RC/JU	Ok
4	PB168238BS	PB168238BS	BN037168.D	04 Jun 2025 12:27		RC/JU	Ok,M
5	PB168238BSD	PB168238BSD	BN037169.D	04 Jun 2025 13:03		RC/JU	Ok,M
6	SSTDCCC0.4	SSTDCCC0.4EC	BN037170.D	04 Jun 2025 13:53		RC/JU	Ok

M : Manual Integration