

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
 Data File : VY022498.D
 Acq On : 02 Jun 2025 14:59
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY060225

Quant Time: Jun 03 03:24:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.448	0.383	14.5	97	0.00
3 P	Chloromethane	1.322	1.259	4.8	107	0.00
4 C	Vinyl Chloride	1.534	1.475	3.8#	103	0.00
5 T	Bromomethane	1.470	1.250	15.0	105	0.00
6 T	Chloroethane	1.053	0.939	10.8	102	0.00
7 T	Trichlorofluoromethane	1.302	1.237	5.0	105	0.00
8 T	Diethyl Ether	0.293	0.260	11.3	93	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.549	0.496	9.7	96	0.00
10 T	Methyl Iodide	0.594	0.533	10.3	86	0.00
11 T	Tert butyl alcohol	0.040	0.033	17.5	87	0.00
12 CM	1,1-Dichloroethene	0.524	0.485	7.4#	98	-0.01
13 T	Acrolein	0.050	0.044	12.0	97	0.00
14 T	Allyl chloride	0.823	0.755	8.3	98	0.00
15 T	Acrylonitrile	0.124	0.108	12.9	90	0.00
16 T	Acetone	0.114	0.101	11.4	91	0.00
17 T	Carbon Disulfide	1.679	1.559	7.1	98	0.00
18 T	Methyl Acetate	0.336	0.262	22.0	82	0.00
19 T	Methyl tert-butyl Ether	1.477	1.299	12.1	92	-0.01
20 T	Methylene Chloride	0.645	0.527	18.3	96	0.00
21 T	trans-1,2-Dichloroethene	0.587	0.540	8.0	99	0.00
22 T	Diisopropyl ether	1.832	1.616	11.8	93	0.00
23 T	Vinyl Acetate	1.084	0.971	10.4	94	0.00
24 P	1,1-Dichloroethane	1.079	0.994	7.9	98	0.00
25 T	2-Butanone	0.164	0.145	11.6	90	0.00
26 T	2,2-Dichloropropane	0.954	0.890	6.7	98	0.00
27 T	cis-1,2-Dichloroethene	0.678	0.624	8.0	96	0.00
28 T	Bromochloromethane	0.471	0.438	7.0	101	0.00
29 T	Tetrahydrofuran	0.106	0.089	16.0	89	0.00
30 C	Chloroform	1.069	0.982	8.1#	97	0.00
31 T	Cyclohexane	1.075	0.940	12.6	97	0.00
32 T	1,1,1-Trichloroethane	0.949	0.885	6.7	97	0.00
33 S	1,2-Dichloroethane-d4	0.559	0.487	12.9	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.298	0.279	6.4	90	0.00
36 T	1,1-Dichloropropene	0.484	0.455	6.0	97	0.00
37 T	Ethyl Acetate	0.216	0.188	13.0	85	0.00
38 T	Carbon Tetrachloride	0.497	0.475	4.4	97	0.00
39 T	Methylcyclohexane	0.651	0.611	6.1	97	0.00
40 TM	Benzene	1.431	1.345	6.0	97	0.00
41 T	Methacrylonitrile	0.126	0.109	13.5	91	-0.01
42 TM	1,2-Dichloroethane	0.391	0.357	8.7	95	0.00
43 T	Isopropyl Acetate	0.467	0.412	11.8	90	0.00
44 TM	Trichloroethene	0.350	0.326	6.9	95	0.00
45 C	1,2-Dichloropropane	0.338	0.315	6.8#	95	0.00
46 T	Dibromomethane	0.190	0.177	6.8	95	0.00
47 T	Bromodichloromethane	0.482	0.452	6.2	96	0.00
48 T	Methyl methacrylate	0.219	0.194	11.4	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
 Data File : VY022498.D
 Acq On : 02 Jun 2025 14:59
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY060225

Quant Time: Jun 03 03:24:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 03 03:22:04 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	88	-0.01
50 S	Toluene-d8	1.206	1.128	6.5	92	0.00
51 T	4-Methyl-2-Pentanone	0.233	0.204	12.4	87	0.00
52 CM	Toluene	0.899	0.848	5.7#	98	0.00
53 T	t-1,3-Dichloropropene	0.452	0.417	7.7	94	0.00
54 T	cis-1,3-Dichloropropene	0.526	0.486	7.6	94	0.00
55 T	1,1,2-Trichloroethane	0.243	0.218	10.3	94	0.00
56 T	Ethyl methacrylate	0.351	0.316	10.0	91	0.00
57 T	1,3-Dichloropropane	0.430	0.393	8.6	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.159	0.155	2.5	101	0.00
59 T	2-Hexanone	0.156	0.137	12.2	88	0.00
60 T	Dibromochloromethane	0.308	0.278	9.7	92	0.00
61 T	1,2-Dibromoethane	0.221	0.203	8.1	95	0.00
62 S	4-Bromofluorobenzene	0.359	0.329	8.4	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.430	0.391	9.1	94	0.00
65 PM	Chlorobenzene	1.106	1.017	8.0	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.339	9.1	94	0.00
67 C	Ethyl Benzene	2.052	1.939	5.5#	98	0.00
68 T	m/p-Xylenes	0.778	0.727	6.6	97	0.00
69 T	o-Xylene	0.729	0.670	8.1	96	0.00
70 T	Styrene	1.206	1.129	6.4	96	0.00
71 P	Bromoform	0.198	0.178	10.1	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	4.105	3.853	6.1	97	0.00
74 T	N-amyl acetate	1.036	0.943	9.0	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.674	0.602	10.7	92	0.00
76 T	1,2,3-Trichloropropane	0.572	0.470	17.8	78	0.00
77 T	Bromobenzene	0.866	0.790	8.8	95	0.00
78 T	n-propylbenzene	5.014	4.683	6.6	96	0.00
79 T	2-Chlorotoluene	2.657	2.470	7.0	96	0.00
80 T	1,3,5-Trimethylbenzene	3.223	3.021	6.3	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.237	0.204	13.9	91	0.00
82 T	4-Chlorotoluene	2.741	2.538	7.4	95	0.00
83 T	tert-Butylbenzene	2.893	2.757	4.7	99	0.00
84 T	1,2,4-Trimethylbenzene	3.223	3.013	6.5	97	0.00
85 T	sec-Butylbenzene	4.430	4.186	5.5	98	0.00
86 T	p-Isopropyltoluene	3.652	3.456	5.4	98	0.00
87 T	1,3-Dichlorobenzene	1.755	1.612	8.1	97	0.00
88 T	1,4-Dichlorobenzene	1.702	1.536	9.8	94	0.00
89 T	n-Butylbenzene	3.464	3.324	4.0	98	0.00
90 T	Hexachloroethane	0.711	0.670	5.8	97	0.00
91 T	1,2-Dichlorobenzene	1.490	1.347	9.6	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.098	0.088	10.2	87	0.00
93 T	1,2,4-Trichlorobenzene	0.813	0.756	7.0	96	0.00
94 T	Hexachlorobutadiene	0.449	0.408	9.1	95	0.00
95 T	Naphthalene	1.560	1.404	10.0	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY060225\
Data File : VY022498.D
Acq On : 02 Jun 2025 14:59
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY060225

Quant Time: Jun 03 03:24:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 03 03:22:04 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.697	0.618	11.3	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6