

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2162 SAS No.: Q2162 SDG No.: Q2162

Instrument ID: MSVOA_Y Calibration Date/Time: 06/03/2025 10:03

Lab File ID: VY022511.D Init. Calib. Date(s): 06/02/2025 06/02/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 11:46 13:39

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	1.322	1.316	0.1	-0.45	20
Vinyl Chloride	1.534	1.560		1.7	20
Bromomethane	1.470	1.393		-5.24	20
Chloroethane	1.053	1.102		4.65	20
Trichlorofluoromethane	1.302	1.272		-2.3	20
1,1,2-Trichlorotrifluoroethane	0.549	0.523		-4.74	20
1,1-Dichloroethene	0.524	0.513		-2.1	20
Acetone	0.114	0.097		-14.91	20
Carbon Disulfide	1.679	1.680		0.06	20
Methyl tert-butyl Ether	1.477	1.494		1.15	20
Methylene Chloride	0.645	0.579		-10.23	20
trans-1,2-Dichloroethene	0.587	0.598		1.87	20
1,1-Dichloroethane	1.079	1.105	0.1	2.41	20
2-Butanone	0.164	0.160		-2.44	20
Carbon Tetrachloride	0.497	0.487		-2.01	20
cis-1,2-Dichloroethene	0.678	0.701		3.39	20
Chloroform	1.069	1.105		3.37	20
1,1,1-Trichloroethane	0.949	0.968		2	20
Methylcyclohexane	0.651	0.616		-5.38	20
Benzene	1.431	1.436		0.35	20
1,2-Dichloroethane	0.391	0.395		1.02	20
Trichloroethene	0.350	0.343		-2	20
1,2-Dichloropropane	0.338	0.343		1.48	20
Bromodichloromethane	0.482	0.494		2.49	20
4-Methyl-2-Pentanone	0.233	0.236		1.29	20
Toluene	0.899	0.901		0.22	20
t-1,3-Dichloropropene	0.452	0.460		1.77	20
cis-1,3-Dichloropropene	0.526	0.531		0.95	20
1,1,2-Trichloroethane	0.243	0.243		0	20
2-Hexanone	0.156	0.155		-0.64	20
Dibromochloromethane	0.308	0.312		1.3	20
Tetrachloroethene	0.430	0.422		-1.86	20
Chlorobenzene	1.106	1.118	0.3	1.09	20
Ethyl Benzene	2.052	2.059		0.34	20
m/p-Xylenes	0.778	0.765		-1.67	20
o-Xylene	0.729	0.713		-2.19	20
Styrene	1.206	1.220		1.16	20
Bromoform	0.198	0.195	0.1	-1.51	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2162 SAS No.: Q2162 SDG No.: Q2162

Instrument ID: MSVOA_Y Calibration Date/Time: 06/03/2025 10:03

Lab File ID: VY022511.D Init. Calib. Date(s): 06/02/2025 06/02/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 11:46 13:39

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	4.105	4.015		-2.19	20
1,1,2,2-Tetrachloroethane	0.674	0.669	0.3	-0.74	20
1,3-Dichlorobenzene	1.755	1.733		-1.25	20
1,4-Dichlorobenzene	1.702	1.659		-2.53	20
1,2-Dichlorobenzene	1.490	1.467		-1.54	20
1,2-Dichloroethane-d4	0.559	0.576		3.04	20
Dibromofluoromethane	0.298	0.300		0.67	20
Toluene-d8	1.206	1.224		1.49	20
4-Bromofluorobenzene	0.359	0.354		-1.39	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.