

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2162 SAS No.: Q2162 SDG NO.: Q2162
Lab File ID: VY022511.D Date Analyzed: 06/03/2025
Instrument ID: MSVOA_Y Time Analyzed: 10:03
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	187797	7.71	325302	8.62	276655	11.42
UPPER LIMIT	375594	8.213	650604	9.115	553310	11.92
LOWER LIMIT	93898.5	7.213	162651	8.115	138328	10.92
EPA SAMPLE NO.						
BP-VPB-182-GW-580-582	262567	7.71	482605	8.61	394621	11.41
VY0603SBL01	287398	7.71	534399	8.62	428748	11.42
VY0603SBS01	197612	7.71	337632	8.62	284268	11.42
VY0603SBSD01	196295	7.71	333757	8.62	277257	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.