

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2162 SAS No.: Q2162 SDG NO.: Q2162
Lab File ID: VX046409.D Date Analyzed: 05/30/2025
Instrument ID: MSVOA_X Time Analyzed: 10:14
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	64536	12.018				
UPPER LIMIT	129072	12.518				
LOWER LIMIT	32268	11.518				
EPA SAMPLE NO.						
BP-VPB-182-TB-20250527	54063	12.02				
BP-VPB-182-GW-600-602	52218	12.02				
BP-VPB-182-GW-600-602MS	62040	12.02				
BP-VPB-182-GW-600-602MSD	57548	12.02				
BP-VPB-182-GW-620-622	45099	12.02				
BP-VPB-182-GW-640-642	47411	12.02				
BP-VPB-182-EB-20250529	46505	12.02				
VX0530WBL01	51208	12.02				
VX0530WBS01	64120	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.