

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2162 SAS No.: Q2162 SDG NO.: Q2162
 Lab File ID: VY022511.D Date Analyzed: 06/03/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	130385	13.346				
UPPER LIMIT	260770	13.846				
LOWER LIMIT	65192.5	12.846				
EPA SAMPLE NO.						
BP-VPB-182-GW-580-582	147101	13.35				
VY0603SBL01	155262	13.35				
VY0603SBS01	131961	13.35				
VY0603SBSD01	127082	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.