

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2162 SAS No.: Q2162 SDG NO.: Q2162
Lab File ID: VY022488.D BFB Injection Date: 06/02/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:31
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	25.6
75	30.0 - 60.0% of mass 95	58.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.5 (0.6) 1
174	50.0 - 100.0% of mass 95	86.2
175	5.0 - 9.0% of mass 174	6.6 (7.6) 1
176	95.0 - 101.0% of mass 174	82.5 (95.6) 1
177	5.0 - 9.0% of mass 176	5.4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY022491.D	06/02/2025	11:46
VSTDIC010	VSTDIC010	VY022492.D	06/02/2025	12:09
VSTDIC020	VSTDIC020	VY022493.D	06/02/2025	12:32
VSTDIC050	VSTDIC050	VY022494.D	06/02/2025	12:54
VSTDIC100	VSTDIC100	VY022495.D	06/02/2025	13:17
VSTDIC150	VSTDIC150	VY022496.D	06/02/2025	13:39