

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04
 Lab Code: CHEM Case No.: Q2163 SAS No.: Q2163 SDG NO.: Q2163
 Lab File ID: VX046409.D Date Analyzed: 05/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:14
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	64536	12.018				
UPPER LIMIT	129072	12.518				
LOWER LIMIT	32268	11.518				
EPA SAMPLE NO.						
250528063-01	53676	12.02				
250528060-03-TRIP-BLANK	51011	12.02				
VX0530WBL01	51208	12.02				
VX0530WBS01	64120	12.02				
VX0530WBSD01	60852	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.