

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2164 SAS No.: Q2164 SDG NO.: Q2164

EPA Sample No.: SSTD020633 Date Analyzed: 06/05/2025

Lab File ID: BP024850.D Time Analyzed: 11:49

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	247295	7.611	1052780	10.38	683253	14.25
UPPER LIMIT	494590	8.111	2105560	10.881	1366510	14.751
LOWER LIMIT	123648	7.111	526390	9.881	341627	13.751
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	234878	7.61	942479	10.38	622427	14.25

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.