

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2164 SAS No.: Q2164 SDG NO.: Q2164

EPA Sample No.: SSTD020633 Date Analyzed: 06/05/2025

Lab File ID: BP024850.D Time Analyzed: 11:49

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1285770	17.051	1359430	21.48	1602490	24.733
UPPER LIMIT	2571540	17.551	2718860	21.98	3204980	25.233
LOWER LIMIT	642885	16.551	679715	20.98	801245	24.233
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	1295550	17.05	1421250	21.47	1655520	24.73

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.