

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Mahesh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133811
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP133838
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060225

Review By	John Carlone	Review On	6/3/2025 9:49:55 AM
Supervise By	Maresh Dadoda	Supervise On	6/3/2025 2:37:54 PM
SubDirectory	VX060225	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134094
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134095,VP134096

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046432.D	02 Jun 2025 09:42		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046433.D	02 Jun 2025 10:21	pH#Lot#V12668	JC/MD	Ok,M
3	VX0602MBL01	VX0602MBL01	VX046434.D	02 Jun 2025 10:51		JC/MD	Ok
4	VX0602WBL01	VX0602WBL01	VX046435.D	02 Jun 2025 11:14		JC/MD	Ok
5	VX0602WBS01	VX0602WBS01	VX046436.D	02 Jun 2025 11:37		JC/MD	Ok,M
6	VX0602WBSD01	VX0602WBSD01	VX046437.D	02 Jun 2025 12:04		JC/MD	Ok,M
7	IBLK	IBLK	VX046438.D	02 Jun 2025 12:27		JC/MD	Ok
8	Q2164-01	STORAGE-BLANK-SO	VX046439.D	02 Jun 2025 12:50	vial A pH<2	JC/MD	Ok
9	Q2164-02	STORAGE-BLANK-WA	VX046440.D	02 Jun 2025 13:14	vial A pH<2	JC/MD	Ok
10	Q2164-03	STORAGE-BLANK-WA	VX046441.D	02 Jun 2025 13:37	vial A pH<2	JC/MD	Ok
11	Q2164-04	STORAGE-BLANK-SAI	VX046442.D	02 Jun 2025 14:00	vial A pH<2	JC/MD	Ok
12	Q2177-08	EB05312025	VX046443.D	02 Jun 2025 14:23	vial A pH<2 EB;Hit of comp.#16	JC/MD	ReRun
13	Q2177-09	TB05312025	VX046444.D	02 Jun 2025 14:47	vial A pH<2 TB	JC/MD	Ok
14	Q2157-01	291	VX046445.D	02 Jun 2025 15:36	cloth material	JC/MD	Ok
15	VX0602MBS01	VX0602MBS01	VX046446.D	02 Jun 2025 15:59		JC/MD	Ok,M
16	VX0602MBSD01	VX0602MBSD01	VX046447.D	02 Jun 2025 16:22		JC/MD	Ok,M
17	Q2155-01	446	VX046448.D	02 Jun 2025 16:46	Need Straight Run	JC/MD	Not Ok
18	Q2155-02	447-SOLID	VX046449.D	02 Jun 2025 17:09	Need Soil Run	JC/MD	Not Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060225

Review By	John Carlone	Review On	6/3/2025 9:49:55 AM
Supervise By	Maresh Dadoda	Supervise On	6/3/2025 2:37:54 PM
SubDirectory	VX060225	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134094		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134095,VP134096		

19	Q2161-01	B27-SOIL-SAMPLE	VX046450.D	02 Jun 2025 17:32		JC/MD	Ok
20	Q2161-02	B28-SOIL-SAMPLE	VX046451.D	02 Jun 2025 17:56	Need Soil Run	JC/MD	Not Ok
21	IBLK	IBLK	VX046452.D	02 Jun 2025 18:19		JC/MD	Ok
22	IBLK	IBLK	VX046453.D	02 Jun 2025 18:42		JC/MD	Ok
23	Q2177-08	EB05312025	VX046454.D	02 Jun 2025 19:06	vial B pH<2	JC/MD	Ok
24	Q2169-01	303-PPR-1	VX046455.D	02 Jun 2025 19:29	Need Straight Run	JC/MD	Not Ok
25	Q2169-03	303-PPR-2	VX046456.D	02 Jun 2025 19:52	Need Straight Run	JC/MD	Not Ok
26	Q2155-01	446	VX046457.D	02 Jun 2025 20:39	Need Soil Run	JC/MD	Not Ok
27	VSTDCCC050	VSTDCCC050EC	VX046458.D	02 Jun 2025 21:03		JC/MD	Ok,M

M : Manual Integration