

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q2164 SAS No.: Q2164 SDG NO.: Q2164
 Lab File ID: VX046433.D Date Analyzed: 06/02/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:21
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	91247	5.54	155497	6.75	135003	10.05
UPPER LIMIT	182494	6.038	310994	7.251	270006	10.549
LOWER LIMIT	45623.5	5.038	77748.5	6.251	67501.5	9.549
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	66133	5.55	131706	6.76	124549	10.05
STORAGE-BLANK-WATER-REF-5	63283	5.54	125909	6.76	119085	10.05
STORAGE-BLANK-WATER-REF-4	69269	5.55	135855	6.76	129857	10.05
STORAGE-BLANK-SAMPLE-MANAGEMENT	63570	5.55	128178	6.76	122500	10.05
VX0602WBL01	66983	5.54	131566	6.76	125222	10.05
VX0602WBS01	88536	5.54	152496	6.76	133074	10.05
VX0602WBSD01	82810	5.54	146139	6.76	127160	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.