



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith SDG No.: Q2176
Contract: CAMP02 Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	104	70 - 130	CV	06/03/2025	12:52	LB135987
CRI01	Aluminum	105	100	105	65 - 135	P	06/10/2025	14:50	LB136097
	Antimony	53.3	50.0	107	65 - 135	P	06/10/2025	14:50	LB136097
	Arsenic	22.9	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Barium	98.7	100	99	65 - 135	P	06/10/2025	14:50	LB136097
	Beryllium	6.22	6.0	104	65 - 135	P	06/10/2025	14:50	LB136097
	Cadmium	6.07	6.0	101	65 - 135	P	06/10/2025	14:50	LB136097
	Calcium	2030	2000	102	65 - 135	P	06/10/2025	14:50	LB136097
	Chromium	10.7	10.0	107	65 - 135	P	06/10/2025	14:50	LB136097
	Cobalt	30.4	30.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Copper	22.7	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Iron	111	100	111	65 - 135	P	06/10/2025	14:50	LB136097
	Lead	13.4	12.0	112	65 - 135	P	06/10/2025	14:50	LB136097
	Magnesium	2040	2000	102	65 - 135	P	06/10/2025	14:50	LB136097
	Manganese	21.0	20.0	105	65 - 135	P	06/10/2025	14:50	LB136097
	Nickel	40.8	40.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Potassium	1990	2000	99	65 - 135	P	06/10/2025	14:50	LB136097
	Selenium	19.5	20.0	98	65 - 135	P	06/10/2025	14:50	LB136097
	Silver	10.6	10.0	106	65 - 135	P	06/10/2025	14:50	LB136097
	Sodium	1890	2000	95	65 - 135	P	06/10/2025	14:50	LB136097
	Thallium	40.8	40.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Vanadium	39.1	40.0	98	65 - 135	P	06/10/2025	14:50	LB136097
	Zinc	44.1	40.0	110	65 - 135	P	06/10/2025	14:50	LB136097