



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	CDM Smith				SDG No.:	Q2176				
Contract:	CAMP02	Lab Code:	CHEM		Case No.:	Q2176	SAS No.:	Q2176		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB25	Mercury	0.076	+/-0.2	U	0.20	CV	06/03/2025	12:46	LB135987	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2176							
Contract:	CAMP02	Lab Code:	CHEM	Case No.:	Q2176	SAS No.:	Q2176		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB90	Mercury	0.076	+/-0.2	U	0.20	CV	06/03/2025	12:50	LB135987
CCB91	Mercury	0.076	+/-0.2	U	0.20	CV	06/03/2025	13:22	LB135987
CCB92	Mercury	0.076	+/-0.2	U	0.20	CV	06/03/2025	13:51	LB135987
CCB93	Mercury	0.076	+/-0.2	U	0.20	CV	06/03/2025	14:18	LB135987
CCB94	Mercury	0.076	+/-0.2	U	0.20	CV	06/03/2025	14:51	LB135987
CCB95	Mercury	0.076	+/-0.2	U	0.20	CV	06/03/2025	15:04	LB135987

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith SDG No.: Q2176
Contract: CAMP02 Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	06/10/2025	14:45	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	14:45	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	14:45	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	14:45	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	14:45	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	14:45	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	14:45	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	14:45	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	14:45	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	14:45	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	14:45	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	14:45	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	14:45	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	14:45	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	14:45	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	14:45	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	14:45	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	14:45	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2176							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2176	SAS No.: Q2176						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	06/10/2025	15:35	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	15:35	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	15:35	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	15:35	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	15:35	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	15:35	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	15:35	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	15:35	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	15:35	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	15:35	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	15:35	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	15:35	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	15:35	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	15:35	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
CCB02	Aluminum	11.3	+/-50	U	100	P	06/10/2025	16:32	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	16:32	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	16:32	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	16:32	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	16:32	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	16:32	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	16:32	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	16:32	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	16:32	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	16:32	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	16:32	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	16:32	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	16:32	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	16:32	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	16:32	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	16:32	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2176							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2176	SAS No.: Q2176						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	06/10/2025	16:32	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	16:32	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
CCB03	Aluminum	11.3	+/-50	U	100	P	06/10/2025	17:40	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	17:40	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	17:40	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	17:40	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	17:40	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	17:40	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	17:40	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	17:40	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	17:40	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	17:40	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	17:40	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	17:40	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	17:40	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	17:40	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
CCB04	Aluminum	11.3	+/-50	U	100	P	06/10/2025	18:32	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	18:32	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	18:32	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	18:32	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	18:32	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	18:32	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	18:32	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	18:32	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	18:32	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	18:32	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2176							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2176	SAS No.: Q2176						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	06/10/2025	18:32	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	18:32	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	18:32	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	18:32	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
CCB05	Aluminum	11.3	+/-50	U	100	P	06/10/2025	19:22	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	19:22	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	19:22	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	19:22	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	19:22	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	19:22	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	19:22	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	19:22	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	19:22	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	19:22	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	19:22	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	19:22	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	19:22	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	19:22	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
CCB06	Aluminum	11.3	+/-50	U	100	P	06/10/2025	20:14	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	20:14	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	20:14	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	20:14	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	20:14	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	20:14	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2176							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2176	SAS No.: Q2176						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	20:14	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	20:14	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	20:14	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	20:14	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	20:14	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	20:14	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	20:14	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	20:14	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
CCB07	Aluminum	11.3	+/-50	U	100	P	06/10/2025	21:08	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	21:08	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	21:08	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	21:08	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	21:08	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	21:08	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	21:08	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	21:08	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	21:08	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	21:08	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	21:08	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	21:08	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	21:08	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	21:08	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
CCB08	Aluminum	11.3	+/-50	U	100	P	06/10/2025	22:08	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	22:08	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2176							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2176	SAS No.: Q2176						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	22:08	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	22:08	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	22:08	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	22:08	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	22:08	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	22:08	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	22:08	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	22:08	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	22:08	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	22:08	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	22:08	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	22:08	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
CCB09	Aluminum	11.3	+/-50	U	100	P	06/10/2025	23:01	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	23:01	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	23:01	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	23:01	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	23:01	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	23:01	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	23:01	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	23:01	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	23:01	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	23:01	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	23:01	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	23:01	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	23:01	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	23:01	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2176							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2176	SAS No.: Q2176						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
CCB10	Aluminum	11.3	+/-50	U	100	P	06/11/2025	00:02	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/11/2025	00:02	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Barium	14.6	+/-50	U	100	P	06/11/2025	00:02	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/11/2025	00:02	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	00:02	LB136097
	Calcium	234	+/-1000	U	2000	P	06/11/2025	00:02	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	00:02	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/11/2025	00:02	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Iron	23.4	+/-50	U	100	P	06/11/2025	00:02	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	00:02	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/11/2025	00:02	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Potassium	918	+/-1000	U	2000	P	06/11/2025	00:02	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/11/2025	00:02	LB136097
	Sodium	868	+/-1000	U	2000	P	06/11/2025	00:02	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Aluminum	11.3	+/-50	U	100	P	06/11/2025	01:39	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/11/2025	01:39	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	01:39	LB136097
	Barium	14.6	+/-50	U	100	P	06/11/2025	01:39	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/11/2025	01:39	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	01:39	LB136097
	Calcium	234	+/-1000	U	2000	P	06/11/2025	01:39	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	01:39	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/11/2025	01:39	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	01:39	LB136097
	Iron	23.4	+/-50	U	100	P	06/11/2025	01:39	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	01:39	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/11/2025	01:39	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/11/2025	01:39	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith		SDG No.: Q2176							
Contract: CAMP02	Lab Code: CHEM	Case No.: Q2176	SAS No.: Q2176						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
	Potassium	918	+/-1000	U	2000	P	06/11/2025	01:39	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/11/2025	01:39	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/11/2025	01:39	LB136097
	Sodium	868	+/-1000	U	2000	P	06/11/2025	01:39	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
CCB12	Aluminum	11.3	+/-50	U	100	P	06/11/2025	02:25	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/11/2025	02:25	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Barium	14.6	+/-50	U	100	P	06/11/2025	02:25	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/11/2025	02:25	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	02:25	LB136097
	Calcium	234	+/-1000	U	2000	P	06/11/2025	02:25	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	02:25	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/11/2025	02:25	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Iron	23.4	+/-50	U	100	P	06/11/2025	02:25	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	02:25	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/11/2025	02:25	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	02:25	LB136097
	Potassium	918	+/-1000	U	2000	P	06/11/2025	02:25	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/11/2025	02:25	LB136097
	Sodium	868	+/-1000	U	2000	P	06/11/2025	02:25	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/11/2025	02:25	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/11/2025	02:25	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	02:25	LB136097