

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CAMP02
ProjectID: South River WM Replacement
Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG No.: Q2176
DataFile: FG015924.D Analyst Name: YP\AJ Analyst Date: 06-04-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	63833620	127667	132460	3.618