

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB053025

Review By	yogesh	Review On	5/30/2025 3:36:07 PM
Supervise By	mohammad	Supervise On	6/3/2025 2:03:52 AM
SubDirectory	FB053025	HP Acquire Method	HP Processing Method
			FB053025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24606,PP24607,PP24608,PP24609,PP24610		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24611		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB031830.D	30 May 2025 10:13		YP/AJ	Ok,M
2	10 GRO STD		FB031831.D	30 May 2025 10:41		YP/AJ	Ok,M
3	20 GRO STD		FB031832.D	30 May 2025 11:08		YP/AJ	Ok
4	50 GRO STD		FB031833.D	30 May 2025 11:36		YP/AJ	Ok
5	100 GRO STD		FB031834.D	30 May 2025 12:12		YP/AJ	Ok
6	FB053025GROICV		FB031835.D	30 May 2025 15:26		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB060325

Review By	yogesh	Review On	6/3/2025 11:24:47 AM
Supervise By	mohammad	Supervise On	6/5/2025 1:35:21 AM
SubDirectory	FB060325	HP Acquire Method	HP Processing Method FB053025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24606,PP24607,PP24608,PP24609,PP24610		
CCC Internal Standard/PEM	PP24615,PP24616,PP24617		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24611		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031856.D	3 Jun 2025 9:12		YP/AJ	Ok
2	VBF0603S1		FB031857.D	3 Jun 2025 9:55		YP/AJ	Ok
3	VBF0603S2		FB031858.D	3 Jun 2025 10:23		YP/AJ	Ok
4	BSF0603S1		FB031859.D	3 Jun 2025 10:51		YP/AJ	Ok
5	Q2150-09		FB031860.D	3 Jun 2025 11:18	Vial- B	YP/AJ	Ok
6	Q2150-10		FB031861.D	3 Jun 2025 11:46	Vial- B, Not Purged	YP/AJ	Not Ok
7	Q2150-04		FB031862.D	3 Jun 2025 12:38	Vial- C,not required	YP/AJ	Not Ok
8	Q2150-10		FB031863.D	3 Jun 2025 13:05	Vial-C	YP/AJ	Ok
9	Q2126-09		FB031864.D	3 Jun 2025 13:33	Vial- A	YP/AJ	Ok,M
10	BSF0603S2		FB031865.D	3 Jun 2025 14:24		YP/AJ	Ok
11	20 PPB GRO STD		FB031866.D	3 Jun 2025 14:52		YP/AJ	Ok
12	Q2176-01		FB031867.D	3 Jun 2025 15:52	Vial-A, Not purge	YP/AJ	Not Ok
13	Q2176-02		FB031868.D	3 Jun 2025 16:20	Vial-A	YP/AJ	Ok
14	Q2176-03		FB031869.D	3 Jun 2025 16:48	Vial-A	YP/AJ	Ok
15	Q2176-04		FB031870.D	3 Jun 2025 17:15	Vial-A	YP/AJ	Ok
16	Q2176-05		FB031871.D	3 Jun 2025 17:43	Vial-A	YP/AJ	Ok
17	Q2176-06		FB031872.D	3 Jun 2025 18:10	Vial-A	YP/AJ	Ok
18	Q2176-07		FB031873.D	3 Jun 2025 18:38	Vial-A	YP/AJ	Ok,M

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB060325

Review By	yogesh	Review On	6/3/2025 11:24:47 AM
Supervise By	mohammad	Supervise On	6/5/2025 1:35:21 AM
SubDirectory	FB060325	HP Acquire Method	HP Processing Method
			FB053025

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24606,PP24607,PP24608,PP24609,PP24610
CCC Internal Standard/PEM	PP24615,PP24616,PP24617
ICV/I.BLK Surrogate Standard	PP24111,PP24611
MS/MSD Standard	
LCS Standard	

19	Q2176-08		FB031874.D	3 Jun 2025 19:06	Vial-A, Not Purged	YP/AJ	Not Ok
20	BSF0603S3		FB031875.D	3 Jun 2025 19:33		YP/AJ	Ok
21	20 PPB GRO STD		FB031876.D	3 Jun 2025 20:01		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB060425

Review By	yogesh	Review On	6/4/2025 10:57:57 AM
Supervise By		Supervise On	
SubDirectory	FB060425	HP Acquire Method	HP Processing Method FB053025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24606,PP24607,PP24608,PP24609,PP24610		
CCC Internal Standard/PEM	PP24618,PP24619,PP24620		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24611		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031877.D	4 Jun 2025 9:36		YP/AJ	Ok
2	VBF0604S1		FB031878.D	4 Jun 2025 10:19		YP/AJ	Ok
3	VBF0604S2		FB031879.D	4 Jun 2025 10:47		YP/AJ	Ok
4	BSF0604S1		FB031880.D	4 Jun 2025 11:14		YP/AJ	Ok
5	Q2176-01		FB031881.D	4 Jun 2025 11:42	Vial-B, Not Purged	YP/AJ	Not Ok
6	Q2176-08		FB031882.D	4 Jun 2025 12:10	Vial-B, Not Purged	YP/AJ	Not Ok
7	Q2176-01		FB031883.D	4 Jun 2025 12:48	Vial-C	YP/AJ	Ok
8	Q2176-08		FB031884.D	4 Jun 2025 13:16	Vial-C	YP/AJ	Ok
9	BSF0604S2		FB031885.D	4 Jun 2025 13:44		YP/AJ	Ok
10	20 PPB GRO STD		FB031886.D	4 Jun 2025 14:11		YP/AJ	Ok
11	VBF0604W1		FB031887.D	4 Jun 2025 15:07		YP/AJ	Ok
12	BSF0604W1		FB031888.D	4 Jun 2025 15:35		YP/AJ	Ok
13	Q2126-09		FB031889.D	4 Jun 2025 16:02	Vial- A	YP/AJ	Ok,NS
14	BSF0604W2		FB031890.D	4 Jun 2025 16:42		YP/AJ	Ok
15	20 PPB GRO STD		FB031891.D	4 Jun 2025 17:41		YP/AJ	Ok

M : Manual Integration