

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL01 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030497.D Time Analyzed: 20:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.692	9.597	9.797	729.950	712.500	2.4
2,4,5-TP(Silvex)	9.398	9.304	9.504	724.280	712.500	1.7
2,4-D	8.506	8.412	8.612	705.370	705.000	0.1
2,4-DB	10.273	10.179	10.379	730.780	712.500	2.6
2,4-DCAA	7.365	7.269	7.469	725.600	750.000	-3.3
DICAMBA	7.557	7.461	7.661	710.910	705.000	0.8
DICHLORPROP	8.273	8.179	8.379	692.230	705.000	-1.8
Dinoseb	11.495	11.402	11.602	717.380	705.000	1.8

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL01 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030497.D Time Analyzed: 20:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.365	10.268	10.468	710.770	712.500	-0.2
2,4,5-TP(Silvex)	9.939	9.842	10.042	709.400	712.500	-0.4
2,4-D	9.032	8.934	9.134	695.700	705.000	-1.3
2,4-DB	10.933	10.836	11.036	693.450	712.500	-2.7
2,4-DCAA	7.772	7.674	7.874	722.520	750.000	-3.7
DICAMBA	7.976	7.877	8.077	701.350	705.000	-0.5
DICHLORPROP	8.696	8.598	8.798	691.250	705.000	-2.0
Dinoseb	11.319	11.221	11.421	702.840	705.000	-0.3