

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030516.D Time Analyzed: 04:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.690	9.597	9.797	707.790	712.500	-0.7
2,4,5-TP(Silvex)	9.395	9.304	9.504	720.680	712.500	1.1
2,4-D	8.504	8.412	8.612	697.490	705.000	-1.1
2,4-DB	10.269	10.179	10.379	681.810	712.500	-4.3
2,4-DCAA	7.364	7.269	7.469	713.260	750.000	-4.9
DICAMBA	7.555	7.461	7.661	714.070	705.000	1.3
DICHLORPROP	8.271	8.179	8.379	693.020	705.000	-1.7
Dinoseb	11.491	11.402	11.602	709.130	705.000	0.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030516.D Time Analyzed: 04:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.365	10.268	10.468	707.320	712.500	-0.7
2,4,5-TP(Silvex)	9.939	9.842	10.042	709.690	712.500	-0.4
2,4-D	9.031	8.934	9.134	727.630	705.000	3.2
2,4-DB	10.932	10.836	11.036	696.340	712.500	-2.3
2,4-DCAA	7.772	7.674	7.874	709.210	750.000	-5.4
DICAMBA	7.976	7.877	8.077	698.330	705.000	-0.9
DICHLORPROP	8.696	8.598	8.798	689.030	705.000	-2.3
Dinoseb	11.318	11.221	11.421	699.890	705.000	-0.7