

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030526.D Time Analyzed: 08:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.688	9.597	9.797	693.230	712.500	-2.7
2,4,5-TP(Silvex)	9.393	9.304	9.504	715.010	712.500	0.4
2,4-D	8.502	8.412	8.612	687.060	705.000	-2.5
2,4-DB	10.268	10.179	10.379	658.560	712.500	-7.6
2,4-DCAA	7.362	7.269	7.469	717.160	750.000	-4.4
DICAMBA	7.554	7.461	7.661	713.070	705.000	1.1
DICHLORPROP	8.270	8.179	8.379	689.740	705.000	-2.2
Dinoseb	11.490	11.402	11.602	712.210	705.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030526.D Time Analyzed: 08:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.363	10.268	10.468	707.860	712.500	-0.7
2,4,5-TP(Silvex)	9.937	9.842	10.042	709.480	712.500	-0.4
2,4-D	9.030	8.934	9.134	722.280	705.000	2.5
2,4-DB	10.930	10.836	11.036	688.890	712.500	-3.3
2,4-DCAA	7.771	7.674	7.874	726.710	750.000	-3.1
DICAMBA	7.974	7.877	8.077	709.610	705.000	0.7
DICHLORPROP	8.694	8.598	8.798	691.730	705.000	-1.9
Dinoseb	11.316	11.221	11.421	708.340	705.000	0.5