

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL05 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030529.D Time Analyzed: 10:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.688	9.597	9.797	696.360	712.500	-2.3
2,4,5-TP(Silvex)	9.393	9.304	9.504	720.870	712.500	1.2
2,4-D	8.502	8.412	8.612	690.150	705.000	-2.1
2,4-DB	10.268	10.179	10.379	655.240	712.500	-8.0
2,4-DCAA	7.361	7.269	7.469	744.640	750.000	-0.7
DICAMBA	7.553	7.461	7.661	720.790	705.000	2.2
DICHLORPROP	8.269	8.179	8.379	693.830	705.000	-1.6
Dinoseb	11.490	11.402	11.602	718.240	705.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL05 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030529.D Time Analyzed: 10:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.362	10.268	10.468	713.650	712.500	0.2
2,4,5-TP(Silvex)	9.937	9.842	10.042	717.800	712.500	0.7
2,4-D	9.030	8.934	9.134	710.700	705.000	0.8
2,4-DB	10.930	10.836	11.036	698.670	712.500	-1.9
2,4-DCAA	7.771	7.674	7.874	730.660	750.000	-2.6
DICAMBA	7.974	7.877	8.077	713.970	705.000	1.3
DICHLORPROP	8.694	8.598	8.798	696.090	705.000	-1.3
Dinoseb	11.316	11.221	11.421	716.500	705.000	1.6