

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL06 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030532.D Time Analyzed: 15:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.688	9.597	9.797	700.170	712.500	-1.7
2,4,5-TP(Silvex)	9.394	9.304	9.504	723.990	712.500	1.6
2,4-D	8.503	8.412	8.612	691.070	705.000	-2.0
2,4-DB	10.268	10.179	10.379	664.630	712.500	-6.7
2,4-DCAA	7.362	7.269	7.469	741.180	750.000	-1.2
DICAMBA	7.554	7.461	7.661	721.180	705.000	2.3
DICHLORPROP	8.270	8.179	8.379	696.700	705.000	-1.2
Dinoseb	11.491	11.402	11.602	723.120	705.000	2.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL06 Date Analyzed: 06/05/2025

Lab Sample No.: HSTDCCC750 Data File : PS030532.D Time Analyzed: 15:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.362	10.268	10.468	712.190	712.500	0.0
2,4,5-TP(Silvex)	9.937	9.842	10.042	713.500	712.500	0.1
2,4-D	9.029	8.934	9.134	708.180	705.000	0.5
2,4-DB	10.929	10.836	11.036	695.800	712.500	-2.3
2,4-DCAA	7.770	7.674	7.874	727.740	750.000	-3.0
DICAMBA	7.974	7.877	8.077	710.730	705.000	0.8
DICHLORPROP	8.694	8.598	8.798	691.060	705.000	-2.0
Dinoseb	11.315	11.221	11.421	713.050	705.000	1.1