

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO111453.D Time Analyzed: 08:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.768	4.673	4.873	543.320	500.000	8.7
Aroclor-1016-2	4.787	4.692	4.892	544.470	500.000	8.9
Aroclor-1016-3	4.844	4.749	4.949	560.720	500.000	12.1
Aroclor-1016-4	4.964	4.869	5.069	549.990	500.000	10.0
Aroclor-1016-5	5.220	5.126	5.326	525.450	500.000	5.1
Aroclor-1260-1	6.259	6.166	6.366	552.270	500.000	10.5
Aroclor-1260-2	6.449	6.355	6.555	576.950	500.000	15.4
Aroclor-1260-3	6.816	6.722	6.922	569.720	500.000	13.9
Aroclor-1260-4	7.076	6.983	7.183	529.440	500.000	5.9
Aroclor-1260-5	7.319	7.225	7.425	525.340	500.000	5.1
Decachlorobiphenyl	8.716	8.624	8.824	47.630	50.000	-4.7
Tetrachloro-m-xylene	3.679	3.582	3.782	52.690	50.000	5.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PO111453.D Time Analyzed: 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.658	4.858	484.590	500.000	-3.1
Aroclor-1016-2	4.772	4.677	4.877	492.050	500.000	-1.6
Aroclor-1016-3	4.947	4.852	5.052	485.980	500.000	-2.8
Aroclor-1016-4	4.989	4.894	5.094	486.580	500.000	-2.7
Aroclor-1016-5	5.202	5.107	5.307	481.390	500.000	-3.7
Aroclor-1260-1	6.232	6.138	6.338	486.550	500.000	-2.7
Aroclor-1260-2	6.420	6.326	6.526	481.620	500.000	-3.7
Aroclor-1260-3	6.572	6.479	6.679	448.480	500.000	-10.3
Aroclor-1260-4	7.042	6.949	7.149	463.240	500.000	-7.4
Aroclor-1260-5	7.283	7.191	7.391	448.590	500.000	-10.3
Decachlorobiphenyl	8.665	8.574	8.774	44.820	50.000	-10.4
Tetrachloro-m-xylene	3.676	3.579	3.779	51.190	50.000	2.4