

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL07 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP072619.D Time Analyzed: 14:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.653	5.552	5.752	408.380	500.000	-18.3
Aroclor-1016-2	5.674	5.574	5.774	446.380	500.000	-10.7
Aroclor-1016-3	5.736	5.636	5.836	438.840	500.000	-12.2
Aroclor-1016-4	5.834	5.734	5.934	447.960	500.000	-10.4
Aroclor-1016-5	6.126	6.027	6.227	453.310	500.000	-9.3
Aroclor-1260-1	7.243	7.147	7.347	454.670	500.000	-9.1
Aroclor-1260-2	7.497	7.401	7.601	456.280	500.000	-8.7
Aroclor-1260-3	7.857	7.760	7.960	575.660	500.000	15.1
Aroclor-1260-4	8.078	7.985	8.185	472.200	500.000	-5.6
Aroclor-1260-5	8.397	8.303	8.503	486.080	500.000	-2.8
Decachlorobiphenyl	10.199	10.113	10.313	50.400	50.000	0.8
Tetrachloro-m-xylene	4.501	4.396	4.596	46.060	50.000	-7.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2176 SAS No.: Q2176 SDG NO.: Q2176

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL07 Date Analyzed: 06/04/2025

Lab Sample No.: AR1660CCC500 Data File : PP072619.D Time Analyzed: 14:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	478.490	500.000	-4.3
Aroclor-1016-2	4.896	4.795	4.995	490.280	500.000	-1.9
Aroclor-1016-3	5.072	4.973	5.173	492.560	500.000	-1.5
Aroclor-1016-4	5.114	5.015	5.215	499.290	500.000	-0.1
Aroclor-1016-5	5.328	5.229	5.429	516.590	500.000	3.3
Aroclor-1260-1	6.359	6.264	6.464	500.890	500.000	0.2
Aroclor-1260-2	6.548	6.453	6.653	478.230	500.000	-4.4
Aroclor-1260-3	6.700	6.606	6.806	510.370	500.000	2.1
Aroclor-1260-4	7.170	7.077	7.277	505.570	500.000	1.1
Aroclor-1260-5	7.412	7.319	7.519	505.920	500.000	1.2
Decachlorobiphenyl	8.811	8.723	8.923	53.050	50.000	6.1
Tetrachloro-m-xylene	3.797	3.691	3.891	49.150	50.000	-1.7